

Installation Note

Receiver Attenuators Upgrade Kit

For E8362B/C PNA Series Microwave Network Analyzers

Network Analyzer Model Number	Upgrade Kit Part Number
E8362B/C	E8362-60107




Agilent Part Number: E8362-90008

Printed in USA December 2007

Supersedes print date: September 2007

© Agilent Technologies, Inc. 2003, 2005, 2007



E8362-90008

Contacting Agilent

Assistance with test and measurements needs and information on finding a local Agilent office are available on the Web at:

www.agilent.com/find/assist

If you do not have access to the Internet, please contact your Agilent field engineer.

NOTE	In any correspondence or telephone conversation, refer to the Agilent product by its model number and full serial number. With this information, the Agilent representative can determine whether your product is still within its warranty period.
-------------	---

About Installing the Upgrade Kit

IMPORTANT	Option 016 can only be installed on analyzers with the option combinations listed below as “Products affected”. If your analyzer does not have the proper options (as listed below), it will be necessary to install those options BEFORE installing this option (Option 016).
------------------	--

Products affectedE8362B/C; without Option 014 E8362B/C; Option 014 (without Option 081) E8362B/C; Option 014 and 081
--

Installation to be performed byAgilent service center or personnel qualified by Agilent

Estimated installation time2.0 hours

Estimated adjustment time0.5 hours

Estimated full instrument calibration time4.5 hours

Description of Option 016 Receiver Attenuators

This option adds a 35-dB step attenuator in the channel A and channel B signal paths.

Items Included in the Upgrade Kit

Check the contents of your kit against this list. If any item is missing or damaged, contact Agilent Technologies. Refer to “Contacting Agilent” on page 2.

Table 1 Contents of Option 016 Upgrade Kit (E8362-60107)

Ref. Desig.	Description	Qty	Part Number
These parts are for all analyzers			
	Installation note (this document)	1	E8362-90008
	Machine screw, M3.0 x 8 CWPNTX (for attaching attenuators)	4	0515-0372
	Cable clamp	1	1400-1439
A43, A44	35-dB step attenuator	2	33325-60011
	Ribbon cable (for A43 and A44 step attenuators)	2	8121-0819
W47	RF cable, A43 channel A step attenuator to A27 channel A mixer	1	E8364-20131
W48	RF cable, A44 channel B step attenuator to A30 channel B mixer	1	E8364-20132
These parts are for analyzers with Option UNL but WITHOUT Option 014			
W7	RF cable, channel R1 attenuator to A28 channel R1 mixer	1	E8362-20039
W8	RF cable, channel R2 attenuator to A29 channel R2 mixer	1	E8362-20040
W45	RF cable, A25 test port 1 coupler to A43 channel A step attenuator	1	E8364-20147
W46	RF cable, A26 test port 2 coupler to A44 channel B step attenuator	1	E8364-20148
These parts are for analyzers WITH Options UNL and 014 but WITHOUT Option 081			
	Lower front panel overlay (Option UNL/016)	1	E8364-80024
W49	RF cable, PORT 1 RCVR A IN to A43 channel A step attenuator	1	E8362-20035
W50	RF cable, PORT 2 RCVR B IN to A44 channel B step attenuator	1	E8362-20036
W70	RF cable, REFERENCE 1 RCVR R1 IN to A28 channel R1 mixer	1	E8362-20041
W71	RF cable, REFERENCE 2 RCVR R2 IN to A29 channel R2 mixer	1	E8362-20038
These parts are for analyzers WITH Options UNL, 014, AND Option 081			
	Lower front panel overlay (Option UNL/014/016)	1	E8364-80024
W49	RF cable, PORT 1 RCVR A IN to A43 channel A step attenuator	1	E8362-20035
W50	RF cable, PORT 2 RCVR B IN to A44 channel B step attenuator	1	E8362-20036
W99	RF cable, A45 reference switch to A28 channel R1 mixer	1	E8362-20037
W71	RF cable, REFERENCE 2 RCVR R2 IN to A29 channel R2 mixer	1	E8362-20038

Installation Procedure for the Upgrade Kit

The network analyzer must be in proper working condition prior to installing this option. Any necessary repairs must be made before proceeding with this installation.

WARNING **This installation requires the removal of the analyzer's protective outer covers. The analyzer must be powered down and disconnected from the mains supply before performing this procedure.**

Electrostatic Discharge Protection


Protection against electrostatic discharge (ESD) is essential while removing or connecting cables or assemblies within the network analyzer.

Static electricity can build up on your body and can easily damage sensitive internal circuit elements when discharged. Static discharges too small to be felt can cause permanent damage. To prevent damage to the instrument:

- *always* wear a grounded wrist strap having a 1 M Ω resistor in series with it when handling components and assemblies.
- *always* use a grounded, conductive table mat while working on the instrument.
- *always* wear a heel strap when working in an area with a conductive floor. If you are uncertain about the conductivity of your floor, wear a heel strap.
- *always* ground yourself before you clean, inspect, or make a connection to a static-sensitive device or test port. You can, for example, grasp the grounded outer shell of the test port or cable connector briefly.

Figure 1 shows a typical ESD protection setup using a grounded mat and wrist strap. Refer to “Tools and Equipment Required for the Installation” on page 5 for part numbers.

Figure 1 ESD Protection Setup



Overview of the Installation Procedure

- Step 1. Remove the Outer Cover.
- Step 2. Remove the Front Panel Assembly (Option 014 Only).
- Step 3. Raise the Receiver Deck.
- Step 4. Remove the Existing Cables.
- Step 5. Install the Attenuators.
- Step 6. Install the Option 016 Cables.
- Step 7. Lower and Fasten the Receiver Deck.
- Step 8. Replace the Lower Front Panel Overlay (Option 014 Only).
- Step 9. Reinstall the Front Panel Assembly and Front Panel Jumpers (Option 014 Only).
- Step 10. Reinstall the Outer Cover.
- Step 11. Enable Option 016.
- Step 12. Perform Post-Upgrade Adjustments and Calibration.

Tools and Equipment Required for the Installation

Description	Qty	Part Number
T-10 TORX driver (set to 9 in-lbs)	1	N/A
T-20 TORX driver (set to 21 in-lbs)	1	N/A
5/16-inch torque wrench (set to 10 in-lbs)	1	N/A
5/16-inch torque wrench (set to 21 in-lbs) (Option 014 only)	1	N/A
ESD grounding wrist strap	1	9300-1367
5-ft grounding cord for wrist strap	1	9300-0980
2 x 4 ft conductive table mat and 15-ft grounding wire	1	9300-0797
ESD heel strap (for use with conductive floors)	1	9300-1308

CAUTION	Use a 5/16-inch torque wrench set to 10 in-lbs on all cable connections except the front-panel connectors to which the front-panel jumpers attach (Option 014). Use a 5/16-inch torque wrench set to 21 in-lbs for these connections.
----------------	--

Equipment Required for Post-Upgrade Adjustments

Equipment Type	Model or Part Number	Alternate Model or Part Number
Power meter	E4418B/E4419B	E4418A/E4419A
Power sensor, 3.5 mm	E4413A	8485A
Adapter, 3.5 mm (f) to 3.5 mm (f)	83059B	85052-60012
RF cable, 3.5 mm (f) to 3.5 mm (f)	85131C	85131E


Step 1. Remove the Outer Cover

CAUTION This procedure is best performed with the analyzer resting on its front handles in the vertical position. *Do not place the analyzer on its front panel without the handles.* This will damage the front panel assemblies.

Refer to Figure 2 for this procedure.

1. Disconnect the power cord (if it has not already been disconnected).
2. With a T-20 TORX driver, remove the strap handles (item ①) by loosening the screws (item ②) on both ends until the handle is free of the analyzer.
3. With a T-20 TORX driver, remove the four rear panel feet (item ③) by removing the center screws (item ④).
4. Slide the four bottom feet (item ⑤) off the cover.
5. Slide the cover off of the frame.

Figure 2 Outer Cover Removal



dt501a

NOTE The figure above shows the E8362B front panel and floppy disk drive. The E8362C front panel has a slightly different appearance and does not include a floppy disk drive.

Step 2. Remove the Front Panel Assembly (Option 014 Only)


Refer to Figure 3 for this procedure.

1. With a 5/16-inch wrench, remove the six front panel semirigid jumper cables (item ①).
2. With a T-10 TORX driver, remove the eight screws (item ②) from the sides of the frame.

CAUTION Before removing the front panel from the analyzer, lift and support the front of the analyzer chassis.

3. Slide the front panel over the test port connectors.
4. Disconnect the front panel interface ribbon cable (item ③) from the A3 front panel interface board. The front panel is now free from the analyzer.

Figure 3 Front Panel Assembly Removal



st536a


NOTE The figure above shows the E8362B front panel and floppy disk drive. The E8362C front panel has a slightly different appearance and does not include a floppy disk drive.

Step 3. Raise the Receiver Deck

Refer to Figure 4 for this procedure.

1. Place the analyzer bottom-side up on a flat surface.
2. With a T-10 TORX driver, remove the four screws, (item ①), that secure the receiver deck.
3. Pull the latch pin (item ②) towards the center of the analyzer to release the receiver deck.
4. Lift the receiver deck to partially raise it, then release the latch pin (item ②). Lift the receiver deck to its fully raised position and ensure that the latch pin latches in the raised position.

Figure 4 Receiver Deck, Raising



Step 4. Remove the Existing Cables


Analyzers WITH Option UNL but WITHOUT Option 014

Refer to Figure 5 for the following procedure.

Remove the following cables:

- W7 E8362-20027 Channel R1 fixed attenuator to A28 channel R1 mixer
- W8 E8362-20028 Channel R2 fixed attenuator to A29 channel R2 mixer
- W9 E8364-20019 A25 test port 1 coupler to A27 channel A mixer
- W10 E8364-20020 A26 test port 2 coupler to A30 channel B mixer

Figure 5 Cable Removal, Analyzers with Option UNL but without Option 014



8362_008_01


Analyzers WITH Options UNL and 014 but WITHOUT Option 081

Refer to Figure 6 for the following procedure.

Remove the following cables:

- W69 E8362-20022 PORT 1 RCVR A IN to A27 channel A mixer
- W72 E8362-20023 PORT 2 RCVR B IN to A30 channel B mixer
- W71 E8362-20015 REFERENCE 2 RCVR R2 IN to A29 channel R2 mixer
- W70 E8362-20014 REFERENCE 1 RCVR R1 IN to A28 channel R1 mixer

Figure 6 Cable Removal, Analyzers with Options UNL and 014 but without Option 081



8362_008_02


Analyzers WITH Options UNL, 014, AND 081

Refer to Figure 7 for the following procedure.

Remove the following cables:

- W69 E8362-20022 PORT 1 RCVR A IN to A27 channel A mixer
- W72 E8362-20023 PORT 2 RCVR B IN to A30 channel B mixer
- W71 E8362-20046 REFERENCE 2 RCVR R2 IN to A29 channel R2 mixer
- W99 E8362-20045 A45 reference switch to A28 channel R1 mixer

Figure 7 Cable Removal, Analyzers with Options UNL, 014, and 081



8362_008_03


Step 5. Install the Attenuators

Refer to Figure 8 for the following procedure.

- Attach one step attenuator to each bracket, as shown, using two M3.0 x 8 screws (provided) for each.

Be careful to position the step attenuators so that the necessary cables can be attached. The end of the step attenuator with the ribbon cable connector must face toward the outside of the analyzer. Review the cable connections in Figure 8 if necessary.

Figure 8 Attenuator Installation



8362_008_4

Step 6. Install the Option 016 Cables


Analyzers WITH Option UNL but WITHOUT Option 014

Refer to Figure 9 for the following procedure. The new parts referenced in this procedure are listed in Table 1 on page 3.

Install the following cables in the order listed:

- Ribbon cable ① 8121-0819 A43 ch A attenuator to A16 motherboard (P1 RCVR ATT)
- Ribbon cable ② 8121-0819 A44 ch B attenuator to A16 motherboard (P2 RCVR ATT)
- W47 E8364-20131 A43 channel A step attenuator to A27 channel A mixer
- W48 E8364-20132 A44 channel B step attenuator to A30 channel B mixer
- W8 E8362-20040 Channel R2 fixed attenuator to A29 channel R2 mixer
- W7 E8362-20039 Channel R1 fixed attenuator to A28 channel R1 mixer
- W46 E8364-20148 A26 test port 2 coupler to A43 channel A step attenuator
- W45 E8364-20147 A25 test port 1 coupler to A44 channel B step attenuator

Figure 9 Cable Installation, Analyzers with Option UNL but without Option 014



8362_008_5


Analyzers WITH Options UNL and 014 but WITHOUT Option 081

Refer to Figure 10 for the following procedure. The new parts referenced in this procedure are listed in Table 1 on page 3.

Install the following cables in the order listed:

- Ribbon cable ① 8121-0819 A43 ch A attenuator to A16 motherboard (P1 RCVR ATT)
- Ribbon cable ② 8121-0819 A44 ch B attenuator to A16 motherboard (P2 RCVR ATT)
- W47 E8364-20131 A43 channel A step attenuator to A27 channel A mixer
- W48 E8364-20132 A44 channel B step attenuator to A30 channel B mixer
- W71 E8362-20038 REFERENCE 2 RCVR R2 IN to A29 channel R2 mixer
- W70 E8362-20041 REFERENCE 1 RCVR R1 IN to A28 channel R1 mixer
- W50 E8362-20036 PORT 2 RCVR B IN to A44 channel B step attenuator
- W49 E8362-20035 PORT 1 RCVR A IN to A43 channel A step attenuator

Figure 10 Cable Installation, Analyzers with Options UNL and 014 but without Option 081



8362_008_06


Analyzers WITH Options UNL, 014, AND 081

Refer to Figure 11 for the following procedure. The new parts referenced in this procedure are listed in Table 1 on page 3.

Install the following cables in the order listed:

- Ribbon cable ① 8121-0819 A43 ch A attenuator to A16 motherboard (P1 RCVR ATT)
- Ribbon cable ② 8121-0819 A44 ch B attenuator to A16 motherboard (P2 RCVR ATT)
- W47 E8364-20131 A43 channel A step attenuator to A27 channel A mixer
- W48 E8364-20132 A44 channel B step attenuator to A30 channel B mixer
- W71 E8362-20038 REFERENCE 1 RCVR R2 IN to A29 channel R2 mixer
- W99 E8362-20037 A45 reference switch to A28 channel R1 mixer
- W50 E8362-20036 PORT 2 RCVR B IN to A44 channel B step attenuator
- W49 E8362-20035 PORT 1 RCVR A IN to A43 channel A step attenuator

Figure 11 Cable Installation, Analyzers with Options UNL, 014, and 081




8362_008_07

Step 7. Lower and Fasten the Receiver Deck

Refer to Figure 12 for this procedure.

1. Pull the latch pin (item ②) toward the center of the analyzer to release the receiver deck.
2. Lift the receiver deck to partially lower it, then release the latch pin (item ②). Lower the receiver deck to its fully lowered position and ensure that the latch pin latches in the lowered position.
3. With a T-10 TORX driver, install the four screws (item ①) to secure the receiver deck.

Figure 12 Receiver Deck, Lowering




Step 8. Replace the Lower Front Panel Overlay (Option 014 Only)

NOTE The new parts referenced in this procedure are listed in Table 1 on page 3.

Refer to Figure 13 for this procedure.

1. From the back side of the front panel, use a blunt object in one of the cutouts in the frame to push the overlay (item ①) and separate it from the front panel.
2. From the front side of the front panel, pull the overlay completely off and discard it.
3. Remove any adhesive remaining on the front panel.
4. Remove the protective backing from the new Option 014 front panel overlay (item ①).
5. Starting from either the left or right side, *loosely* place the overlay in the recess on the lower front panel, ensuring that it fits tightly against the recess edges.
6. Once the overlay is in place, press it firmly onto the frame to secure it.

Figure 13 Lower Front Panel Overlay Replacement



dt510a

NOTE The figure above shows the E8362B front panel and floppy disk drive. The E8362C front panel has a slightly different appearance and does not include a floppy disk drive.


Step 9. Reinstall the Front Panel Assembly and Front Panel Jumpers (Option 014 Only)

CAUTION Before installing the front panel assembly onto the analyzer, lift and support the front of the analyzer chassis.

Refer to Figure 14 for this procedure.

1. Tighten all 12 of the front-panel feed-through connectors using a 5/16-inch torque wrench set to 21-in lbs.
2. Reconnect the ribbon cable (item ①) to the A3 front panel interface board.
3. Slide the front panel over the test port connectors being careful to align the power switch and floppy disk drive to their corresponding front panel cutouts. Ensure that the ribbon cable ① is located below the fan to prevent it from being damaged by the fan blades.
4. With a T-10 TORX driver, install the eight screws (item ②) in the sides of the frame.
5. Install the six semirigid jumpers (item ③) on the front panel and tighten to 10-in lbs.

Figure 14 Front Panel Assembly Reinstallation



dt511a

NOTE	The previous figure shows the E8362B front panel and floppy disk drive. The E8362C front panel has a slightly different appearance and does not include a floppy disk drive.
-------------	--


Step 10. Reinstall the Outer Cover

CAUTION	This procedure is best performed with the analyzer resting on its front handles in the vertical position. <i>Do not place the analyzer on its front panel without the handles.</i> This will damage the front panel assemblies.
----------------	---

Refer to Figure 15 for this procedure.

1. Slide the cover over the analyzer frame.
2. With a T-20 TORX driver, install the four rear panel feet (item ①) by installing the center screws (item ②).
3. Slide the four bottom feet (item ③) into position on the cover.
4. With a T-20 TORX driver, install the strap handles (item ④) by installing the screws (item ⑤) on both ends of the handle.

Figure 15 Outer Cover Reinstallation



dt512a

NOTE	The previous figure shows the E8362B front panel and floppy disk drive. The E8362C front panel has a slightly different appearance and does not include a floppy disk drive.
-------------	--

Step 11. Enable Option 016

Procedure Requirements

- The analyzer must be powered up and operating to perform this procedure.
- The Network Analyzer program must be running.
- A mouse is required.

Enable Option 016

1. On the analyzer's **System** menu, point to **Service**, and then click **Option Enable**.
2. In the **Select Desired Option** list, click **016 - Receiver Attenuators**.
3. Click **Enable**.
4. Click **Yes** in answer to the displayed question in the **Restart Analyzer?** box.
5. When the installation is complete, click **Exit**.

Verify that Option 016 is Enabled

1. On the analyzer's **Help** menu, click **About Network Analyzer**.
2. Verify that "016" is listed after "Options:" in the display.
3. Click **OK** when done.

NOTE	If Option 016 has not been enabled, perform "Enable Option 016" again. If the option is still not enabled, contact Agilent Technologies. Refer to "Contacting Agilent" on page 2.
-------------	---

Step 12. Perform Post-Upgrade Adjustments and Calibration

Adjustments

The following adjustments must be made due to the hardware changes of the analyzer.

- source calibration
- receiver calibration

These adjustments are described in the PNA service guide and in the PNA on-line HELP. A list of equipment required to perform these adjustments can be found at “Equipment Required for Post-Upgrade Adjustments” on page 5.

Calibration

The analyzer should now operate and phase lock over its entire frequency range.

If you experience difficulty with the basic functioning of the analyzer, contact Agilent. Refer to “Contacting Agilent” on page 2.

- Although the analyzer functions, its performance relative to its specifications has not been verified.
- It is recommended that a full instrument calibration be performed using the PNA performance test software.
- Refer to the analyzer’s service guide for information on the performance test software.

