



N1027A-76A Electrical Clock Recovery Phase Matching Kit

CAUTION

Always use the 8 lb-in torque wrench when making electrical connections.

For information on *using* the electrical clock recovery phase matching kit, refer to the N1076A user's guide, 86100D FlexDCA's online help, or N1010A FlexDCA's online help.





Kit Parts

ID	Part	Qty	Part Number
1	Coaxial cable, 2.92 mm (m) to 2.92 mm (m), 21.5 in. length	2	N1075-60015
2	Coaxial cable, 2.92 mm (m) to 2.92 mm (m), 1 m length	2	N1075-60015
3	Delay-line assembly, 18.2 ns	1	N1027-60015
4	Microwave pick-off tee, 2.4 mm connectors, matched pair	1	N1027-2P2
5	Microwave equalizer, 0.5 to 18 GHz, 6 dB negative slope, SMA	2	N1027A-EQ6
6	Microwave equalizer, 0.5 to 18 GHz, 9 dB negative slope, SMA	2	N1027A-EQ9
7	T-20 Torx screwdriver	1	8710-1615
8	T-10 Torx screwdriver	1	8710-1623
9	Torque wrench, 8 lb-in, 5/16 Inch	1	8710-1765

This information subject to change without notice.
 © Keysight Technologies 2016
 Edition 1 September 2016
 Printed in Malaysia



N1076-90002
www.keysight.com