

Keysight N2125A/6A/7A and N2136A Calibration Modules

Notices

© Keysight Technologies 2019, 2021

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

N2125-97001

Edition

Second edition, October 2021

Available in electronic format only

Published by:

Keysight Technologies, Inc.
1900 Garden of the Gods Road
Colorado Springs, CO 80907 USA

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Keysight disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Keysight shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Keysight and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

U.S. Government Rights

The Software is “commercial computer software,” as defined by Federal Acquisition Regulation (“FAR”) 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement (“DFARS”) 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at

<http://www.keysight.com/find/sweula>.

The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFAR 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

Contents

1 Calibration Module Overview

- Calibration Module Models / 6
- What Is a Calibration? / 9
- When to Perform a Calibration / 9
- Inspecting the Package Contents / 10
- Safety and Regulatory Information / 11
- Cleaning and Handling the Calibration Module / 12

2 Using the Calibration Modules

- Calibrating an Infiniium UXR Oscilloscope / 14
 - Additional Equipment Required / 14
 - Calibration Module Components / 15
 - Applying Torque / 16
 - Calibration Time / 17
 - Calibration Procedure / 17
- Calibrating an M8131A Digitizer / 19
 - Calibration Time / 19
 - Calibration Procedure / 19

1 Calibration Module Overview

Calibration Module Models / 6

What Is a Calibration? / 9

When to Perform a Calibration / 9

Inspecting the Package Contents / 10

Safety and Regulatory Information / 11

Cleaning and Handling the Calibration Module / 12

This chapter introduces the four calibration modules used to calibrate the Keysight Infiniium UXR real-time oscilloscopes and the M8131A digitizer module.

Calibration Module Models

The N2125A, N2126A, and N2127A calibration modules are used to calibrate Infiniium UXR oscilloscopes.

N2125A 1.0 mm	N2126A 1.85 mm	N2127A 3.5 mm
UXR1104A, UXR1102A	UXR0704A, UXR0702A	UXR0334A
UXR1004A, UXR1002A	UXR0594A, UXR0592A	UXR0254A
UXR0804A, UXR0802A	UXR0504A, UXR0502A	UXR0204A
UXR0704AP, UXR0702AP	UXR0404A, UXR0402A	UXR0164A
UXR0594AP, UXR0592AP		UXR0134A
UXR0404AP, UXR0402AP		
UXR0254AP, UXR0252AP		

The N2136A calibration module is used to calibrate M8131A digitizers. The M8131A digitizers can be ordered in following configurations:

Description	Product Number
1 channel, 4 GSa/s, 1.6 GHz bandwidth, single ended	M8131A-011*
2 channel, 4 GSa/s, 1.6 GHz bandwidth, single ended	M8131A-012*
4 channel, 4 GSa/s, 1.6 GHz bandwidth, single ended	M8131A-014*
1 channel, 16 GSa/s, 6.5 GHz bandwidth, single ended	M8131A-061*
2 channel, 16 GSa/s, 6.5 GHz bandwidth, single ended	M8131A-062*
4 channel, 16 GSa/s, 6.5 GHz bandwidth, single ended	M8131A-064*
1 channel, 32 GSa/s, 12.5 GHz bandwidth, single ended	M8131A-131*
2 channel, 32 GSa/s, 12.5 GHz bandwidth, single ended	M8131A-132*
1 channel, 16 GSa/s, differential input	M8131A-FD1**
2 channel, 16 GSa/s or 1 channel, 32 GSa/s, diff. input	M8131A-FD2**
4 channel, 16 GSa/s or 2 channel, 32 GSa/s, diff. input	M8131A-FD4**

* Options -011,-012,-014, -061,-062,-064,-131,-132 require the N2136A for calibration.
 ** Options -FD1, FD2, FD4 do NOT require the N2136A for calibration.



Figure 1 N2125A 1.0 mm Calibration Module



Figure 2 N2126A 1.85 mm Calibration Module



Figure 3 N2127A 3.5 mm Calibration Module




Figure 4 N2136A 2.92 mm Calibration Module

What Is a Calibration?

Calibration of an oscilloscope or digitizer is simply a self-adjustment. The purpose of running a calibration is to reach performance optimization. Running a calibration performs a complete calibration of the entire oscilloscope or digitizer, including all channels installed in it, so it can be referenced back to NIST-characterized standards.

Calibration of UXR oscilloscopes

The calibration of a UXR oscilloscope is referred to as a frame calibration. It is a rigorous, complete set of calcs. It includes interleave correction, time scale, and 50 Ω input resistance calibrations, among others. To perform a frame calibration, the N2125A, N2126A, or N2127A calibration module is required.

All of the oscilloscope's calibration factors are stored into flash memory (NAND) on the backplane board.

Calibration of M8131A digitizers

Calibration of M8131A digitizer modules is required to compensate certain hardware distortions as well as calibrate the DC levels.

When to Perform a Calibration

For UXR oscilloscopes, you need to perform a calibration:

- When it has been more than 1 year since the last calibration.
- When the oscilloscope's operating temperature (after a 30-minute warm-up period) is more than ± 5 °C different from that of the last calibration.

For M8131A digitizers, you need to perform a calibration:

- When it has been more than 1 year since the last calibration.
- When the digitizer's operating temperature (after a 15-minute warm-up period) is more than ± 10 °C different from that of the last calibration.
- When the digitizer's number of available channels gets increased by a new license/option.

Inspecting the Package Contents








Verify that you received the items listed below for your calibration module.

Description	N2125A	N2126A	N2127A	N2136A	Part /Model No.
Calibration module assembly, 1.0 mm	X				N2125A
Calibration module assembly, 1.85 mm		X			N2126A
Calibration module assembly, 3.5 mm			X		N2127A
Calibration module assembly, M8131A				X	N2136A
Adapter, 3.5 mm (f) to (f)	X	X	X	X	5061-5311
Calibration cable, coaxial 50-ohm	X	X	X		8121-3052
Torque wrench, 8 in-lbs, 5/16 inch				X	8710-1765
Functional Test Certificate	X	X	X	X	9320-6636
Finding Product Manual sheet	X	X	X	X	9320-6794
Keysight safety leaflet	X	X	X	X	9320-6797
China RoHS addendum	X	X	X	X	9320-6679
South Korean Class A EMC declaration	X	X	X	X	5061-7383

Safety and Regulatory Information

Familiarize yourself with safety markings and instructions before you use the calibration module. This product has been designed and tested in accordance with international standards.

These symbols may be used on your calibration module or its packaging.

Symbol	Description
	The RCM mark is a registered trademark of the Australian Communications and Media Authority.
	<p>The CE mark is a registered trademark of the European Community. This CE mark shows that the product complies with all the relevant European Legal Directives.</p> <p>ICES/NMB-001 indicates that this ISM device complies with the Industry Canadian Interference-Causing Equipment Standard (ICES-001). Cet appareil ISM est conforme a la norme NMB-001 du Canada.</p> <p>ISM GRP.1 Class A indicates that this is an Industrial Scientific and Medical Group 1 Class A product.</p>
	<p>The crossed out wheeled bin symbol indicates that separate collection for waste electric and electronic equipment (WEEE) is required, as obligated by the EU DIRECTIVE and other National legislation.</p> <p>Please refer to keysight.com/go/takeback to understand your Trade in options with Keysight in addition to product takeback instructions.</p>
	This symbol indicates the time period during which no hazardous or toxic substance elements are expected to leak or deteriorate during normal use. Forty years is the expected useful life of the product.
	<p>The symbol on all primary and secondary packaging indicates compliance to China standard GB 18455-2001.</p> <p>To return unwanted products, contact your local Keysight office.</p>
	South Korean Certification (KC) mark; includes the marking's identifier code.
	<p>This mark denotes compliance with the essential requirements of the following applicable UK regulations:</p> <ul style="list-style-type: none"> ▪ Electromagnetic Compatibility Regulations 2016 No. 1091 (as amended) ▪ Electrical Equipment (Safety) Regulations 2016 No. 1101 (as amended) ▪ The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 No. 3032 (as amended)

Cleaning and Handling the Calibration Module

WARNING

Keep isopropyl alcohol away from heat, sparks, and flame. Store in a tightly closed container. It is extremely flammable. In case of fire, use alcohol foam, dry chemical, or carbon dioxide; water may be ineffective.

Use isopropyl alcohol with adequate ventilation and avoid contact with eyes, skin, and clothing. It causes skin irritation, may cause eye damage, and is harmful if swallowed or inhaled. It may be harmful if absorbed through the skin. Wash thoroughly after handling.

In case of spill, soak up with sand or earth. Flush spill area with water.

Dispose of isopropyl alcohol in accordance with all applicable federal, state, and local environmental regulations.

Inspect the calibration module regularly to prevent any surface contamination. If the module requires cleaning:

- 1 Disconnect the calibration module from the oscilloscope.
- 2 Gently clean the external parts of the module with a soft cloth dampened with isopropyl alcohol, preferably 99.5%. Do not try to clean internal parts.

To keep the calibration module in top condition:

- Store the module in its protective case when not in use.
- Be careful not to stretch any cables when using the module.
- Use the wrench that came with the oscilloscope or calibration module to torque connections as instructed.
- Do not bend or twist the cables.

2 Using the Calibration Modules

Calibrating an Infiniium UXR Oscilloscope / 14

Calibrating an M8131A Digitizer / 19

Calibrating an Infiniium UXR Oscilloscope

The UXR calibration modules are hot-swappable. That is, they can be connected and disconnected while the UXR oscilloscope is powered.

Additional Equipment Required

In addition to the calibration module and the accessories that came with it, you need the following equipment to calibrate a UXR oscilloscope. The adapters were included with the UXR oscilloscope.

Description	Critical specifications	Recommended part number or product
Adapter: 3.5 mm (f) to (f)	No substitute	5061-5311
Adapter: 1.0 mm ruggedized (f) to 2.92 mm (f) for models with 1.0 mm connector types	No substitute	5067-1393
Adapter: 2.92 mm to 2.40 mm 50 Ω for models with 1.85 mm connector types	No substitute	1250-3782
GPS 10 MHz frequency reference	Output Frequency: 10 MHz, < ± 0.1 ppb with no aging	Fluke Electronics 910R GPS Rubidium Frequency Reference

Calibration Module Components

The N2125A and N2126A calibration modules consist of an AutoProbe pod, a remote head, a main cable, and an RF cable.




Figure 5 N2125A calibration module connected to UXR oscilloscope with 1.0 mm connector types




Figure 6 N2126A calibration module connected to UXR oscilloscope with 1.85 mm connector types

The N2127A calibration module is a single unit.




Figure 7 N2127A calibration module connected to UXR oscilloscope with 3.5 mm connector types

Applying Torque

CAUTION

Over-tightening cables can damage the input connectors on the UXR oscilloscope.

For UXR models with 1.85 mm connector types, use the 5/16-inch torque wrench to tighten the connections to 8-in-lbs. Use the dual-ended 6 mm and 7 mm wrench to hold the mating connections in place while turning the connector.

For UXR oscilloscope models with 1.0 mm connector types, use the dual-ended torque wrench. One end is calibrated to 4 in-lbs for use when mating to standard 1.0 mm connectors; the other end is calibrated to 10 in-lbs for use only when mating to Keysight ruggedized 1.0 mm connectors. The N2125A has a ruggedized connector, so be sure to use the ruggedized end of the wrench. Follow the usage instructions in the *Keysight Infiniium UXR Real-Time Oscilloscopes User's Guide*.

Calibration Time

It takes about 1-1/2 hours to calibrate the first channel, and about an hour per each additional channel, including the time required to change cables from channel to channel.

Calibration Procedure


- 1 Remove the connector savers from all channels.
- 2 Connect the calibration module to the oscilloscope. For N2125A and N2126A modules, connect the components in this order:
 - a Attach the AutoProbe pod to the AutoProbe 3 interface for channel 1.
 - b Connect the RF cable to Aux Out.
 - c Attach the remote head to channel 1, using the wrench to apply proper torque.
- 3 Let the oscilloscope and calibration module warm up before running the cal.

NOTE

Both the calibration module and the oscilloscope must be warmed up (with the Infiniium application running) for at least 30 minutes at ambient temperature before starting the calibration procedure. Failure to allow warm up may result in inaccurate calibration.

- 4 Click **Utilities > Calibration...**
- 5 Clear the **Cal Memory Protect** check box. You cannot run a calibration if this box is selected.

Clear this check box before beginning calibration



- 6 To run a calibration without a time scale cal, select only the **Calibration** check box. If you also want to run a time scale calibration, you will need additional equipment.

- 7 A dialog box appears, showing an estimated time to complete the calibration. Click **OK**.
- 8 Click **Start Frame Cal**, then follow the instructions on the screen.
 - You must use the appropriate calibration module for your oscilloscope, and use the appropriate calibration cable with adapters when prompted by the software. Failure to use the appropriate parts will result in an inaccurate calibration.
 - As you move the calibration module from one channel to the next, first disconnect the remote head and the AutoProbe pod from the channel just calibrated. When connecting the calibration module to the next channel, connect the AutoProbe pod first, then connect the remote head.
 - A Passed/Failed indication appears for each calibration section.
 - If any section has failed, wait until the calibration is complete and then select the **Enable Details** check box for information on the failures. Also check the calibration cables.
- 9 When the calibration procedure is complete, click **Close**.

Calibrating an M8131A Digitizer

The N2136A calibration module is hot-swappable. That is, it can be connected and disconnected while the M8131A digitizer is powered.

Calibration Time

The time it takes to calibrate the M8131A digitizer module depends on the model:

- 4 GSa/s digitizer with single-ended input (options -011, -012, -014) takes about 20 minutes per channel.
- 16 GSa/s digitizer with single-ended input (options -061, -062, -064) takes about 20 minutes per channel.
- 32 GSa/s digitizer with single-ended input (options -131, -132) takes about 40 minutes per channel.


NOTE

The M8131A 16/32 GSa/s digitizer with differential input (options -FD1, -FD2, -FD4) does not require a calibration module. Refer to the *M8131A User's Guide* for calibration information.

Calibration Procedure

- 1 Disconnect all cables from the M8131A digitizer.
- 2 Let the digitizer warmed up for 15 minutes before running the calibration.

- 3 Open the Calibration dialog box (**Utilities > Calibration**).




- 4 Click **Start Calibration**, then follow the instructions on the screen.

The Process Result field will indicate that the calibration has been started. You can click **Abort** at any time. If the calibration module is not properly connected, a dialog box will appear, requesting you to verify the connection.

- 5 When calibration is complete for the first channel, follow the on-screen instructions for calibrating the remaining channels.

If any calibration fails, the calibration terminates with the following message (in Calibration Logs):



- 6 When all channels have completed calibration, click **Save** to save the calibration data to a text file.

2 Using the Calibration Modules



This information is subject to change without notice.

© Keysight Technologies, Inc. 2019, 2021

Second edition, October 2021

N2125-97001

www.keysight.com