


**Agilent EPM Series Power Meters:  
E4418A/B and E4419A/B**

**Installation Note**


**Instructions for installing  
the latest power meter firmware**



**Agilent Technologies**

Document Part Number E4418-90043 Printed in the UK August 2001	
---	--



Firmware Upgrade Kit Part Number E4418-61035	
---	--



Notice.

The information contained in this document is subject to change without notice.

Agilent Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Windows 95 and Windows NT are U.S. registered trademarks of Microsoft Corporation.

## Agilent EPM Series Power Meters: E4418A/B and E4419A/B

Product Affected: . . . . .	Agilent EPM Series Power Meters: E4418A/B and E4419A/B.
Serial Numbers: . . . . .	All
Options: . . . . .	All
To Be Performed By: . . . . .	(X) Agilent Service Center (X) Personnel Qualified by Agilent (X) Customer
Estimated Installation Time: . . . .	10 minutes via GPIB, 30 minutes via RS-232
Estimated Verification Time: . . .	1 minute

### A. Introduction

Agilent EPM power meter firmware is upgraded by downloading new firmware files from disc via computer GPIB or RS-232 interface. This requires the upgraded firmware files and the *Agilent EPM and EPM-P Dload program* for Windows 95 and NT. Both are supplied with firmware upgrade kit E4418-61035 as shown below in Table 1. They can also be obtained from the internet—for further information refer to section B. “Obtaining The Firmware And Download Program From The Internet,” on page 5.

This installation note details how to perform a firmware upgrade.

#### A.1 Firmware Upgrade Kit E4418-61035 Parts List

**Table 1**                      **Parts List**

Item	Quantity	Description	Part Number
1	1	Set of 3 discs containing the firmware files and download program.	E4418-61037
2	1	Installation Note (this document).	E4418-90043

## Agilent EPM Series Power Meters: E4418A/B and E4419A/B

### A.2 Tools Required

**Table 2** Tools Required for GPIB or RS-232 Connection

Quantity	Description	Part Number
	None	

### A.3 Hardware Required

**Table 3** Equipment Required for GPIB Connection

Quantity	Description	Part Number
1	PC with a GPIB interface	
1	GPIB cable	10833B

**Table 4** Equipment Required for RS-232 Connection

Quantity	Description	Part Number
1	PC with an RS-232 interface	
1	Serial RS-232 cable 9-pin (female) to 9-pin (female)	5182-4794

### A.4 Software Required

- The PC must be running either Windows 95<sup>®</sup> or Windows NT<sup>®</sup> 4.0.
- The PC must have the VISA library installed. (Typically installed during the GPIB card driver installation)

## B. Obtaining The Firmware And Download Program From The Internet

---

**NOTE:** If you have purchased Firmware Upgrade Kit E4418-61035 ignore this section and go straight to C. "Installing The Download Program" on page 6.

---

Make sure that you have three blank formatted discs available before beginning the following steps.

1. Access the following URL:  
`http://www.agilent.com/find/powermeters`
2. Select the EPM Series Power Meter Firmware Upgrade link, under the Additional Information section.
3. Download the download program:
  - Place a blank formatted disc in drive A.
  - Select the Setupload.exe link at the right of the page.
  - When the download dialog box appears, select drive A.  
The file setupload.exe is displayed in the File name window.
  - Click the **Save** button to start the file transfer.  
The Saving Location window opens showing the file transfer progress. When file transfer is complete the download dialog box closes.
  - Remove the floppy disc and label it appropriately.
4. Download the firmware:
  - Place a blank formatted disc in drive A.
  - For E4418A/B firmware, select the e4418\_a1nnnn.exe link. For E4419A/B firmware, select the e4419\_a2nnnn.exe link.

---

**NOTE:** The value of nnnn changes with each firmware release.

E4418A/B firmware must be used in E4418A/B single-channel power meters only. All E4418A/B firmware filenames begin with e4418\_a1.

E4419A/B firmware must be used in E4419A/B dual-channel power meters only. All E4419A/B firmware filenames begin with e4419\_a2.

---

- When the download dialog box appears, select drive A.
  - Click the **Save** button to start the file transfer.  
The Saving Location window appears, showing the file transfer progress. When file transfer is complete the download dialog box closes.
  - Remove the floppy disc and label it appropriately.
5. Decompress the downloaded firmware files:
    - Ensure that the firmware disc is in drive A.
    - From the Windows task bar select **Start, Run...** and click on the **Browse...** button.
    - Select the required firmware file. Click on **OK**.
    - When the WinZip Self-Extractor window opens, select **Unzip**.
    - Close the WinZip Self-Extractor window when complete.

## C. Installing The Download Program

This procedure shows how to install the *Agilent EPM and EPM-P Dload program* for Windows 95 and NT on a PC running Windows 95<sup>®</sup> or Windows NT<sup>®</sup> 4.0.

1. Insert the download program disc into drive A.
2. From the Windows task bar select **Start, Run...** and enter `A:\setupdload.exe`
3. Click on **OK**.
4. To verify the installation, select **Start, Programs, Agilent EPM and EPM-P Dload** from the Windows task bar. Check that the program option **Agilent EPM and EPM-P Dload** is displayed as shown below:



## D. Instrument Self Test

Use this procedure to check that the power meter's components are functioning correctly.

1. Connect the power cord to the power meter and press the power switch. The front panel display and the green power LED light up when the power meter is switched on. The instrument performs a power on Self Test. If the self test is not successful the ERR (error) annunciator turns on.
2. Perform a full Instrument Self Test by pressing **System**, **More**, **Service**, **Self Test**, **Instrument Self Test**. The power meter now carries out a full functional test.
3. If the ERR annunciator is turned on, review the error message queue by pressing **System**, **Error List** and use **Next** to scroll through the error messages.

## E. Upgrading Firmware

This procedure shows how to download new firmware into the power meter using a PC running Windows 95<sup>®</sup> or Windows NT<sup>®</sup> 4.0.

1. Connect the computer to the power meter:
  - If using GPIB, connect the computer to the power meter via the GPIB cable.
  - If using RS-232, connect the computer to the power meter via the RS-232 cable.

---

**NOTE:** The COM2 RS-232 port is normally used but COM1 may be used if it is available. No other programs can make use of this port at any time during the procedure

---

2. Ensure that the power meter is turned on.
3. Insert the appropriate firmware disc into drive A.
4. Run *Agilent EPM Dload for Windows 95,NT*:
  - Select **Start, Programs, Agilent EPM and EPM-P Dload** from the Windows task bar
  - Select **Agilent EPM and EPM-P Dload**
5. Ensure that the Recover from failed download box is *not* checked.
6. In the Temporary file box, enter a location where the download program can store data (for example, calibration tables) while upgrading your power meter's firmware. Location C:\info.tmp is acceptable for most computers. This file is used for recovery purposes if the upgrade fails for any reason. It is deleted automatically after a successful upgrade. If the upgrade fails, the file is left to enable recovery as described in Section F. "Recovering From a Failed Upgrade," on page 8.
7. Select the VISA address of your power meter. Either:
  - a Click on the down arrow beside the Instrument addr box.  
After approximately 15 seconds you are presented with a list of all EPM series power meters connected to your computer. Select the meter to be upgraded.

or

  - b Otherwise, enter your power meter's VISA address into the Instrument addr box.
    - For GPIB, this is in the form:  
GPIB<board>::<address>::INSTR  
e.g. GPIB0::13::INSTR
    - For RS-232 this is in the form:  
ASRL<board>::INSTR  
e.g. ASRL2::INSTR
8. Select the firmware file to be used:
  - Click on the **Browse** button next to the Download file box.
  - Click on the Look in box and select 3<sup>1</sup>/<sub>2</sub> Floppy.
  - Select the appropriate firmware file.
  - Click on Open.

## Agilent EPM Series Power Meters: E4418A/B and E4419A/B

---

**NOTE:** E4418A/B firmware must be used in E4418A/B single-channel power meters only. All E4418A/B firmware filenames begin with e4418\_a1.

E4419A/B firmware must be used in E4419A/B dual-channel power meters only. All E4419A/B firmware filenames begin with e4419\_a2.

---

9. Download the firmware:

- Click on the Download button.
- Follow the instructions given on the screen.

10. Verify that the new firmware has been downloaded. On the power meter press , , **Service**, **Version**. The new firmware revision and firmware date is displayed.

11. To check that the power meter is operating correctly, repeat D. "Instrument Self Test," on page 6.

### F. Recovering From a Failed Upgrade

It is unlikely that your upgrade will fail. However, if the upgrade is not successful, two things may happen:

- Your power meter may stop working.
- You may lose calibration tables and other data.

To return the power meter to normal operation:

1. Follow steps 1 to 5 of E. "Upgrading Firmware," on page 7.
2. Ensure that the Recover from failed download box *is* checked.
3. Ensure that the Temporary file box contains the location where you stored the temporary file.
4. Follow E. "Upgrading Firmware," on page 7 instructions from step 8.  
Note that if the power meter is not functioning its VISA address cannot be detected automatically and must be entered manually in the Instrument addr box.

---

**NOTE:** The temporary file is *not* over-written if the Recover from failed download box *is* checked. It is deleted afterwards if the upgrade is successful. If the upgrade is unsuccessful then the file is unaltered, enabling another attempt to be made.

---



## **G. Troubleshooting Hints**

### ***G.1 “dloadwin32.exe - Unable to locate DLL***

***The dynamic link library VISA32.DLL could not be found in the specified path...”***

Ensure that VISA is installed correctly on your computer.

### ***G.2 I clicked on the ‘down’ arrow beside Instrument addr, and my power meter was not on the list.***

Ensure that your power meter is connected to the PC. The VISA library must be set up for the GPIB interface card. Also ensure that the power meter’s programming language is set to SCPI (see the User’s Guide for how to do this).

### ***G.3 “Could not open VISA session with instrument at <address>”***

Ensure that you have entered the correct address for your power meter. The VISA library must be set up for the GPIB interface card.

### ***G.4 “Instrument returns identity <ID>***

***This download program can only be used with EPM power meters”***

Ensure that the address entered is correct, and that there are no other instruments connected using this address.

### ***G.5 “File <filename> contains errors”***

The firmware file which you have selected is not a valid dload file, or does not exist. Check that you have entered the correct path

## Agilent EPM Series Power Meters: E4418A/B and E4419A/B

### **Agilent Technologies Software License Agreement.**

ATTENTION: USE OF THE SOFTWARE IS SUBJECT TO THE AGILENT SOFTWARE LICENSE TERMS SET FORTH BELOW. USING THE SOFTWARE INDICATES YOUR ACCEPTANCE OF THESE LICENSE TERMS. IF YOU DO NOT ACCEPT THESE LICENSE TERMS, YOU MUST RETURN THE SOFTWARE FOR A FULL REFUND. IF THE SOFTWARE IS SUPPLIED WITH ANOTHER PRODUCT, YOU MAY RETURN THE ENTIRE UNUSED PRODUCT FOR A FULL REFUND.

### **AGILENT SOFTWARE LICENSE TERMS.**

The following terms govern your use of the enclosed Software unless you have a separate written agreement with Agilent.

#### **License Grant.**

Agilent grants you a license to Use one copy of the Software. "Use" means storing, loading, installing, executing or displaying the Software. You may not modify the Software or disable any licensing or control features of the Software. If the Software is licensed for "concurrent use", you may not allow more than the maximum number of authorized users to Use the Software concurrently.

#### **Ownership.**

The Software is owned and copyrighted by Agilent or its third party suppliers. Your license confers no title or ownership in the Software and is not a sale of any rights in the Software. Agilent's third party suppliers may protect their rights in the event of any violation of these License Terms.

#### **Copies and Adaptations.**

You may only make copies or adaptations of the Software for archival purposes or when copying or adaptation is an essential step in the authorized Use of the Software. You must reproduce all copyright notices in the original Software on all copies or adaptations. You may not copy the Software onto any bulletin board or similar system.

#### **No Disassembly or Decryption.**

You may not disassemble or decompile the Software unless Agilent's prior written consent is obtained. In some jurisdictions, Agilent's consent may not be required for disassembly or recompilation. Upon request, you will provide Agilent with reasonably detailed information regarding any disassembly or decompilation. You may not decrypt the Software unless decryption is a necessary part of the operation of the Software.

#### **Transfer**

Your license will automatically terminate upon any transfer of the Software. Upon transfer, you must deliver the Software, including any copies and related documentation, to the transferee. The transferee must accept these License Terms as a condition to the transfer.

**Termination.**

Agilent may terminate your license upon notice for failure to comply with any of these License Terms. Upon termination, you must immediately destroy the Software, together with all copies, adaptations and merged portions in any form.

**Export Requirements.**

You may not export or re-export the Software or any copy or adaptation in violation of any applicable laws or regulations.

**U.S. Government Restricted Rights.**

The Software and documentation have been developed entirely at private expense and are provided as “Commercial Computer Software” or “restricted computer software”. They are delivered and licensed as “commercial computer software” as defined in DFARS 252.227-7013 (Oct. 1988), DFARS 252.211-7015 (May 1991) or DFARS 252.227-7014 (Jun. 1995), as a “commercial item” as defined in FAR 2.101 (a), or as “Restricted computer software” as defined in FAR 52.227-19 (Jun. 1987) (or any equivalent agency regulation or contract clause), whichever is applicable. You have only those rights provided for such Software and Documentation by the applicable FAR or DFARS clause or the Agilent standard software agreement for the product.

## Agilent EPM Series Power Meters: E4418A/B and E4419A/B