

Keysight D9010XAUC XAUI Compliance Application

Notices

© Keysight Technologies, Inc. 2007-2019

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Revision

Version 01.40.0000

Edition

June 4, 2019

Available in electronic format only

Published by:

Keysight Technologies, Inc.
1900 Garden of the Gods Road
Colorado Springs, CO 80907 USA

Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Keysight disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Keysight shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Keysight and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology License

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at www.keysight.com/find/sweula. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFAR 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

In This Book

This book is your guide to programming the Agilent Technologies D9010XAUC XAUI Compliance Application.

- **Chapter 1**, “Introduction to Programming,” starting on page 7, describes compliance application programming basics.
- **Chapter 2**, “Configuration Variables and Values,” starting on page 9, **Chapter 3**, “Test Names and IDs,” starting on page 13, and **Chapter 4**, “Instruments,” starting on page 17 provide information specific to programming the D9010XAUC XAUI Compliance Application.

How to Use This Book

Programmers who are new to compliance application programming should read all of the chapters in order. Programmers who are already familiar with this may review chapters 2, 3, and 4 for changes.

Contents

In This Book / 3

1 Introduction to Programming

Remote Programming Toolkit / 8

2 Configuration Variables and Values

3 Test Names and IDs

4 Instruments

Index

1 Introduction to Programming

Remote Programming Toolkit / 8

This chapter introduces the basics for remote programming a compliance application. The programming commands provide the means of remote control. Basic operations that you can do remotely with a computer and a compliance app running on an oscilloscope include:

- Launching and closing the application.
- Configuring the options.
- Running tests.
- Getting results.
- Controlling when and where dialogs get displayed
- Saving and loading projects.

You can accomplish other tasks by combining these functions.

Remote Programming Toolkit

The majority of remote interface features are common across all the Keysight Technologies, Inc. family of compliance applications. Information on those features is provided in the N5452A Compliance Application Remote Programming Toolkit available for download from Keysight here: www.keysight.com/find/rpi. The D9010XAUC XAUI Compliance Application uses Remote Interface Revision 6.00. The help files provided with the toolkit indicate which features are supported in this version.

In the toolkit, various documents refer to "application-specific configuration variables, test information, and instrument information". These are provided in Chapters 2, 3, and 4 of this document, and are also available directly from the application's user interface when the remote interface is enabled (View>Preferences::Remote tab::Show remote interface hints). See the toolkit for more information.

2 Configuration Variables and Values

The following table contains a description of each of the D9010XAUC XAUI Compliance Application options that you may query or set remotely using the appropriate remote interface method. The columns contain this information:

- GUI Location – Describes which graphical user interface tab contains the control used to change the value.
- Label – Describes which graphical user interface control is used to change the value.
- Variable – The name to use with the SetConfig method.
- Values – The values to use with the SetConfig method.
- Description – The purpose or function of the variable.

For example, if the graphical user interface contains this control on the **Set Up** tab:

- Enable Advanced Features

then you would expect to see something like this in the table below:

Table 1 Example Configuration Variables and Values

GUI Location	Label	Variable	Values	Description
Set Up	Enable Advanced Features	EnableAdvanced	True, False	Enables a set of optional features.

and you would set the variable remotely using:

ARSL syntax

```
arsl -a ipaddress -c "SetConfig 'EnableAdvanced' 'True'"
```

C# syntax

```
-----
remoteAte.SetConfig("EnableAdvanced", "True");
```

Here are the actual configuration variables and values used by this application:

NOTE

Some of the values presented in the table below may not be available in certain configurations. Always perform a "test run" of your remote script using the application's graphical user interface to ensure the combinations of values in your program are valid.

NOTE

The file, "ConfigInfo.txt", which may be found in the same directory as this help file, contains all of the information found in the table below in a format suitable for parsing.

Table 2 Configuration Variables and Values

GUI Location	Label	Variable	Values	Description
Configure	Amplitude Test Mode	AmpTestMode	AVERAGE, WORSTCASE	Select the test pattern to use for all tests.
Configure	Bit Error Ratio Level, BER	BER	E6, E7, E8, E9, E10, E11, E12, E13, E14, E15, E16, E17, E18	Select the number of BER for jitter tests.
Configure	D+ Channel	DPChan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for D+.
Configure	D- Channel	DNChan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for D-.
Configure	Data Lane	DataLaneS	Lane 0, Lane 1, Lane 2, Lane 3	Select data lane under test.
Configure	Lagging ISI Filter Coefficients	LagCo	0, 1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18	Select the number of lagging ISI Filter coefficients for jitter tests.
Configure	Lane 0 Channel	Lane0Chan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for Lane 0.
Configure	Lane 1 Channel	Lane1Chan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for Lane 1.
Configure	Lane 2 Channel	Lane2Chan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for Lane 2.
Configure	Lane 3 Channel	Lane3Chan	CHAN1, CHAN2, CHAN3, CHAN4	Select the channel number for Lane 3.
Configure	Leading ISI Filter Coefficients	LeadCo	0, -1, -2, -3, -4, -5, -6, -7, -8, -9, -10, -11	Select the number of leading ISI Filter coefficients for jitter tests.

Table 2 Configuration Variables and Values (continued)

GUI Location	Label	Variable	Values	Description
Configure	Mask Test Averaging	Avg_Mask	0, 2, 4, 8, 16, 32, 64, 128, 256	Select the averaging number for Mast Test.
Configure	Measurement Thresholds	Threshold	thrd1, thrd2	Select the measurement threshold level for transition time and Driver Output Amplitude Test (Far End).
Configure	Memory Depth	MemDepth	262144, 524288, 320000, 640000, 1025000, 1280000, 1920000, 2000000, 4000000, 5000000, 8000000, 419430, 838861, 512000, 1024000, 1640000, 2048000, 3072000, 3200000, 6400000, 8000000, 12800000	Please select memory depth of the oscilloscope to acquire data.
Configure	Number of UI	NoOfUI	312500, 318750, 375000, 500000, 1000000	Select the number of UI to capture data.
Configure	PLL Bandwidth	PLLbandwidth	1, 0	Select the PLL Bandwidth.
Configure	Report Test Time	TimeReport	true, false	Choose whether to report time-stamp and test time for each test.
Configure	Sampling Rate	SamplingRate	20E+09, 40E+09, 32E+09, 64E+09	Select the Sampling Rate to use.
Configure	Test Pattern	Pattern	55, CJPAT, HFTP, LFTP, MFTP, CRPAT	Select the test pattern in use for all Clause 47 and 54 tests.
Run Tests	Event	RunEvent	(None), Fail, Margin < N, Pass	Names of events that can be used with the StoreMode=Event or RunUntil RunEventAction options
Run Tests	RunEvent=Margin < N: Minimum required margin %	RunEvent_Margin < N_MinPercent	Any integer in range: 0 <= value <= 99	Specify N using the 'Minimum required margin %' control.

Table 2 Configuration Variables and Values (continued)

GUI Location	Label	Variable	Values	Description
Set Up	DeviceSpec	cmbConnectionType	Differential, Single-ended	
Set Up	DeviceType	DeviceType	XAUI, 10GBASE-CX4	DeviceType
Set Up	cbLane0	cbLane0	0.0, 1.0	cbLane0
Set Up	cbLane1	cbLane1	0.0, 1.0	cbLane1
Set Up	cbLane2	cbLane2	0.0, 1.0	cbLane2
Set Up	cbLane3	cbLane3	0.0, 1.0	cbLane3
Set Up	cmbBitRate	cmbBitRate	10 GbE XAUI 3.125Gbps, 10 GFC XAUI 3.1875Gbps, XAUI Based 3.75Gbps, OBSAI RP3 768Mbps, OBSAI RP3 1.536Gbps, OBSAI RP3 3.072Gbps, SRIO Short 1.25Gbps, SRIO Short 2.5Gbps, SRIO Short 3.125Gbps, SRIO Long 1.25Gbps, SRIO Long 2.5Gbps, SRIO Long 3.125Gbps, CPRI-LV 614.4 Mbps, CPRI-LV 1.2288Gbps, CPRI-LV 2.4576Gbps, CPRI-LV 3.072Gbps, CPRI-HV 614.4 Mbps, CPRI-HV 1.2288Gbps, CPRI-HV 2.4576Gbps	cmbBitRate
Set Up	pcboOverallDeviceID	pcboOverallDeviceID		pcboOverallDeviceID
Set Up	txtOverallUserComment	txtOverallUserComment	(Accepts user-defined text)	txtOverallUserComment

3 Test Names and IDs

The following table shows the mapping between each test's numeric ID and name. The numeric ID is required by various remote interface methods.

- Name – The name of the test as it appears on the user interface **Select Tests** tab.
- Test ID – The number to use with the RunTests method.
- Description – The description of the test as it appears on the user interface **Select Tests** tab.

For example, if the graphical user interface displays this tree in the **Select Tests** tab:

- All Tests
 - Rise Time
 - Fall Time

then you would expect to see something like this in the table below:

Table 3 Example Test Names and IDs

Name	Test ID	Description
Fall Time	110	Measures clock fall time.
Rise Time	100	Measures clock rise time.

and you would run these tests remotely using:

```
ARSL syntax
-----
arsl -a ipaddress -c "SelectedTests '100,110'"
arsl -a ipaddress -c "Run"
```

```
C# syntax
-----
remoteAte.SelectedTests = new int[] {100,110};
remoteAte.Run();
```

Here are the actual Test names and IDs used by this application:

NOTE

The file, "TestInfo.txt", which may be found in the same directory as this help file, contains all of the information found in the table below in a format suitable for parsing.

Table 4 Test IDs and Names

Name	TestID	Description
Baud Rate Test	100	The baud rate of the device under test (DUT) must be within the conformance limits specified in 47.3.3 of IEEE 802.3.
CX4 Baud Rate Test	1000	The baud rate of the device under test (DUT) must be within the conformance limits specified in 54.6.3.3 of IEEE 802.3.
CX4 Fall Time Test	1500	The fall time of the device under test (DUT) must be within the conformance limits specified in 54.6.3.7 of IEEE 802.3.
CX4 Rise Time Test	1400	The rise time of the device under test (DUT) must be within the conformance limits specified in 54.6.3.7 of IEEE 802.3.
Deterministic Jitter Test	1700	The deterministic transmit jitter of the device under test (DUT) must be within the conformance limits specified in 54.6.3.8 of IEEE 802.3.
Deterministic Jitter Test (Far-end)	610	The DUT must conform to the jitter requirements specified in clause 47.3.3.5 of IEEE 802.3
Deterministic Jitter Test (Near-end)	600	The DUT must conform to the jitter requirements specified in clause 47.3.3.5 of IEEE 802.3
Differential Output Amplitude Test	1200	The differential output amplitude of the device under test (DUT) must be within the conformance limits specified in 54.6.3.4 of IEEE 802.3.
Differential Output Template Test	1900	The DUT's transmitter meets the eye template requirements specified in clause 54.3.3.5. of IEEE 802.3
Driver Eye Template Test (Far-end)	410	The DUT's transmitter meets the eye template requirements specified in clause 47.3.3.5. of IEEE 802.3 (require either Near-end or Far-end test only)
Driver Eye Template Test (Near-end)	400	The DUT's transmitter meets the eye template requirements specified in clause 47.3.3.5. of IEEE 802.3 (require either Near-end or Far-end test only)
Driver Output Amplitude Test (Far-end)	210	The driver differential output amplitude of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.
Driver Output Amplitude Test (Near-end)	200	The driver differential output amplitude of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.
Driver Single-ended Output Swing Maximum Absolute Test (Tx+)	301	The single-ended output swing of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.

Table 4 Test IDs and Names (continued)

Name	TestID	Description
Driver Single-ended Output Swing Maximum Absolute Test (Tx-)	303	The single-ended output swing of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.
Driver Single-ended Output Swing Minimum Absolute Test (Tx+)	302	The single-ended output swing of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.
Driver Single-ended Output Swing Minimum Absolute Test (Tx-)	304	The single-ended output swing of the device under test (DUT) must be within the conformance limits specified in clause 47.3.3.2 of IEEE 802.3.
Fall Time Test	951	The fall time of the device under test (DUT) must be within the conformance limits - value > 60ps.
Fall Time Test	950	The fall time of the device under test (DUT) must be within the conformance limits specified in 47.3.3.3 of IEEE 802.3.
Lane-to-Lane Amplitude Difference	1250	The differential output amplitude of the device under test (DUT) must be within the conformance limits specified in 54.6.3.4 of IEEE 802.3. It will report the worst case result.
Random Jitter Test	1600	The random transmit jitter of the device under test (DUT) must be within the conformance limits specified in 54.6.3.8 of IEEE 802.3.
Rise Time Test	901	The rise time of the device under test (DUT) must be within the conformance limits - value > 60ps.
Rise Time Test	900	The rise time of the device under test (DUT) must be within the conformance limits specified in 47.3.3.3 of IEEE 802.3.
Total Jitter Test	1800	The total transmit jitter of the device under test (DUT) must be within the conformance limits specified in 54.6.3.8 of IEEE 802.3.
Total Jitter Test (Far-end)	510	The DUT must conform to the jitter requirements specified in clause 47.3.3.5 of IEEE 802.3
Total Jitter Test (Near-end)	500	The DUT must conform to the jitter requirements specified in clause 47.3.3.5 of IEEE 802.3

3 Test Names and IDs

4 Instruments

The following table shows the instruments used by this application. The name is required by various remote interface methods.

- Instrument Name – The name to use as a parameter in remote interface commands.
- Description – The description of the instrument.

For example, if an application uses an oscilloscope and a pulse generator, then you would expect to see something like this in the table below:

Table 5 Example Instrument Information

Name	Description
scope	The primary oscilloscope.
Pulse	The pulse generator used for Gen 2 tests.

and you would be able to remotely control an instrument using:

ARSL syntax (replace [description] with actual parameter)

```
-----  
arsl -a ipaddress -c "SendScpiCommandCustom 'Command=[scpi  
command];Timeout=100;Instrument=pulsegen'"
```

```
arsl -a ipaddress -c "SendScpiQueryCustom 'Command=[scpi  
query];Timeout=100;Instrument=pulsegen'"
```

C# syntax (replace [description] with actual parameter)

```
-----  
SendScpiCommandOptions commandOptions = new SendScpiCommandOptions();  
commandOptions.Command = "[scpi command]";  
commandOptions.Instrument = "[instrument name]";  
commandOptions.Timeout = [timeout];  
remoteAte.SendScpiCommand(commandOptions);
```

```
SendScpiQueryOptions queryOptions = new SendScpiQueryOptions();  
queryOptions.Query = "[scpi query]";  
queryOptions.Instrument = "[instrument name]";
```

```
queryOptions.Timeout = [timeout];  
remoteAte.SendScpiQuery(queryOptions);
```

Here are the actual instrument names used by this application:

NOTE

The file, "InstrumentInfo.txt", which may be found in the same directory as this help file, contains all of the information found in the table below in a format suitable for parsing.

Table 6 Instrument Names

Instrument Name	Description
Infiniium	The primary oscilloscope.

Index

C

configuration variables and values, [9](#)
copyright, [2](#)

I

IDs and names of tests, [13](#)
instrument names, [17](#)

N

names and IDs of tests, [13](#)
names of instruments, [17](#)
notices, [2](#)

P

programming, introduction to, [7](#)

R

Remote Programming Toolkit, [8](#)

T

test names and IDs, [13](#)

V

variables and values, configuration, [9](#)

W

warranty, [2](#)

