

Instrument Software Update Instructions

Keysight X-Series Signal Analyzers, Non-Touch (N9000A, N9010A, N9020A, N9030A) With WinXP operating system

➔ Upgrading to A.12.09 and above ←

Purpose

The purpose of this document is to outline the instrument software update procedure for all non-touch X-Series Signal Analyzers (CXA, EXA, MXA, and PXA) running the Windows XP operating system from the revision that is currently installed, to the latest revision available. The advantages of updating the instrument software with the current revision include:

1. Keeps the analyzer current and up to date.
2. Updating the advanced measurement applications where appropriate.
3. Updating the PC board software where appropriate.

An update to the instrument software revision **does not** require a new license key for the measurement application(s), so long as the application(s) were licensed prior to the update. This procedure will not update the 89601A or 89601B VSA measurement software.

To determine what revision of instrument software is currently installed, press [**System**], {**Show**}, {**System**} on the instrument, look for the "Instrument S/W Revision" number (for example: "A.10.52"). If the installed revision is the same as the latest revision on the web, there is no need to update the Software.

Windows Service Pack 3

X-Series analyzers with Windows XP must have Windows Service Pack 3 installed before upgrading to the latest instrument software. These instructions include how to identify if your analyzer has Service Pack 3, and steps to download Service Pack 3 from the Microsoft® website and install it on your instrument if needed. All X-Series analyzers running Windows XP that have instrument software version A.08.00 or later will already have this Service Pack installed.

To determine if you need to download Service Pack 3: Go to **Start**→**My Computer** right click and go to Properties.

File Download

1. Go to http://www.keysight.com/find/Xseries_software
2. Select the appropriate link in the Software Upgrade section of the table for the x-Series analyzer that is being updated.
3. Select the Previous Versions tab to access the files for analyzers running the Windows XP OS.
4. Download the executable file by clicking on the Download button and then selecting the appropriate instrument model number. Save the file to a location of your choice.

If you do not have the ability to download files directly to the C: drive of the instrument being updated, store the file on a large capacity USB storage device for transfer to the instrument. (The file size is ~1 GByte.)

Update Procedure

The default user account is ***instrument***, which does not have the required permissions to install the instrument software updates. For the process outlined below, the user must be logged in as the ***administrator***.

The new X-Series automated instrument software upgrade process has an install wizard which removes the old software version and installs new software version without manual steps. The removal of the old software version and the installation of the new software version completes in approximately 40 minutes.

1. If the analyzer is not already running, turn the analyzer on and allow to completely boot up
2. Close the analyzer application by pressing [**File**], {**Exit**}
3. Connect a USB keyboard and mouse to the instrument.
4. Log out of the default user (***instrument***) by selecting **Start**→**Log Off**→**Log Off**.
5. At the login prompt enter:
 - User name: **administrator**
 - Default Password: **agilent4u**

Note: If a “Found New Hardware” wizard window pops up, select “No Not at this Time” to bypass the activity.

Service Pack 3 Requirement: – Instruments that are being updated from a version prior to A.08.00 (i.e. A.04.05) to instrument software version A.08.00 or later need to perform steps A-F below.

- A. Insert the USB storage device containing the Windows XP Service Pack (KB936929) file ("SP3 Install Package") into one of the available USB ports on the instrument.
- B. Navigate to the SP3 Install Package on the USB storage device. Click on the SP3 Install Package to run it.

- C. Follow the on-screen prompts to complete the installation. This installation will take between 30 and 45 minutes.
- D. When the installation has finished, select "Yes, I want to restart my computer now" and then select "Finish".
- E. After the instrument restarts, you will be asked to either, "*Help protect my PC by turning on Automatic Updates now*" Or "*Not right now*". Select "**Not right now**" and click "Next" in the lower-right hand corner.
- F. Click **Start**→**Log Off**→**Log Off**

At the log in prompt, enter:


- User Name: **administrator**
- Default Password: **agilent4u**

Instrument Software Install

If you downloaded the instrument software update file to the instrument via LAN, access that file from the location that it was stored and proceed to step 1 below.


If you downloaded the instrument software update file to a USB storage device, plug that into one of the USB ports on the front of the instrument. Access the file for the removable drive, then proceed to step 1.

1. Double click on the file named "**XSA_Installer_A.XX.XX._WinXP.exe**" for the version to which you want to upgrade.
2. A message appears, reminding you that you must be logged in as **administrator** and the Windows XP Service Pack 3 is required. Select Yes




3. A window appears showing the extraction of the software installer. When the extraction process is complete, there will be another window that will show the progress of the software removal and installation. There is no interaction required from this point forward.


Extraction window



If you are upgrading from very old instrument software versions, you may see a message such as the one below:



If you choose to uninstall the 89600A Application, you may see a message such as the following:



If you see this message, please refer to the following FAQ:


[What to do when, "Unwise.exe file missing," error message is given?](#)

Otherwise, you should see the Instrument Software update window:



The old instrument software is being removed. The "Uninstall in Progress" text will roll across the screen indicating software activity. There will be several windows appearing and disappearing. It will display "Completed Successfully" when this step is finished.


NO interaction with the instrument is required.




The new instrument software is being installed. The “Installation in Progress” text will roll across the screen indicating software activity. There will be several windows appearing and disappearing during this time. NO interaction with the instrument is required. Please be patient the total software installation process can take 35 to 40 minutes. It will display “Completed Successfully” when the software update is complete. You will be prompted to restart your instrument at the end.

Example of one of the windows you will see during the software update process:

FPGA Programming: DO NOT power off the instrument at this time for **ANY** reason. If this process is interrupted, the instrument may end up in an unusable state, and will likely need to be sent to Keysight for repair before it will be usable again.



Software upgrade “Completed Successfully”



4. When the Software installation is complete. You will be asked to reboot the instrument. If no action is taken the instrument will automatically reboot in 5 minutes.

Note: When you press “Finish” you will see this window message indicating that the system is shutting down. This is a normal Window warning message.



5. After the instrument restarts it will boot up in the default “instrument “account, perform the alignments and display the measurement screen.
6. Check for any errors and verify that the new software version is installed. Press [**System**], {**Show**}, {**System**} on the instrument, look for the “Instrument S/W Revision”

| Agilent IQ Analyzer (Basic) - Complex Spectrum | | |
|--|--------------------------------------|-----------------------------|
| RFIO1 | 50 Ω | DC |
| SENSE:INT | | RF OFF |
| ALIGN OFF | | |
| Center Freq 1.00000000 GHz | | |
| PNO: Wide | | Center Freq: 1.00000000 GHz |
| IFGain:Low | | Trig: Free Run |
| | | Avg Hold:>25/25 |
| Atten: 6 dB (Elec 0) | | |
| Agilent Technologies | | |
| Agilent EXT | Agilent EXT Test Set | |
| Product Number | E6607A | |
| Serial Number | MY49390107 | |
| Instrument S/W Revision | A.09.47 | |
| Computer Name | A-E6607A-90107 | |
| IP Address | 127.0.0.1 | |
| Host ID | E6607A,MY49390107 | |
| mDNS Enabled | Yes | |
| mDNS Host Name | A-E6607A-90107 | |
| mDNS Service Name | Agilent E6607A Test Set - MY49390107 | |

Note: If the instrument does not reboot, remove the USB storage device and cycle power. The boot order for the instrument may have been changed in the instrument BIOS previously, and the instrument may be instructed to boot from USB first, and the internal disk drive second. If this is the case, failure to remove the USB device will result in an error, as the instrument attempts to boot from it.