

Name: Calibration Laboratory of Keysight Technologies (China) Co., Ltd.

Address: No.3, Wangjing North Road, Chaoyang District, Beijing, China

Registration No. CNAS L0640

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS


Effective Date: 2024-01-18 Expiry Date: 2030-01-21



SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
未分组							
1	*RF/MW Vector Signal Generator	Power Level	RF/MW Vector Signal Generator Test Procedure CAL-CHA001-01	20 dBm~30 dBm , 9 kHz~14 GHz	U=0.16 dB		
				20 dBm~30 dBm ,14 GHz~18 GHz	U=0.17 dB		
				15 dBm~20 dBm ,9 kHz~100 kHz	U=0.10 dB		
				15 dBm~20 dBm ,100 kHz~8 GHz	U=0.053 dB		
				15 dBm~20 dBm ,8 GHz~24 GHz	U=0.081 dB		
				15 dBm~20 dBm ,24 GHz~67 GHz	U=0.12 dB		
				15 dBm~20 dBm ,9 kHz~100 kHz	U=0.097 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				10 dBm~15 dBm ,100 kHz~8 GHz	U=0.051 dB		
				10 dBm~15 dBm ,8 GHz~24 GHz	U=0.079 dB		
				10 dBm~15 dBm ,24 GHz~67 GHz	U=0.12 dB		
				3 dBm~10 dBm ,9 kHz~100 kHz	U=0.058 dB		
				3 dBm~10 dBm ,100 kHz~8 GHz	U=0.043 dB		
				3 dBm~10 dBm ,8 GHz~24 GHz	U=0.075 dB		
				3 dBm~10 dBm ,24 GHz~67 GHz	U=0.12 dB		
				-3 dBm~3 dBm ,9 kHz~12.4 GHz	U=0.03 dB		
				-3 dBm~3 dBm ,12.4 GHz~18 GHz	U=0.034 dB		
				-3 dBm~3 dBm ,18 GHz~24 GHz	U=0.073 dB		
				-3 dBm~3 dBm ,24 GHz~26.5 GHz	U=0.088 dB		
				-3 dBm~3 dBm ,26.5 GHz~33 GHz	U=0.10 dB		
				-3 dBm~3 dBm ,33 GHz~67 GHz	U=0.12 dB		
				-20 dBm~-3 dBm ,9 kHz~100 kHz	U=0.13 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~-3 dBm ,100 kHz~8 GHz	U=0.047 dB		
				-20 dBm~-3 dBm ,8 GHz~12.4 GHz	U=0.061 dB		
				-20 dBm~-3 dBm ,12.4 GHz~14 GHz	U=0.070 dB		
				-20 dBm~-3 dBm ,14 GHz~18 GHz	U=0.074 dB		
				-20 dBm~-3 dBm ,18 GHz~24 GHz	U=0.077 dB		
				-20 dBm~-3 dBm ,24 GHz~26.5 GHz	U=0.092 dB		
				-20 dBm~-3 dBm ,26.5 GHz~33 GHz	U=0.10 dB		
				-20 dBm~-3 dBm ,33 GHz~67 GHz	U=0.12 dB		
				-30 dBm~-20 dBm ,9 kHz~100 kHz	U=0.16 dB		
				-30 dBm~-20 dBm ,100 kHz~8 GHz	U=0.047 dB		
				-30 dBm~-20 dBm ,8 GHz~12.4 GHz	U=0.061 dB		
				-30 dBm~-20 dBm ,12.4 GHz~14 GHz	U=0.070 dB		
				-30 dBm~-20 dBm ,14 GHz~18 GHz	U=0.074 dB		
				-30 dBm~-20 dBm ,18 GHz~24 GHz	U=0.077 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-30 dBm~-20 dBm ,24 GHz~26.5 GHz	U=0.092 dB		
				-30 dBm~-20 dBm ,26.5 GHz~33 GHz	U=0.10 dB		
				-30 dBm~-20 dBm ,33 GHz~67 GHz	U=0.12 dB		
				-90 dBm~-30 dBm ,9 kHz~100 kHz	U=0.16 dB		
				-90 dBm~-30 dBm ,100 kHz~8 GHz	U=0.048 dB		
				-90 dBm~-30 dBm ,8 GHz~12.4 GHz	U=0.062 dB		
				-90 dBm~-30 dBm ,12.4 GHz~14 GHz	U=0.071 dB		
				-90 dBm~-30 dBm ,14 GHz~18 GHz	U=0.075 dB		
				-90 dBm~-30 dBm ,18 GHz~24 GHz	U=0.078 dB		
				-90 dBm~-30 dBm ,24 GHz~26.5 GHz	U=0.093 dB		
				-90 dBm~-30 dBm ,26.5 GHz~33 GHz	U=0.10 dB		
				-90 dBm~-30 dBm ,33 GHz~50 GHz	U=0.12 dB		
				-110 dBm~-90 dBm ,9 kHz~100 kHz	U=0.16 dB		
				-110 dBm~-90 dBm ,100 kHz~8 GHz	U=0.086 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-110 dBm~-90 dBm ,8 GHz~12.4 GHz	U=0.069 dB		
				-110 dBm~-90 dBm ,12.4 GHz~14 GHz	U=0.077 dB		
				-110 dBm~-90 dBm ,14 GHz~18 GHz	U=0.081 dB		
				-110 dBm~-90 dBm ,18 GHz~24 GHz	U=0.084 dB		
				-110 dBm~-90 dBm ,24 GHz~26.5 GHz	U=0.098 dB		
				-110 dBm~-90 dBm ,26.5 GHz~33 GHz	U=0.11 dB		
				-110 dBm~-90 dBm ,33 GHz~40 GHz	U=0.12 dB		
				-110 dBm~-90 dBm ,40 GHz~50 GHz	U=0.14 dB		
				-130 dBm~-110 dBm ,9 kHz~100 kHz	U=0.16 dB		
				-130 dBm~-110 dBm ,100 kHz~8 GHz	U=0.071 dB		
				-130 dBm~-110 dBm ,8 GHz~12.4 GHz	U=0.083 dB		
				-130 dBm~-110 dBm ,12.4 GHz~14 GHz	U=0.089 dB		
				-130 dBm~-110 dBm ,14 GHz~18 GHz	U=0.092 dB		
				-130 dBm~-110 dBm ,18 GHz~24 GHz	U=0.096 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-130 dBm~-110 dBm ,24 GHz~26.5 GHz	U=0.11 dB		
				-130 dBm~-110 dBm .26.5 GHz~33 GHz	U=0.12 dB		
				-130 dBm~-110 dBm ,33 GHz~40 GHz	U=0.13 dB		
				-130 dBm~-110 dBm ,40 GHz~50 GHz	U=0.14 dB		
				-140 dBm~-130 dBm ,9 kHz~100 kHz	U=0.17 dB		
				-140 dBm~-130 dBm ,100 kHz~8 GHz	U=0.077 dB		
				-140 dBm~-130 dBm ,8 GHz~24 GHz	U=0.10 dB		
				-140 dBm~-130 dBm ,24 GHz~50 GHz	U=0.15 dB		
		Timebase Frequency		10 MHz	$U_{rel}=1 \times 10^{-11}$		
		Harmonic Distortion		-146 dBc~-25 dBc ,Carrier 0.5 MHz~1 MHz	U=0.13 dB		
				-146 dBc~-30dBc ,Carrier 1 MHz~0.7 GHz	U=0.61 dB		
				-146 dBc~-30dBc ,Carrier 0.7 GHz~1 GHz	U=0.52 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-146 dBc~-30 dBc Carrier 1.0 GHz~10 GHz	U=0.80 dB		
				-146 dBc~-30 dBc Carrier 10 GHz~20 GHz	U=0.90 dB		
				-146 dBc~-30 dBc Carrier Frequency: 20 GHz~50 GHz	U=1.4 dB		
		Spurious		-146 dBc~-30dBc ,9 kHz~2 GHz	U=0.089 dB		
				-146 dBc~-30dBc ,2 GHz~6 GHz	U=0.096 dB		
				-146 dBc~-30 dBc ,6 GHz~26.5 GHz	U=0.089 dB		
				-146 dBc~-30 dBc ,6 GHz~50 GHz	U=0.090 dB		
		Pulse Modulation On/Off Ratio		80 dB~120 dB,9 kHz~50 GHz	U=0.089 dB		
		FM Deviation		10 Hz~100 kHz,Carrier : 9 kHz~50 GHz,Mod Rate: 1 kHz	U _{rel} =0.99%		
		FM Frequency Response		0 dB~1 dB,Carrier : 9 kHz~50 GHz,Mod Rate 80 Hz~100 kHz	U=0.12 dB		
		FM Distortion		0.1%~1%, Carrier : 10 MHz~3.3 GHz, Mod Rate 1 kHz	U=0.002%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Residual FM	ilac-M	0.1 Hz~100 Hz, Carrier 250 kHz~250 MHz	U=0.06 Hz		
				0.1 Hz~100 Hz, Carrier 250 MHz~500 MHz	U=0.03 Hz		
				0.1 Hz~100 Hz, Carrier 500 MHz~ 1000 MHz	U=0.06 Hz		
				0.1 Hz~100 Hz, Carrier 1000 MHz~2000 MHz	U=0.12 Hz		
				0.1 Hz~100 Hz, Carrier 2000 MHz~4000 MHz	U=0.24 Hz		
				0.1 Hz~100 Hz, Carrier 4000 MHz~6000 MHz	U=0.48 Hz		
		Phase Modulation Deviation	ilac-M	1 rad~100 rad, Carrier : 10 MHz~3.3 GHz, Mod Rate 200 Hz ~ 20 kHz	U _{rel} =0.13%		
		Phase Modulation Frequency Response		0 dB~3 dB,Carrier 500 MHz~1000 MHz, Mod rate 80 Hz~100 kHz	U=0.12 dB		
		Phase Modulation Frequency Response		0 dB~3 dB,Carrier 500 MHz~1000 MHz, mod rate 100 kHz~100 MHz	U=0.14 dB		
		Phase Modulation Distortion		0.03%~1%,Carrier Frequency: 500 MHz~ 700 MHz,Phase deviation:(1~90) rad	U=0.004%		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AM		0.03%~1%, Carrier Frequency: 700 MHz~ 1000 MHz, Phase deviation: (1~90) rad	$U=0.007\%$		
				AM depth 5%~99%, Carrier Frequency: 100 kHz~ 10 MHz, Mod Rate 50 Hz ~ 10 kHz	$U_{rel}=0.75\%$		
				AM depth 20%~99%, Carrier Frequency: 10 MHz~3 GHz, Mod Rate 50 Hz ~ 100 kHz	$U_{rel}=0.5\%$		
				AM depth 20%~99%, Carrier Frequency: 3 GHz~26.5 GHz, Mod Rate 50 Hz ~ 100 kHz	$U_{rel}=1.5\%$		
				AM depth 20%~99%, Carrier Frequency: 26.5 GHz~ 31.15 GHz, Mod Rate 50 Hz ~ 100 kHz	$U_{rel}=1.9\%$		
				AM depth 20%~99%, Carrier Frequency: 31.15 GHz~ 50 GHz, Mod Rate 50 Hz ~ 100 kHz	$U_{rel}=6\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AM Distortion		0.01%~3%,Carrier Frequency: 100 kHz~50 GHz, Mod Rate 20 Hz~100 kHz	U=0.006%		
		AM Frequency Response		-3 dB~3 dB,Carrier Frequency: 100 kHz~50 GHz,Mod Rate 10 Hz~20 kHz	U=0.025 dB		
				-3 dB~3 dB,Carrier Frequency: 100 kHz~50 GHz,Mod Rate 20 kHz~100 kHz	U=0.12 dB		
		DC FM Carrier Offset		-0.5%~0.5%,Carrier Frequency: 200 MHz~4500 MHz,Deviation 1 kHz~500 kHz	U=0.0051%		
				-0.5%~0.5%,Carrier Frequency: 200 MHz~4500 MHz,Deviation 500 kHz ~4 MHz	U=0.0001%		
		Amplitude Switching Speed		10 μs~500 μs	U=0.91 μs		
				500 μs~10 ms	U=1.3 μs		
		Burst Modulation On/Off Ratio		70 dB~130 dB,Carrier Frequency: 0.25 MHz~4 GHz	U=0.08 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Digital Modulation Power Relative to CW		0 dBm~15 dBm,Carrier Frequency: 9 kHz~50 GHz	U=0.002 dB		
		GSM Phase Error RMS		0.1°~1°,Carrier Frequency: 800 MHz~2200 MHz	U=0.02 °		
		GSM Phase Error Peak		0.1°~1°,Carrier Frequency: 800 MHz~2200 MHz	U=0.12 °		
		EDGE EVM		0.1%~5%,Carrier Frequency: 800 MHz~2200 MHz	U=0.025%		
		CDMA RHO		0.1~1,Carrier Frequency: 800 MHz~2200 MHz	U=0.0002		
		CDMA ACP		-80 dB~-30 dB,Carrier Frequency: 800 MHz~2200 MHz	U=0.12 dB		
		CDMA EVM		0.1%~5%,Carrier Frequency: 800 MHz~2200 MHz	U=0.023%		
		WCDMA EVM		0.1%~1%,Carrier Frequency: 800 MHz~2200 MHz	U=0.03%		
		WCDMA ACP		-80 dB~-30 dB,Carrier Frequency: 800 MHz~2200 MHz(include ACLR)	U=0.30 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		LTE ACP		-80 dB~ -30 dB,Carrier Frequency: 800 MHz~ 2200 MHz	U=0.50 dB		
		DECT Position Deviation Error		0.1 kHz~3 kHz,Carrier Frequency: 800 MHz~ 2200 MHz	U=0.12 kHz		
		Custom EVM		0.1%~5%,Carrier Frequency: 100 MHz~44 GHz (π/4QPSK、 PHS、PDC、NADC、 TETRA、BPSK、 QPSK、QAM16、 QAM64、QAM256)	U=0.01%		
		2FSK Freq Shift Error		0.1%~5%,Carrier Frequency: 100 MHz~ 6000 MHz	U=0.048%		
		GMSK Phase Error RMS		0.1° ~1° ,Carrier Frequency: 800 MHz~ 6000 MHz	U=0.013 °		
		Frequency Switching Speed		180 ns~100 μs,Carrier Frequency: 9 kHz~50 GHz	U=4 ns		
				100 μs~10 ms,Carrier Frequency: 9 kHz~50 GHz	U=50 μs		
				10 ms~100 ms,Carrier Frequency: 9 kHz~50 GHz	U=0.8 ms		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pulse Modulation Level		0 dBm~9 dBm,Carrier Frequency: 9 kHz~50 GHz	U=0.028 dB		
		Pulse Modulation Rise/Fall Time		20 ps~10 ns,Carrier Frequency: 9 kHz~50 GHz	U=0.41 ns		
		Residual Phase Noise		-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 100 MHz~3.2 GHz,Deviation: 1 Hz~10 Hz	U=2.8 dB		
				-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 100 MHz~3.2 GHz,Deviation: 10 Hz~100 Hz	U=2.5 dB		
				-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 100 MHz~3.2 GHz,Deviation: 100 Hz~10 kHz	U=2.4 dB		
				-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 100 MHz~3.2 GHz,Deviation: 10 kHz~1 MHz	U=2.3 dB		




No. CNAS L0640

在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Single Sideband Phase Noise		-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 1 MHz~20 GHz,Deviation: 1 Hz~10 Hz	U=2.9 dB		
				-150 dBc/Hz~-20 dBc/Hz,Carrier Frequency: 1 MHz~20 GHz,Deviation: 10 Hz~100 Hz	U=2.6 dB		
				-160 dBc/Hz~-25 dBc/Hz,Carrier Frequency: 1 MHz~20 GHz,Deviation: 100 Hz~10 kHz	U=2.4 dB		
				-170 dBc/Hz~-25 dBc/Hz,Carrier Frequency: 1 MHz~20 GHz,Deviation: 10 kHz~1 MHz	U=2.3 dB		
				-170 dBc/Hz~-25 dBc/Hz,Carrier Frequency: 10 MHz~200 MHz,Deviation: 1 MHz~100 MHz	U=4.6 dB		
		Sub-Harmonic Spurious		-146 dBc~-40 dBc,载波频率: 9 kHz~50 GHz	U=0.75 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Swept Frequency Accuracy		9 kHz~50 GHz	$U_{rel}=3.4 \times 10^{-6}$		
2	*RF/MW Frequency Synthesizer	Power Level	RF/MW Frequency synthesizer Test Procedure CAL-CHA001-02	0 dBm~18 dBm,100 kHz~4.2 GHz	$U=0.08$ dB		
				0 dBm~13 dBm,4.2 GHz~6 GHz	$U=0.09$ dB		
				0 dBm~10 dBm,6 GHz~10 GHz	$U=0.13$ dB		
				0 dBm~10 dBm,18 GHz~26.5 GHz	$U=0.18$ dB		
				0 dBm~10 dBm,26.5 GHz~36 GHz	$U=0.25$ dB		
				0 dBm~10 dBm,36 GHz~50 GHz	$U=0.42$ dB		
				10 dBm~20 dBm,6 GHz~10 GHz	$U=0.19$ dB		
				10 dBm~20 dBm,10 GHz~18 GHz	$U=0.22$ dB		
				0 dBm~-127 dBm,3 MHz~1 GHz	$U=0.11$ dB+0.005 dB/10 dB		
				0 dBm~-127 dBm,1 GHz~6.0 GHz	$U=0.02$ dB+0.005 dB/10 dB		
				0 dBm~-90 dBm,6 GHz~26.5 GHz	$U=0.18$ dB+0.005 dB/10 dB		
				0 dBm~-90 dBm,26.5 GHz~33.5 GHz	$U=0.25$ dB+0.005 dB/10 dB		



No. CNAS L0640

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
				0 dBm~-20 dBm,33.5 GHz~50 GHz	U=0.25 dB			
					-20 dBm~-70 dBm,33.5 GHz~50 GHz	U=0.41 dB		
					-70 dBm~-90 dBm,33.5 GHz~50 GHz	U=0.42 dB		
		Frequency		10 MHz	U _{rel} =1.0x10 ⁻¹¹			
		Harmonic Distortion		-146 dBc~-30 dBc ,1 MHz~2 GHz	U=0.5 dB			
					-146 dBc~-30 dBc ,2 GHz~4 GHz	U=0.6 dB		
					-146 dBc~-30dBc ,4 GHz~6.0 GHz	U=1.2 dB		
					-146 dBc~-30dBc ,6.0 GHz~7.0 GHz	U=1.5 dB		
					-146 dBc~-30 dBc ,7.0 GHz~20 GHz	U=1.8 dB		
		Spurious		-130 dBc~-50 dBc ,184 kHz~1.0 MHz	U=1.2 dB			
					-130 dBc~-50 dBc ,1.0 MHz~41.4 MHz	U=1.1 dB		
					-130 dBc~-50 dBc ,41.1 MHz~1.602 GHz	U=0.8 dB		
					-130 dBc~-50 dBc ,1.602 GHz~1.925 GHz	U=1.7 dB		
					-130 dBc~-50 dBc ,1.925 GHz~3.0 GHz	U=2.0 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-130 dBc ~ -50 dBc ,3.0 GHz ~ 6.0 GHz	U=2.0 dB		
				-130 dBc ~ -50 dBc ,6.0 GHz ~ 13.5 GHz	U=1.8 dB		
				-130 dBc ~ -50 dBc ,13.5 GHz ~ 20 GHz	U=1.9 dB		
				-130 dBc ~ -50 dBc ,20 GHz ~ 40 GHz	U=2.0 dB		
				-130 dBc ~ -50 dBc ,40 GHz ~ 50 GHz	U=1.8 dB		
		FM Deviation		5 kHz ~ 100 kHz, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 6 GHz	U _{rel} =0.96%		
				0.8 kHz ~ 100 kHz, Mod Rate 10 kHz, Carrier frequency 0.25 MHz ~ 6 GHz	U _{rel} =2.0%		
				25 kHz ~ 350 kHz, Mod Rate 20 kHz, Carrier frequency 10 MHz ~ 26.5 GHz	U _{rel} =0.86%		
				25 kHz ~ 375 kHz, Mod Rate 100 kHz, Carrier frequency 10 MHz ~ 26.5 GHz	U _{rel} =0.96%		
				500 kHz, Mod Rate 100 kHz, Carrier frequency 10 MHz ~ 26.5 GHz	U _{rel} =0.11%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				1 MHz~2 MHz, Mod Rate 100 kHz, Carrier frequency 10 MHz ~ 26.5 GHz	$U_{rel}=0.24\%$		
				2.4 MHz, Mod Rate 1 MHz, Carrier frequency 10 MHz ~ 26.5 GHz	$U_{rel}=0.054\%$		
				5 MHz, Mod Rate 1 MHz, Carrier frequency 10 MHz ~ 26.5 GHz	$U_{rel}=0.24\%$		
				8 MHz, Mod Rate 1 MHz, Carrier frequency 10 MHz ~ 26.5 GHz	$U_{rel}=1.4\%$		
		Phase Modulation Deviation		0.1 rad~1.0 rad, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=1.0\%$		
				1 rad~1.5 rad, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=0.6\%$		
				1.5 rad~2.0 rad, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=0.5\%$		
				2 rad~5.5 rad, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=0.2\%$		
				100 rad, Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=2.9\%$		




No. CNAS L0640

第 18 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AM		10%~90%,Mod Rate 1 kHz, Carrier frequency 0.25 MHz ~ 26.5 GHz	$U_{rel}=1.1\%$		
		AM Distortion		1%~6% ,AM: 30 %~90%, Mod Rate 1 kHz,Carrier frequency 0.25 MHz ~ 26.5 GHz	$U=0.47\%$		
		AM Frequency Response		0 dB~3 dB,Mod Rate 20 Hz~1 kHz,Carrier frequency 0.25 MHz ~ 26.5 GHz	$U=0.08$ dB		
				0 dB~3 dB,Mod Rate 1 kHz~10 kHz,Carrier frequency 0.25 MHz ~ 26.5 GHz	$U=0.02$ dB		
				0 dB~3 dB,Mod Rate 10 kHz~40 kHz,Carrier frequency 0.25 MHz ~ 26.5 GHz	$U=0.08$ dB		
				0 dB~3 dB,Mod Rate 40 kHz~250 kHz,Carrier frequency 0.25 MHz ~ 26.5 GHz	$U=0.087$ dB		
		Single Side Band Phase Noise		-160 dBc/Hz~-25 dBc/Hz,Carrier Frequency: 100 MHz~20 GHz,Deviation: 10 Hz~1 MHz	$U=1.0$ dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
3	*RF/MW Signal Generator	ALC Linearity	RF/MW Signal Generator Test Procedure CAL-CHA001-03	0 dBm~13 dBm,100 kHz~2.5 MHz	U=0.11 dB		
				0 dBm~13 dBm,2.5 MHz~500 MHz	U=0.12 dB		
				0 dBm~13 dBm,500 MHz~20 GHz	U=0.20 dB		
				0.01%~3%,Carrier Frequency: 100 MHz~20 GHz,AM:10%, Mod rate 1kHz	U=0.03%		
		0.01%~3%,Carrier Frequency: 1.5 MHz~5 MHz,AM:30%, Mod rate 1kHz		U=0.12%			
		0.01%~3%,Carrier Frequency: 5 MHz~3 GHz,AM:30%, Mod rate 1kHz		U=0.075%			
		0.01%~3%,Carrier Frequency: 3 GHz~6 GHz,AM:30%, Mod rate 1kHz		U=0.15%			
		0.01%~3%,Carrier Frequency: 1.5 MHz~5 MHz,AM:80%, Mod rate 1kHz		U=0.37%			
0.01%~3%,Carrier Frequency: 5 MHz~3 GHz,AM:80%, Mod rate 1kHz	U=0.12%						
		AM Distortion					




No. CNAS L0640

第 20 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AM Frequency Response		0.01%~3%, Carrier Frequency: 3 GHz~6 GHz, AM: 80%, Mod rate 1 kHz	$U=0.24\%$		
				0 dB~3 dB, Carrier Frequency: 100 MHz~ 2.7 GHz, Mod Rate: 10 Hz~10 kHz	$U=0.027\text{ dB}$		
				0 dB~3 dB, Carrier Frequency: 100 MHz~ 2.7 GHz, Mod Rate: 10 kHz~50 kHz	$U=0.04\text{ dB}$		
				0 dB~3 dB, Carrier Frequency: 2.7 GHz~20 GHz, Mod Rate: 10 Hz~ 50 kHz	$U=0.13\text{ dB}$		
		AM		1%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	$U=0.08\%$		
				1%~2%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	$U=0.10\%$		
				2%~5%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	$U=0.14\%$		
				5%~10%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	$U=0.10\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				10%~15%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.17%		
				15%~20%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.20%		
				20%~30%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.26%		
				30%~50%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.31%		
				50%~60%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.48%		
				60%~70%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.39%		
				70%~80%, Carrier Frequency: 100 kHz~6 GHz, Mod rate: 1 kHz	U=0.27%		
				10%~80%,Carrier Frequency: 20 GHz, Mod rate: 1 kHz	U=1.3%		
		Analog Ramp Sweep		10 MHz~20 GHz	U=0.001 GHz		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Power Level		-120.1 dBm~0 dBm,Carrier Frequency: 1 MHz~1275 MHz	U=0.082 dB		
				-119 dBm~-96.1 dBm,Carrier Frequency: 1275 MHz~5400 MHz	U=0.20 dB		
				-96.1 dBm~0 dBm,Carrier Frequency: 1275 MHz~5400 MHz	U=0.15 dB		
				0 dBm~13 dBm,Carrier Frequency: 1 MHz~3000 MHz	U=0.13 dB		
				0 dBm~11 dBm,Carrier Frequency: 3000 MHz~6000 MHz	U=0.14 dB		
				-10 dBm~6 dBm,Carrier Frequency: 6 GHz~20 GHz	U=0.31 dB		
	Carrier Frequency			10 kHz~5.4 GHz	$U_{rel}=1.9 \times 10^{-7}$		
	Carrier Frequency Offset			-13.5 MHz~13.5 MHz,Carrier Frequency: 1350 MHz	U=0.40 kHz		
	Carrier Frequency Offset with FM	-50 Hz~50 Hz,Carrier Frequency: 2 GHz	U=1.0 Hz				




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		EXT Mod ALC Flatness		-1 dB~1 dB,Frequency: 10 Hz~ 20 Hz,Amplitude: 1 V	U=0.0055 dB		
				-1 dB~1 dB,Frequency: 20 Hz~ 50 Hz,Amplitude: 1 V	U=0.0058 dB		
				-1 dB~1 dB,Frequency: 50 Hz~ 20kHz Hz,Amplitude: 1 V	U=0.0076 dB		
				-1 dB~1 dB,Frequency: 20 kHz~ 200kHz Hz,Amplitude: 1 V	U=0.029 dB		
				-1 dB~1 dB,Frequency: 200 kHz~500kHz Hz,Amplitude: 1 V	U=0.089 dB		
				-1 dB~1 dB,Frequency: 10 Hz~ 20 Hz,Amplitude: 0.7 V	U=0.0056 dB		
				-1 dB~1 dB,Frequency: 20 Hz~ 50 Hz,Amplitude: 0.7 V	U=0.0055 dB		
				-1 dB~1 dB,Frequency: 50 Hz~ 20kHz Hz,Amplitude: 0.7 V	U=0.008 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-1 dB~1 dB, Frequency: 20 kHz~200kHz Hz, Amplitude: 0.7 V	U=0.029 dB		
				-1 dB~1 dB, Frequency: 200 kHz~500kHz Hz, Amplitude: 0.7 V	U=0.089 dB		
				-1 dB~1 dB, Frequency: 10 Hz~20 Hz, Amplitude: 1.4 V	U=0.0085 dB		
				-1 dB~1 dB, Frequency: 20 Hz~50 Hz, Amplitude: 1.4 V	U=0.0075 dB		
				-1 dB~1 dB, Frequency: 50 Hz~20kHz Hz, Amplitude: 1.4 V	U=0.0092 dB		
				-1 dB~1 dB, Frequency: 20 kHz~200kHz Hz, Amplitude: 1.4 V	U=0.035 dB		
				-1 dB~1 dB, Frequency: 200 kHz~500kHz Hz, Amplitude: 1.4 V	U=0.095 dB		
		FM Frequency Response		0 dB~3 dB, Carrier Frequency: 35.5 MHz, Mod Rate: 30 Hz	U=0.55 dB		




No. CNAS L0640

第 25 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB~3 dB,Carrier Frequency: 35.5 MHz,Mod Rate: 30 Hz~100 kHz	U=0.22 dB		
				0 dB~3 dB,Carrier Frequency: 2 GHz,Mod Rate: 20 kHz~200 kHz	U=0.09 dB		
		FM Distortion		0.01%~3%,Carrier Frequency: 21.1 MHz~ 39.6 MHz,Deviation : 210 kHz~396 kHz,, Mod Rate 1 kHz	U=0.37%		
				0.01%~0.5%,Carrier Frequency: 2 GHz,Deviation : 1 kHz~50 kHz, Mod Rate 1 kHz	U=0.03%		
		FM Deviation		1 kHz~396 kHz,Carrier Frequency: 100 kHz~6 GHz, Mod Rate 1 kHz	U _{rel} =1.0%		
		Frequency Setting Time		0.01 ms~2 ms,Carrier Frequency: 100 kHz~6 GHz	U=0.01 ms		
		Harmonic Spurious		-70 dBc~-40 dBc,Carrier Frequency: 10 kHz ~ 100 kHz	U=0.98 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 100 kHz ~ 5 MHz	U=0.49 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-70 dBc~-40 dBc,Carrier Frequency: 5 MHz ~ 1.5 GHz	U=0.31 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 1.5 GHz ~ 3 GHz	U=0.74 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 3 GHz ~ 3.3 GHz	U=1.3 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 3.3 GHz ~ 6.6 GHz	U=1.6 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 6.6 GHz ~ 20 GHz	U=2.9 dB		
		IF Input		-18 dBm, IF Frequency: 10 MHz~700 MHz,Carrier Frequency: 10 MHz ~ 2 GHz	U=0.23 dB		
				-18 dBm, IF Frequency: 10 MHz~700 MHz,Carrier Frequency: 2 GHz ~ 6 GHz	U=0.74 dB		
				-18 dBm, IF Frequency: 10 MHz~700 MHz,Carrier Frequency: 6 GHz ~ 12 GHz	U=1.4 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-18 dBm, IF Frequency: 10 MHz~700 MHz,Carrier Frequency: 12 GHz ~ 16 GHz	U=1.7 dB		
				-18 dBm, IF Frequency: 10 MHz~700 MHz,Carrier Frequency: 16 GHz ~ 20 GHz	U=1.8 dB		
		IQ Error Vector		0.01%~1.5%,Carrier Frequency: 100 kHz~6 GHz	U=0.74%		
		IQ Input Impedance		1~1.2,Frequency: 1 MHz	U=0.007		
				1~1.2,Frequency: 50 MHz	U=0.004		
		IQ Residual Carrier And Leakage		-70 dBc~-40 dBc,Carrier Frequency: 100 kHz~6 GHz	U=0.17 dB		
		Level Frequency Response		0 dB~1 dB, Frequency: 300 kHz~4 GHz,5 dBm	U=0.13 dB		
					0 dB~1 dB, Frequency: 4 GHz~6 GHz,5 dBm	U=0.14 dB	
		Level Settling Time		3 ms~5 ms,Amplitude: -130 dBm~0 dBm,Frequency: 30 MHz~5 GHz	U=0.0062 ms		
		Mod Gen Distortion		0.01%~ 0.1%,Frequency: 100 Hz~20 kHz	U=0.013%		




No. CNAS L0640

第 28 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.01%~ 0.1%,Frequency: 20 kHz~100 kHz	U=0.032%		
		MOD GEN Frequency Response		2 dB,Frequency: 10 Hz~20 Hz	U=0.0055 dB		
				2 dB,Frequency: 20 Hz~100 Hz	U=0.0054 dB		
				2 dB,Frequency: 100 Hz~10 kHz	U=0.0057 dB		
				2 dB,Frequency: 10 kHz~30 kHz	U=0.0075 dB		
				2 dB,Frequency: 30 kHz~70 kHz	U=0.0098 dB		
				2 dB,Frequency: 70 kHz~100 kHz	U=0.029 dB		
				2 dB,Frequency: 100 kHz~300 kHz	U=0.089 dB		
				3 dB,Frequency: 300 kHz~1 MHz	U=0.15 dB		
				3 dB,Frequency: 1 MHz~4 MHz	U=0.43 dB		
				3 dB,Frequency: 4 MHz~10 MHz	U=1.6 dB		
		MOD GEN Output Level		10 mV, Frequency: 1 kHz	U=0.0057 mV		
				30 mV, Frequency: 1 kHz	U=0.0071 mV		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				100 mV, Frequency: 1 kHz	U=0.13 mV		
				100 mV~300 mV, Frequency: 1 kHz	U=0.59 mV		
				300 mV~1 V, Frequency: 1 kHz	U=1.1 mV		
				1 V~3 V, Frequency: 1 kHz	U=5.9 mV		
		Maximum Leveled Output Power		10 dBm~20 dBm, Frequency: 10 MHz~100 MHz	U=0.22 dB		
				10 dBm~20 dBm, Frequency: 100 MHz~2 GHz	U=0.19 dB		
				10 dBm~20 dBm, Frequency: 2 GHz~20 GHz	U=0.31 dB		
		MOD ALC Distortion		0.01%~0.1%, Frequency: 50 Hz~20 kHz, Amplitude: 1V	U=0.013%		
		MOD GEN Frequency		1 kHz~40 kHz	U=0.00061 Hz		
				40 kHz~1 MHz	U=0.0015 Hz		
		Non-Harmonic Spurious		-70 dBc~-40 dBc, Frequency: 10 kHz~1 MHz	U=1.0 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-70 dBc~-40 dBc, Frequency: 1 MHz~3 GHz	U=0.21 dB		
				-70 dBc~-40 dBc, Frequency: 3 GHz~6.6 GHz	U=0.74 dB		
				-70 dBc~-40 dBc, Frequency: 6.6 GHz~13.2 GHz	U=1.4 dB		
				-70 dBc~-40 dBc, Frequency: 13.2 GHz~18 GHz	U=1.7 dB		
				-70 dBc~-40 dBc, Frequency: 18 GHz~20 GHz	U=1.8 dB		
		Output Impedance		1~1.9, Frequency: 300 kHz~6 GHz	U=0.021		
		Phase MOD Distortion		0.01%~3%, Mod Rate: 1 kHz, Phase: 10 rad	U=0.37%		
		Phase MOD Flatness		0 dB~3 dB, Mod Rate: 0.1 kHz~10 kHz, Phase: 10 rad, Carrier Frequency 10MHz~6 GHz	U=0.19 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB~3 dB ,Mod Rate: 10 kHz~20 kHz,Phase: 0.3 rad~10 rad, Carrier Frequency 10MHz~6 GHz	U=0.45 dB		
		Phase MOD ON AM		0.001 rad~0.1 rad,Carrier Frequency: 500 MHz,AM: 30%,Mod Rate: 1 kHz	U=0.031 rad		
		Phase MOD		0.5 rad,500 MHz	U=0.015 rad		
				0.625 rad,500 MHz	U=0.019 rad		
				1 rad,500 MHz	U=0.01 rad		
				2.5 rad,500 MHz	U=0.025 rad		
				5 rad,500 MHz	U=0.05 rad		
				0.625 rad,70.687 MHz	U=0.019 rad		
				10.0 rad, 46.8 MHz~6 GHz	U=0.1 rad		
		Pulse Generator Pulse Width		1 μs~100 μs	U=0.01 ns		
		Pulse MOD Video Crosstalk		-30 dBc,Carrier Frequency: 1000 MHz	U=0.17 dB		
		Pulse Modulation On/Off Ratio		78 dB,Carrier Frequency: 50 MHz~6 GHz	U=0.16 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pulse Modulation: Rise/Fall Time		0.01 ns~12 ns, Carrier Frequency: 500 MHz~20 GHz	U=0.01 ns		
		Residual FM		0.01 Hz~4 Hz, Carrier Frequency: 1 GHz	U=0.61 Hz		
		Residual AM		0.001%~0.02%, Carrier Frequency: 100 MHz~6 GHz	U=0.001%		
		Single Sideband Phase Noise		-110 dBc/Hz~-70 dBc/Hz, Carrier Frequency: 10 MHz~20 GHz	U=0.17 dB		
				-110 dBc/Hz~-70 dBc/Hz, Carrier Frequency: 10 MHz~20 GHz	U=0.28 dB		
				-110 dBc/Hz~-70 dBc/Hz, Carrier Frequency: 10 MHz~20 GHz	U=0.37 dB		
		Sub-Harmonic Spurious		-20 dBc~-80 dBc, Carrier Frequency: 1 GHz~3 GHz	U=0.31 dB		
				-20 dBc~-80 dBc, Carrier Frequency: 3 GHz~6 GHz	U=0.5 dB		
				-20 dBc~-80 dBc, Carrier Frequency: 6 GHz~6.6 GHz	U=1.3 dB		




在线扫码获取验证

No. CNAS L0640


第 33 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITED CERTIFICATE	-20 dBc~-80 dBc,Carrier Frequency: 6.6 GHz~ 13.2 GHz	U=1.6 dB		
				-20 dBc~-80 dBc,Carrier Frequency: 13.2 GHz~ 20 GHz	U=2.9 dB		
		Synchronous AM With FM		0.001%~0.1% ,Carrier Frequency: 10 MHz~6 GHz	U=0.007%		
		Synchronous PHIM With AM		0.001 rad~0.1 rad ,Carrier Frequency: 100 kHz~6 GHz	U=0.004 rad		
		Wideband Noise		-120 dBc/Hz~-160 dBc/Hz,Carrier Frequency: 20 MHz~3 GHz	U=0.49 dB		
				-120 dBc/Hz~-160 dBc/Hz,Carrier Frequency: 3 GHz~6 GHz	U=1.8 dB		
4	*High Precision Digital Multimeter	DC Voltage	High precision Digital Multimeter Test Procedure CAL-CHA002-01	100 mV	U=0.00032 mV		
				1 V	U=0.0000026 V		
				10 V	U=0.000025 V		
				100 V	U=0.00026 V		
				1000 V	U=0.0059 V		
		DC Voltage Offset		0V, 100 mV Range	U=0.00016 mV		



No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Voltage		0V, 1 V Range	U=0.00015 mV		
				0V, 10 V Range	U=0.00032 mV		
				0V, 100 V Range	U=0.014 mV		
				0V, 1000 V Range	U=0.041 mV		
				10 mV ,1 kHz	U=0.00089 mV		
				10 mV ,20 kHz	U=0.0012 mV		
				10 mV ,100 kHz	U=0.0085 mV		
				10 mV ,300 kHz	U=0.066 mV		
				10 mV ,1 MHz	U=0.021 mV		
				10 mV ,1 MHz~4 MHz	U=0.14 mV		
				100 mV ,1 kHz	U=0.0046 mV		
				100 mV ,20 kHz	U=0.0069 mV		
				100 mV ,100 kHz	U=0.036 mV		
				100 mV ,300 kHz	U=0.068 mV		
				100 mV ,1 MHz	U=0.17 mV		
				100 mV ,4 MHz~8 MHz	U=0.67 mV		
				100 mV ,10 MHz	U=2.3 mV		
				1 V ,1 kHz	U=0.000040 V		
				1 V ,20 kHz	U=0.000063 V		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				1 V ,50 kHz	U=0.00013 V		
				1 V ,100 kHz	U=0.00020 V		
				1 V ,300 kHz	U=0.00060 V		
				1 V ,500 kHz~1 MHz	U=0.0017 V		
				1 V ,1 MHz~8 MHz	U=0.0067 V		
				1 V ,10 MHz	U=0.025 V		
				3 V , 100 kHz	U=0.00056 V		
				3 V ,2 MHz~8 MHz	U=0.021 V		
				3 V ,10 MHz	U=0.075 V		
				10 V ,10 Hz	U=0.00055 V		
				10 V ,20 Hz	U=0.00038 V		
				10 V ,40 Hz	U=0.00031 V		
				10 V ,200 Hz~500 Hz	U=0.00050 V		
				10 V ,1 kHz	U=0.00040 V		
				10 V ,10 kHz	U=0.00062 V		
				10 V ,20 kHz	U=0.00064 V		
				10 V ,50 kHz	U=0.0012 V		
				10 V ,100 kHz	U=0.0016 V		
				10 V ,300 kHz	U=0.0051 V		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date		
				10 V ,500 kHz	U=0.016 V				
						10 V ,1 MHz	U=0.019 V		
						100 V ,1 kHz	U=0.0041 V		
						100 V ,20 kHz	U=0.0079 V		
						100 V ,50 kHz	U=0.0075 V		
						100 V ,100 kHz	U=0.023 V		
						700 V ,1 kHz	U=0.060 V		
		AC Current		10 μ A ,1 kHz	U=0.0074 μA				
						100 μ A ,1 kHz	U=0.017 μA		
						1 mA ,1 kHz	U=0.00011 mA		
						10 mA,1 kHz	U=0.0011 mA		
						100 mA,1 kHz	U=0.011 mA		
						1 A,1 kHz	U=0.00021 A		
		DC Current		100 μ A	U=0.0013 μA				
						1 mA	U=0.0000096 mA		
						10 mA	U=0.000094 mA		
						100 mA	U=0.0022 mA		
		DC Current Offset		1 A	U=0.000036 A				
						0A, 100 μ A Range	U=0.00012 μA		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	0A, 1 mA Range	$U=0.0000012 \text{ mA}$		
				0A, 10 mA Range	$U=0.000012 \text{ mA}$		
				0A, 100 mA Range	$U=0.00012 \text{ mA}$		
				0A, 1 A Range	$U=0.0000015 \text{ A}$		
		Resistance		10 Ω	$U=0.000086 \Omega$		
				100 Ω	$U=0.00083 \Omega$		
				1 k Ω	$U=0.0000074 \text{ k } \Omega$		
				10 k Ω	$U=0.000074 \text{ k } \Omega$		
				100 k Ω	$U=0.00074 \text{ k } \Omega$		
				1 M Ω	$U=0.0000077 \text{ M } \Omega$		
				10 M Ω	$U=0.00014 \text{ M } \Omega$		
				100 M Ω	$U=0.0018 \text{ M } \Omega$		
		Resistance Offset		0 Ω , 10 Ω Range	$U=0.000021 \Omega$		
		Frequency		1 Hz	$U=0.000020 \text{ Hz}$		
10 MHz	$U=0.000058 \text{ MHz}$						
5	*Digital Multimeter	DC Voltage	Digital Multimeter Test Procedure CAL-CHA002-02	10 mV~220 mV	$U=0.00077\% \times R_d + 0.4 \mu \text{ V}$		
				220 mV~2.2 V	$U=0.00051\% \times R_d + 0.7 \mu \text{ V}$		
				2.2 V~22 V	$U=0.00036\% \times R_d + 3 \mu \text{ V}$		
				22 V~ 220 V	$U=0.00051\% \times R_d + 40 \mu \text{ V}$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Voltage		220 V~1100 V	$U=0.00067\% \times R_d + 0.4 \text{ mV}$		
				10 mV~22 mV, 20 Hz~40 Hz	$U=0.0095\% \times R_d + 4 \text{ } \mu\text{V}$		
				10 mV~22 mV, 40 Hz~20 kHz	$U=0.0084\% \times R_d + 4 \text{ } \mu\text{V}$		
				10 mV~22 mV, 20 kHz~50 kHz	$U=0.021\% \times R_d + 4 \text{ } \mu\text{V}$		
				10 mV~22 mV, 50 kHz~100 kHz	$U=0.053\% \times R_d + 5 \text{ } \mu\text{V}$		
				22 mV~220 mV, 10 Hz~20 Hz	$U=0.025\% \times R_d + 13 \text{ } \mu\text{V}$		
				22 mV~220 mV, 20 Hz~40 Hz	$U=0.0095\% \times R_d + 7 \text{ } \mu\text{V}$		
				22 mV~220 mV, 40 Hz~20 kHz	$U=0.0060\% \times R_d + 7 \text{ } \mu\text{V}$		
				22 mV~220 mV, 20 kHz~50 kHz	$U=0.013\% \times R_d + 7 \text{ } \mu\text{V}$		
				22 mV~220 mV, 50 kHz~100 kHz	$U=0.033\% \times R_d + 18 \text{ } \mu\text{V}$		
				22 mV~220 mV, 100 kHz~300 kHz	$U=0.069\% \times R_d + 21 \text{ } \mu\text{V}$		
				22 mV~220 mV, 300 kHz~500 kHz	$U=0.15\% \times R_d + 27 \text{ } \mu\text{V}$		
				22 mV~220 mV, 500 kHz~1 MHz	$U=0.29\% \times R_d + 48 \text{ } \mu\text{V}$		
				220 mV~2.2 V, 10 Hz~20 Hz	$U=0.025\% \times R_d + 42 \text{ } \mu\text{V}$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				220 mV~2.2 V, 20 Hz~40 Hz	$U=0.0095\% \times R_d + 16 \mu V$		
				220 mV~2.2 V, 40 Hz~20 kHz	$U=0.0044\% \times R_d + 8 \mu V$		
				220 mV~2.2 V, 20 kHz~50 kHz	$U=0.0071\% \times R_d + 11 \mu V$		
				220 mV~2.2 V, 50 kHz~100 kHz	$U=0.0090\% \times R_d + 32 \mu V$		
				220 mV~2.2 V, 100 kHz~300 kHz	$U=0.035\% \times R_d + 84 \mu V$		
				220 mV~2.2 V, 300 kHz~500 kHz	$U=0.11\% \times R_d + 2 mV$		
				220 mV~2.2 V, 500 kHz~1 MHz	$U=0.18\% \times R_d + 3 mV$		
				2.2 V~22 V, 10 Hz~20 Hz	$U=0.025\% \times R_d + 42 \mu V$		
				2.2 V~22 V, 20 Hz~40 Hz	$U=0.0095\% \times R_d + 16 \mu V$		
				2.2 V~22 V, 40 Hz~20 kHz	$U=0.0044\% \times R_d + 8 \mu V$		
				2.2 V~22 V, 20 kHz~50 kHz	$U=0.0071\% \times R_d + 11 \mu V$		
				2.2 V~22 V, 50 kHz~100 kHz	$U=0.0090\% \times R_d + 32 \mu V$		
				2.2 V~22 V, 100 kHz~300 kHz	$U=0.034\% \times R_d + 84 \mu V$		
				2.2 V~22 V, 300 kHz~500 kHz	$U=0.011\% \times R_d + 2.1 mV$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				2.2 V~22 V,500 kHz~1 MHz	$U=0.18\% \times R_d+3.1 \text{ mV}$		
				22 V~220 V,20 Hz~40 Hz	$U=0.0095\% \times R_d+1.5 \text{ mV}$		
				22 V~220 V, 40 Hz~20 kHz	$U=0.0055\% \times R_d+0.6 \text{ mV}$		
				22 V~220 V,20 kHz~50 kHz	$U=0.0084\% \times R_d+1.1 \text{ mV}$		
				22 V~220 V, 50 kHz~100 kHz	$U=0.016\% \times R_d+2.6 \text{ mV}$		
				220 V~1000 V,40 Hz~50 Hz	$U=0.0095\% \times R_d+4.3 \text{ mV}$		
				220 V~1000 V, 50 Hz~1 kHz	$U=0.0072\% \times R_d+3.5 \text{ mV}$		
		AC Current		0.022 mA~0.22 mA,40 Hz~1 kHz	$U=0.011\% \times R_d+8 \text{ nA}$		
				0.022 mA~0.22 mA,1 kHz~5 kHz	$U=0.029\% \times R_d+12 \text{ nA}$		
				0.022 mA~0.22 mA,5 kHz~10 kHz	$U=0.11\% \times R_d+69 \text{ nA}$		
				0.22 mA~2.2 mA,10 Hz~20 Hz	$U=0.026\% \times R_d+42 \text{ nA}$		
				0.22 mA~2.2 mA,20 Hz~40 Hz	$U=0.017\% \times R_d+37 \text{ nA}$		
				0.22 mA~2.2 mA,40 Hz~1 kHz	$U=0.011\% \times R_d+37 \text{ nA}$		
				0.22 mA~2.2 mA,1 kHz~5 kHz	$U=0.021\% \times R_d+0.11 \mu \text{ A}$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.22 mA~2.2 mA, 5 kHz~10 kHz	$U=0.11\% \times R_d + 0.69 \mu A$		
				2.2 mA~22 mA, 10 Hz~20 Hz	$U=0.026\% \times R_d + 42 nA$		
				2.2 mA~22 mA, 20 Hz~40 Hz	$U=0.017\% \times R_d + 37 nA$		
				2.2 mA~22 mA, 40 Hz~1 kHz	$U=0.011\% \times R_d + 37 nA$		
				2.2 mA~22 mA, 1 kHz~5 kHz	$U=0.021\% \times R_d + 0.58 \mu A$		
				2.2 mA~22 mA, 5 kHz~10 kHz	$U=0.11\% \times R_d + 5.0 \mu A$		
				22 mA~220 mA, 10 Hz~20 Hz	$U=0.026\% \times R_d + 4.2 \mu A$		
				22 mA~220 mA, 20 Hz~40 Hz	$U=0.017\% \times R_d + 3.7 \mu A$		
				22 mA~220 mA, 40 Hz~1 kHz	$U=0.011\% \times R_d + 2.7 \mu A$		
				22 mA~220 mA, 1 kHz~5 kHz	$U=0.021\% \times R_d + 3.7 \mu A$		
				22 mA~220 mA, 5 kHz~10 kHz	$U=0.11\% \times R_d + 10 \mu A$		
				220 mA~2.2 A, 20 Hz~1 kHz	$U=0.026\% \times R_d + 37 \mu A$		
				220 mA~2.2 A, 1 kHz~5 kHz	$U=0.046\% \times R_d + 80 \mu A$		
				220 mA~2.2 A, 5 kHz~10 kHz	$U=0.73\% \times R_d + 0.16 mA$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Current	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	2.2 A~11 A,40 Hz~1 kHz	$U=0.048\% \times R_d+0.18 \text{ mA}$		
				2.2 A~11 A,1 kHz~5 kHz	$U=0.10\% \times R_d+0.40 \text{ mA}$		
				0.01 mA~0.22 mA	$U=0.0040\% \times R_d+6.2 \text{ nA}$		
				0.22 mA~2.2 mA	$U=0.0036\% \times R_d+7.1 \text{ nA}$		
				2.2 mA~22 mA	$U=0.0036\% \times R_d+40 \text{ nA}$		
				22 mA~220 mA	$U=0.0046\% \times R_d+0.72 \mu \text{ A}$		
				220 mA~2.2 A	$U=0.0082\% \times R_d+12 \mu \text{ A}$		
				2.2 A~11 A	$U=0.036\% \times R_d+0.48 \text{ mA}$		
		Resistance		1 Ω	$U=0.23 \text{ m}\Omega$		
				10 Ω	$U=0.23 \text{ m}\Omega$		
				100 Ω	$U=1 \text{ m}\Omega$		
				1 k Ω	$U=6.6 \text{ m}\Omega$		
				10 k Ω	$U=66 \text{ m}\Omega$		
				100 k Ω	$U=0.86 \Omega$		
				1 M Ω	$U=13 \Omega$		
				10 M Ω	$U=0.4 \text{ k}\Omega$		
				100 M Ω	$U=10 \text{ k}\Omega$		
		11 Ω~99.9 Ω	$U=0.0023\% \times R_d+0.9 \text{ m}\Omega$				




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				110 Ω ~ 0.99 k Ω	$U=0.0029\% \times R_d + 0.3 \text{ m}\Omega$		
				1.1 k Ω ~ 9.9 k Ω	$U=0.0029\% \times R_d + 0.3 \text{ m}\Omega$		
				11 k Ω ~ 99.9 k Ω	$U=0.0029\% \times R_d + 0.54 \text{ m}\Omega$		
				110 k Ω ~ 0.99 M Ω	$U=0.0033\% \times R_d + 4.7 \text{ m}\Omega$		
				1.1 M Ω ~ 9.9 M Ω	$U=0.0033\% \times R_d + 4.7 \text{ m}\Omega$		
				11 M Ω ~ 32.9 M Ω	$U=0.0063\% \times R_d + 41 \text{ m}\Omega$		
				33 M Ω ~ 99.9 M Ω	$U_{\text{rel}}=0.017\%$		
				1 G Ω	$U_{\text{rel}}=0.5\%$		
		Frequency		1 Hz ~ 1 MHz	$U=0.00026\% \times R_d + 0.001 \text{ Hz}$		
		Capacitance		1 μ F 100Hz	$U=0.46 \text{ nF}$		
				1 nF 1kHz	$U=13 \text{ pF}$		
				5 nF 1kHz	$U=20 \text{ pF}$		
				10 nF 1kHz	$U=33 \text{ pF}$		
				0.1 μ F 1kHz	$U=0.28 \text{ nF}$		
				9 μ F ~ 10 μ F 100Hz	$U=29 \text{ nF}$		
		0.1 mF 50Hz	$U=0.43 \text{ μ F}$				
		1 mF 20Hz	$U=4.3 \text{ μ F}$				
		Temperature		-200 ° C ~ -100 ° C	$U=0.33 \text{ ° C}$		
				-100 ° C ~ -30 ° C	$U=0.29 \text{ ° C}$		




No. CNAS L0640

第 44 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	-25 ° C~120 ° C	$U=0.18 \text{ } ^\circ \text{C}$		
				120 ° C~1000 ° C	$U=0.34 \text{ } ^\circ \text{C}$		
				1000 ° C~1372 ° C	$U=0.36 \text{ } ^\circ \text{C}$		
		Conductance		10 nS~50 nS	$U_{\text{rel}}=0.088\%$		
		Diode Voltage		1 V	$U=0.00013 \text{ V}$		
				3 V	$U=0.00054 \text{ V}$		
6	*Spectrum Analyzer	Time Base Frequency	Spectrum Analyzer Test Procedure CAL-CHA003-01	10 MHz	$U_{\text{rel}}=1.0 \times 10^{-11}$		
		Frequency Readout		500 MHz~1.505 GHz	$U=0.001 \text{ Hz}$		
				1.505 GHz~17.5 GHz	$U=1.1 \text{ Hz}$		
				17.5 GHz~21 GHz	$U=4 \text{ Hz}$		
				21 GHz~45 GHz	$U=5 \text{ Hz}$		
		Count		1 GHz	$U=0.001 \text{ Hz}$		
				1.5 GHz ~21 GHz	$U=0.02 \text{ Hz}$		
				21 GHz~45 GHz	$U=0.5 \text{ Hz}$		
		Frequency Span		10 kHz~1 MHz	$U_{\text{rel}}=0.022\%$		
				1 MHz~17 MHz	$U_{\text{rel}}=0.02\%$		
				17 MHz	$U_{\text{rel}}=0.0041\%$		
				17 MHz~100 MHz	$U_{\text{rel}}=0.02\%$		
100 MHz~3590 MHz	$U_{\text{rel}}=0.0082\%$						



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				3.59 GHz~26.5 GHz	$U_{rel}=0.023\%$		
		Resolution Bandwidth Switching Uncertainty		-1 dB~1 dB,100 Hz~300 Hz	$U=0.0058$ dB		
				-1 dB~1 dB,300 Hz~8 MHz	$U=0.002$ dB		
				-1 dB~1 dB,8 MHz~40 MHz	$U=0.0058$ dB		
				-1 dB~1 dB,40 MHz~80 MHz	$U=0.0058$ dB		
			Resolution Bandwidth		30 Hz	$U_{rel}=6.0\%$	
				100 Hz	$U_{rel}=0.65\%$		
				300 Hz	$U_{rel}=0.31\%$		
				1 kHz	$U_{rel}=0.29\%$		
				3 kHz	$U_{rel}=0.33\%$		
				9 kHz	$U_{rel}=1.7\%$		
				10 kHz	$U_{rel}=0.11\%$		
				30 kHz	$U_{rel}=0.033\%$		
				100 kHz	$U_{rel}=0.057\%$		
				120 kHz	$U_{rel}=1.8\%$		
				300 kHz	$U_{rel}=0.02\%$		
			1 MHz	$U_{rel}=0.01\%$			
			2 MHz	$U_{rel}=1.1\%$			




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				3 MHz	$U_{rel}=0.19\%$		
				5 MHz	$U_{rel}=1.3\%$		
				10 MHz	$U_{rel}=0.06\%$		
				20 MHz	$U_{rel}=0.03\%$		
				50 MHz	$U_{rel}=0.46\%$		
		Residual FM		0.01 Hz~250 Hz, Center Frequency 7 MHz	$U=0.10$ Hz		
				0.01 Hz~250 Hz, Center Frequency 500 MHz	$U=0.9$ Hz		
				0.01 Hz~250 Hz, Center Frequency 678.5 MHz	$U=0.48$ Hz		
				0.01 Hz~250 Hz, Center Frequency 1000 MHz	$U=0.19$ Hz		
				0.01 Hz~250 Hz, Center Frequency 1678.5 MHz~2200 MHz	$U=0.94$ Hz		
		Fast Time Domain Amplitude		0.01 Hz~250 Hz, Center Frequency 2900 MHz	$U=5.9$ Hz		
				1 μ s~1 s	$U_{rel}=0.029\%$		
		Noise Sidebands		-80 dBc/Hz~-170 dBc/Hz, Freq Deviation:98 Hz~100 Hz, Carrier: 500MHz~1 GHz	$U=0.48$ dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-80 dBc/Hz~-155 dBc/Hz , 100 Hz~10 kHz Deviation,Center Frequency: 500 MHz~1 GHz	U=0.12 dB		
				-80 dBc/Hz~-155 dBc/Hz , 10 kHz~30 kHz Deviation,Center Frequency: 500 MHz~1 GHz	U=0.05 dB		
				-80 dBc/Hz~-155 dBc/Hz , Deviation:30 kHz~1 MHz , Center Frequency: 500 MHz~1 GHz	U=0.31 dB		
				-80 dBc/Hz~-170 dBc/Hz , 1 MHz~10 MHz Deviation, Center Frequency: 500 MHz~1 GHz	U=0.56 dB		
				-80 dBc/Hz~-120 dBc/Hz, Deviation:100 kHz ~ 999 kHz,Center Frequency: 2 GHz	U=0.08 dB		
				-80 dBc/Hz~-140 dBc/Hz, Deviation:999 kHz ~ 1 MHz ,Center Frequency: 2 GHz	U=0.09 dB		
		Sweep Time		0.1 ms ~ 20 ms	U _{rel} =0.26 %		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				20 ms~100 s	$U_{rel}=0.04\%$		
		Reference Level		0 dBm~-10 dBm, 50MHz	$U=0.019$ dB		
				-10 dBm~-50 dBm, 50MHz	$U=0.025$ dB		
				-50 dBm~-80 dBm, 50MHz	$U=0.095$ dB		
				-80 dBm~-90 dBm, 50MHz	$U=0.20$ dB		
			Displayed Average Noise Level		-151 dBm~-90dBm, 20 Hz~10 kHz, RBW 1 Hz	$U=0.40$ dB	
				-156 dBm~-90dBm, 10 kHz~100 kHz, RBW 1 Hz	$U=0.33$ dB		
				-162 dBm~-90dBm, 100 kHz~200 kHz, RBW 1 Hz	$U=0.33$ dB		
				-168 dBm~-90dBm, 200 kHz~500 kHz, RBW 1 Hz	$U=0.33$ dB		
				-172 dBm~-90dBm, 500 kHz~500 MHz, RBW 1 Hz	$U=0.33$ dB		
				-170 dBm~-90dBm, 500 MHz~2 GHz, RBW 1 Hz	$U=0.33$ dB		



No. CNAS L0640

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-167 dBm~-90dBm, 2 GHz~6.6 GHz, RBW 1 Hz	U=0.33 dB		
				-166 dBm~-90dBm, 6.6 GHz~8.4 GHz, RBW 1 Hz	U=0.33 dB		
				-165 dBm~-90dBm, 8.4 GHz~13.6 GHz, RBW 1 Hz	U=0.33 dB		
				-164 dBm~-90dBm, 13.6 GHz~20 GHz, RBW 1 Hz	U=0.33 dB		
				-162 dBm~-90dBm, 20 GHz~26.5 GHz, RBW 1 Hz	U=0.33 dB		
				-161 dBm~-90dBm, 26.5 GHz~32 GHz, RBW 1 Hz	U=0.33 dB		
				-160 dBm~-90dBm, 32 GHz~34 GHz, RBW 1 Hz	U=0.33 dB		
				-158 dBm~-90dBm, 34 GHz~40 GHz, RBW 1 Hz	U=0.33 dB		
				-157 dBm~-90dBm, 40 GHz~46 GHz, RBW 1 Hz	U=0.33 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Residual Response	ilac-M	-155 dBm~-90dBm, 46 GHz~50 GHz, RBW 1 Hz	U=0.33 dB		
				-80 dB~-160 dB ,150 kHz~1.25 MHz	U=0.87 dB		
				-80 dB~-160 dB ,1.25 MHz~50 MHz	U=0.44 dB		
				-80 dB~-160 dB ,50 MHz~3.9 GHz	U=0.33 dB		
				-80 dB~-160 dB ,3.9 GHz~35 GHz	U=0.53 dB		
				-80 dB~-160 dB ,35 GHz~45 GHz	U=1.2 dB		
		Input Attenuation Switching Uncertainty	ilac-M	0 dB~6 dB,50 MHz	U=0.019 dB		
				6 dB~20 dB,50 MHz	U=0.023 dB		
				20 dB~70 dB,50 MHz	U=0.028 dB		
				70 dB~75 dB,50 MHz	U=0.043 dB		
				0 dB~30 dB,128 MHz	U=0.019 dB		
				30 dB~60 dB,128 MHz	U=0.029 dB		
				60 dB~70 dB,128 MHz	U=0.035 dB		
				70 dB~75 dB,128 MHz	U=0.043 dB		
		Other Input Related	ilac-M	10 dB~70 dB,2.9 GHz	U=0.18 dB		
				-130 dBc~-40 dBc,1 MHz~10 MHz	U=0.10 dB		




No. CNAS L0640

第 51 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Spurious Responses		-130 dBc~-40 dBc,11 MHz~3.1 GHz	U=0.17 dB		
				-130 dBc~-40 dBc,3.1 GHz~7 GHz	U=0.19 dB		
				-130 dBc~-40 dBc,7 GHz~13 GHz	U=0.23 dB		
				-130 dBc~-40 dBc,13 GHz~39.1 GHz	U=0.27 dB		
				-130 dBc~-40 dBc,39 GHz~50 GHz	U=0.52 dB		
		Second Harmonic Distortion		-120 dBc~-40 dBc,9 MHz~28 MHz	U=0.42dB		
				-120 dBc~-40 dBc,28 MHz~3490 MHz	U=0.18 dB		
				-120 dBc~-40 dBc,3490 MHz~15.5 GHz	U=0.27 dB		
				-120 dBc~-40 dBc,15.5 GHz~31 GHz	U=0.32 dB		
		Third Order Intermodulation Distortion		-100 dBc~-40 dBc,10 MHz~28 MHz	U=0.18 dB		
				-100 dBc~-40 dBc,28 MHz~3 GHz	U=0.062 dB		
				-100 dBc~-40 dBc,3 GHz~4 GHz	U=0.075 dB		
				-100 dBc~-40 dBc,4 GHz~40 GHz	U=0.061 dB		
				-100 dBc~-40 dBc,40 GHz~46 GHz	U=1.5 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Gain Compression		-100 dBc~-40 dBc,46 GHz~50 GHz	U=1.6 dB		
				0 dB~1 dB,50 MHz~2 GHz	U=0.04 dB		
				0 dB~1 dB,2 GHz~14 GHz	U=0.12 dB		
				0 dB~1 dB,14 GHz~ 19.9 GHz	U=0.14 dB		
		Frequency Response		-5 dB~5 dB,3 Hz~20 Hz	U=0.062 dB		
				-5 dB~5 dB,20 Hz~9 kHz	U=0.056 dB		
				-5 dB~5 dB,9 kHz~300 kHz	U=0.062 dB		
				-5 dB~5 dB,0.300 MHz~3550 MHz	U=0.051 dB		
				-5 dB~5 dB,3.55 GHz~ 18 GHz	U=0.064 dB		
				-5 dB~5 dB,18 GHz~ 26.5 GHz	U=0.095 dB		
				-5 dB~5 dB,26 GHz~ 39.5 GHz	U=0.18 dB		
				-5 dB~5 dB,39.5 GHz~ 50 GHz	U=0.24 dB		
		Absolute Amplitude		-83 dBm~-80 dBm,50 MHz	U=0.076 dB		
				-80 dBm~-10 dBm,50 MHz	U=0.045 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-10 dBm~-1 dBm,50 MHz	U=0.23 dB		
				-30 dBm,128 MHz	U=0.076 dB		
				-20 dBm ~ -40 dBm,64 MHz	U=0.06 dB		
				-10 dBm ~ -20 dBm,64 MHz	U=0.093 dB		
		Display Scale Fidelity		-90 dBm ~ -75 dBm,5 MHz	U=0.043 dB		
				-75 dBm~-65 dBm,5 MHz	U=0.036 dB		
				-65 dBm~0 dBm,5 MHz	U= 0.017 dB		
				-90 dBm~-85 dBm, 50 MHz	U=0.021 dB		
				-85 dBm~-45 dBm, 50 MHz	U=0.009 dB		
				-45 dBm~-11 dBm, 50 MHz	U=0.006 dB		
				-11 dBm~0 dBm, 50 MHz	U=0.018 dB		
				-90 dBm~-85 dBm,128 MHz	U=0.074 dB		
				-85 dBm~-80 dBm,128 MHz	U=0.025 dB		
				-80 dBm~0 dBm,128 MHz	U=0.021 dB		
		Input Port Return Loss		6 dB~60 dB,10 MHz~3.5 GHz	U=0.030 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-M	6 dB~60 dB,3.5 GHz~6 GHz	U=0.053 dB		
				6 dB~60 dB,6 GHz~9 GHz	U=0.067 dB		
				6 dB~60 dB,9 GHz~20 GHz	U=0.042 dB		
				6 dB~60 dB,20 GHz~30 GHz	U=0.060 dB		
				6 dB~60 dB ,30 GHz~35 GHz	U=0.072 dB		
				6 dB~60 dB,35 GHz~40 GHz	U=0.091 dB		
		Tracking GEN Level		0 dBm~-40 dBm ,128 MHz	U=0.088 dB		
				-10 dBm, 64 MHz	U=0.14 dB		
		Tracking GEN Frequency Response		-3 dB~3 dB ,9 kHz~80 kHz	U=0.089 dB		
				-3 dB~3 dB ,80 kHz~2 GHz	U=0.094 dB		
				-3 dB~3 dB ,2 GHz~3 GHz	U=0.11 dB		
				-3 dB~3 dB,3 GHz~5 GHz	U=0.11 dB		
				-3 dB~3 dB,5 GHz~7 GHz	U=0.076 dB		
		Tracking Gen LO		-3 dB~3 dB,7 GHz~7.5 GHz	U=0.14 dB		
				-16 dBm~-90 dBm ,9 kHz~2.9 GHz	U=0.15 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Feedthrough Amplitude		-16 dBm~-90 dBm ,2.9 GHz~3 GHz	U=3.0 dB		
		Tracking GEN		-17 dBm~-18 dBm ,50 MHz	U=0.10 dB		
		Absolute Amplitude, Vernier		-18 dBm~-26 dBm ,50 MHz	U=0.008 dB		
		Tracking GEN		-146 dBc~-15 dBc ,9 kHz~850 MHz	U=0.36 dB		
		Harmonic Distortion		-146 dBc~-15 dBc ,850 MHz~1.5 GHz	U=2.3 dB		
		Tracking GEN Non-Harmonic Distortion		-146 dBc~-23 dBc ,9 kHz~1.8 GHz	U=0.36 dB		
				-146 dBc~-23 dBc ,1.8 GHz~3 GHz	U=2.6 dB		
		Audio - DC Offset		4 V , 4 V Range	U=0.1 mV		
				0.4 V, 0.4 V Range	U=0.057 mV		
		Audio - Residual Noise		0.1 μV~5 mV, 4 V Range	U=1.6 μV		
				0.1 μV~1 mV, 0.4 V Range	U=0.58 μV		
		Audio - DC Voltage Measurement		4 V, 4 V Range	U=1.6 mV		
				0.4 V, 4 V Range	U=0.36 mV		
				-4 V, 4 V Range	U=1.6 mV		
				-0.4 V, 4 V Range	U=0.36 mV		
				400 mV, 0.4 V Range	U=0.077 mV		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
				100 mV, 0.4 V Range	U=0.084 mV			
				400 mV, 0.4 V Range	U=0.077 mV			
				100 mV, 0.4 V Range	U=0.084 mV			
		Audio - AC Voltage Measurement Uncertainty		3 V,20 Hz,4 V Range	U=0.71 mV			
					3 V,1 kHz ,4 V Range	U=0.48 mV		
					3 V,20 kHz ,4 V Range	U=0.72 mV		
					3 V,100 kHz,4 V Range	U=3.0 mV		
					2 V,300 kHz,4 V Range	U=8.1 mV		
					2 V,1 MHz ,4 V Range	U=25 mV		
					0.3 V,1 kHz,4 V Range	U=0.085mV		
					0.3 V,300 kHz,4 V Range	U=1.1 mV		
					0.3 V,20 Hz,0.4 V Range	U=0.071 mV		
					0.3 V,1 kHz ,0.4 V Range	U=0.048 mV		
					0.3 V,20 kHz ,0.4 V Range	U=0.072mV		
					0.3 V,100 kHz,0.4 V Range	U=0.30 mV		
					0.3 V,300 kHz ,0.4 V Range	U=1.1 mV		
					0.3 V,1 MHz,0.4 V Range	U=3.5 mV		
					0.1 V,1 kHz,0.4 V Range	U=0.012 mV		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.1 V, 300 kHz, 0.4 V Range	U=0.35 mV		
		Audio - Inherent Distortion		-60 dB, 100 Hz, 3 V	U=0.35 dB		
				-60 dB, 1 kHz~20 kHz, 3 V	U=0.10 dB		
				-60 dB, 80 kHz, 3 V	U=0.13 dB		
				-60 dB, 20 kHz~80 kHz, 0.6 V	U=0.18 dB		
				-60 dB, 20 kHz, 0.3 V	U=0.13 dB		
				-60 dB, 80 kHz, 0.3 V	U=0.31 dB		
			Audio - Distortion Measurement		-50 dB~-20 dB, 100 Hz~50 kHz	U=0.14 dB	
		Audio - SINAD Measurement		-40 dB~20 dB, 1 kHz	U=0.14 dB		
		Audio - Filter Frequency Response		0 dB~0.05 dB, 50 Hz HPF, Frequency: 200 Hz	U=0.006 dB		
				0 dB~1 dB, 50 Hz HPF, Frequency: 20 Hz	U=0.17 dB		
				0 dB~0.05 dB, 300 Hz HPF, Frequency: 1 kHz	U=0.006 dB		
				0 dB~1 dB, 300 Hz HPF, Frequency: 100 Hz	U=0.11 dB		
				0 dB~0.05 dB, 3 kHz LPF, Frequency: 1 kHz	U=0.006 dB		
				0 dB~1 dB, 3 kHz LPF, Frequency: 6 kHz	U=0.12 dB		




No. CNAS L0640

第 58 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				300 kHz,0.3 V	U=0.000057 kHz		
				990 kHz,0.3 V	U=0.000024 kHz		
				300 kHz,0.1 V	U=0.000012 kHz		
				1.01 kHz,0.1 V	U=0.00000057 kHz		
				20 Hz,0.1 V	U=0.001 kHz		
		AM - Modulation Depth		5%, Carrier Frequency:12.5 MHz,Mod Rate: 1 kHz	U=0.001%		
				50%, Carrier Frequency:12.5 MHz,Mod Rate: 1 kHz	U=0.006%		
				99%, Carrier Frequency:12.5 MHz,Mod Rate: 1 kHz	U=0.006%		
				5%,Carrier Frequency:12.5 MHz,Mod Rate: 90 Hz	U=0.001%		
				50%, Carrier Frequency:12.5 MHz,Mod Rate: 90 Hz	U=0.006%		
				99%, Carrier Frequency:12.5 MHz,Mod Rate: 90 Hz	U=0.006%		
				5%, Carrier Frequency:12.5 MHz,Mod Rate: 150 Hz	U=0.001%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date						
				50%, Carrier Frequency:12.5 MHz,Mod Rate: 150 Hz	U=0.006%								
				99%, Carrier Frequency:12.5 MHz,Mod Rate: 150 Hz	U=0.006%								
				5%, Carrier Frequency:100 kHz,Mod Rate: 1 kHz	U=0.001%								
				50%, Carrier Frequency:100 kHz,Mod Rate: 1 kHz	U=0.006%								
				99%, Carrier Frequency:100 kHz,Mod Rate: 1 kHz	U=0.006%								
				47%~53%, Carrier Frequency:12.5 MHz,Mod Rate: 50 Hz~100 kHz	U=0.01%								
		AM - Flatness						47%~53%, Carrier Frequency:100 kHz,Mod Rate: 50 Hz~10 kHz	U=0.01%				
								AM - Inherent Distortion				-50.5 dB,Carrier Frequency:12.5 MHz,Mod Rate:50 Hz ,99%	U=0.72 dB
												-50.5 dB,Carrier Frequency:12.5 MHz,Mod Rate:1 kHz ,99%	U=0.39 dB



No. CNAS L0640


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-50.5 dB,Carrier Frequency:12.5 MHz,Mod Rate:10 kHz ,99 %	U=0.15 dB		
				-50.5 dB,Carrier Frequency:12.5 MHz,Mod Rate:50 kHz ,99 %	U=0.24 dB		
				-50.5 dB,Carrier Frequency:100 kHz,Mod Rate:50 Hz ,99 %	U=0.93 dB		
				-50.5 dB,Carrier Frequency:100 kHz,Mod Rate:1 kHz ,5 %	U=3.0 dB		
				-50.5 dB,Carrier Frequency:100 kHz,Mod Rate:1 kHz ,99 %	U=0.59 dB		
				-50.5 dB,Carrier Frequency:100 kHz,Mod Rate:10 kHz ,99 %	U=0.55 dB		
		AM - FM Rejection		0.01%~0.2%,Carrier Frequency: 100 kHz~ 600 MHz	U=0.006%		
		AM - Residual AM		0.01%~0.01%,Carrier Frequency: 300 kHz~3 GHz	U=0.0006%		
		FM - Frequency Deviation		5.52 kHz,Carrier Frequency: 200 kHz,Mod Rate: 1 kHz	U=0.001 kHz		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				55.2 kHz, Carrier Frequency: 200 kHz, Mod Rate: 10 kHz	U=0.006 kHz		
				50 kHz, Carrier Frequency: 200 kHz, Mod Rate: 50 Hz	U=0.006 kHz		
				5.52 kHz, Carrier Frequency: 12.0 MHz, Mod Rate: 1 kHz	U=0.001 kHz		
				55.2 kHz, Carrier Frequency: 12.0 MHz, Mod Rate: 10 kHz	U=0.006 kHz		
				552 kHz, Carrier Frequency: 12.0 MHz, Mod Rate: 100 kHz	U=0.06 kHz		
				481 kHz, Carrier Frequency: 12.0 MHz, Mod Rate: 200 kHz	U=0.6 kHz		
		FM - Inherent FM Distortion		-50.5 dB, Carrier Frequency: 200 kHz, Mod Rate: 1 kHz Deviation: 50 kHz	U=0.13 dB		
				-50.5 dB, Carrier Frequency: 200 kHz, Mod Rate: 5 kHz Deviation: 50 kHz	U=0.12 dB		
				-50.5 dB, Carrier Frequency: 200 kHz, Mod Rate: 10 kHz Deviation: 50 kHz	U=0.14 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:1 kHz Deviation: 50 kHz	U=1.5 dB		
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:10 kHz Deviation: 50 kHz	U=1.5 dB		
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:1 kHz Deviation: 500 kHz	U=0.19 dB		
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:20 kHz Deviation: 500 kHz	U=0.30 dB		
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:50 kHz Deviation: 500 kHz	U=0.18 dB		
				-50.5 dB,Carrier Frequency:12.0 MHz,Mod Rate:100 kHz Deviation: 500 kHz	U=0.13 dB		
		FM - AM Rejection		0.1 Hz~20 Hz,Carrier Frequency:100 kHz~ 12.5 MHz	U=0.057 Hz		
		FM - Residual FM		0.1 Hz~1 Hz,Carrier Frequency:300 kHz	U=0.058 Hz		
				0.1 Hz~1 Hz,Carrier Frequency:600 MHz	U=0.066 Hz		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.1 Hz~2.5 Hz,Carrier Frequency:3 GHz	U=0.086 Hz		
		PM - Phase Deviation		5.52 rad,Carrier Frequency:12 MHz, Mod Rate:100 kHz	U=0.00057 rad		
				20 rad,Carrier Frequency:12 MHz, Mod Rate:20 kHz	U=0.0057 rad		
				400 rad,Carrier Frequency:12 MHz, Mod Rate:1 kHz	U=0.057 rad		
		PM - Inherent PM Distortion		-60 dB,500 rad,Mod Rate:1 kHz	U=0.23 dB		
				-60 dB,25 rad,Mod Rate:20 kHz	U=0.10 dB		
				-60 dB,10 rad,Mod Rate:50 kHz	U=0.09 dB		
				-60 dB,5 rad,Mod Rate:100 kHz	U=0.11 dB		
		PM - AM Rejection		0 rad~0.02 rad,Carrier Frequency:12.5 MHz	U=0.001 rad		
		Tracking GEN IQ Modulation Power Level		-3 dBm~ 3 dBm,+90 ° ,-90° ,Carrier Frequency: 1001 MHz	U=0.75 dB		
		Tracking GEN IQ Modulation Harmonic		-60 dBc~-30 dBc,+90 ° , Frequency: 1000 MHz	U=0.10 dB		
					-60 dBc~-30 dBc,+90 ° , Frequency 999 MHz	U=0.10 dB	




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-60 dBc~-30 dBc, -90° Frequency: 1000 MHz	U=0.10 dB		
				-60 dBc~-30 dBc,-90° Frequency: 999 MHz	U=0.10 dB		
		Tracking GEN AM Modulation		0 dB~3 dB,Carrier Frequency: 1 GHz	U=0.75 dB		
				0 dB~2 dB,Carrier Frequency:1001 MHz~ 999 MHz	U=0.085 dB		
		Tracking GEN FM Modulation		0 dB~3 dB,Carrier Frequency: 1 GHz	U=0.75 dB		
				0 dB~1 dB,Carrier Frequency:1001 MHz~ 999 MHz	U=0.085 dB		
				3~5,Resolution Bandwidth 10 Hz~300 Hz	U=0.028		
				3~5,Resolution Bandwidth 1 kHz	U=0.0069		
		IF Bandwidth - Shape Factor		3~5,Resolution Bandwidth 3 kHz	U=0.028		
				3~5,Resolution Bandwidth 10 kHz	U=0.01		
				3~5,Resolution Bandwidth 30 kHz	U=0.028		
				3~5,Resolution Bandwidth 100 kHz	U=0.01		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				3~5,Resolution Bandwidth 300 kHz、1 MHz	U=0.02		
				3~5,Resolution Bandwidth 3MHz、10 MHz	U=0.01		
				3~5,Resolution Bandwidth 20 MHz	U=0.34		
				3~5,Resolution Bandwidth 50 MHz	U=0.84		
		Calibrator Output Level		-30 dBm,128 MHz	U=0.076 dB		
				-10 dBm,300 MHz	U=0.11 dB		
				-20 dBm,50 MHz	U=0.18 dB		
				-40 dBm,120 MHz	U=0.12 dB		
				0 dBm,100 MHz	U=0.17 dB		
		Noise Density		-170 dBm/Hz~-120 dBm/Hz, 1.8 GHz ~ 42 GHz	U=0.57 dB		
		IF Frequency Response		-1 dB~1 dB, Carrier 1.825 GHz ,-30 dBm~0 dBm, Span 1 kHz ~ 10 MHz	U=0.011 dB		
				-1 dB~1 dB, Carrier 1.825 GHz~42.25 GHz, -30 dBm~0 dBm, Span 10 MHz~80 MHz	U=0.04 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-1 dB~1 dB, Carrier 1.825 GHz~42.25 GHz, 30 dBm~0 dBm, Span 80 MHz~255 MHz	U=0.08 dB		
				-1 dB~1 dB, Carrier 1.825 GHz ~ 5.95 GHz,-30 dBm~0 dBm, Span 255 MHz ~ 509 MHz	U=0.09 dB		
				-1 dB~1 dB, Carrier 1.825 GHz ~ 5.95 GHz,-30 dBm~0 dBm, Span 255 MHz ~ 510 MHz	U=0.11 dB		
		Power Bandwidth		-1 dB~1 dB,Carrier 50 MHz, -30 dBm~0 dBm, RBW 3 Hz~1 MHz	U=0.001 dB		
		Audio Distortion		0.01%~0.30%, Amplitude: 0.11 V~2.8 V, Frequency: 20 Hz~ 250 Hz	U=0.003%		
				0.01%~0.30%, Amplitude: 0.11 V~2.8 V, Frequency: 250 Hz~ 25 kHz	U=0.002%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.01%~0.30%, Amplitude: 0.11 V~2.8 V, Frequency: 25 kHz~ 75 kHz	$U=0.004\%$		
		Audio Amplitude		0.055 V ~ 0.79 V, 0.038 kHz	$U_{rel}=0.015\%$		
				0.055 V~0.79 V, 0.0381 kHz ~0.74 kHz	$U_{rel}=0.012\%$		
				0.055 V~0.79 V, 0.74 kHz~243.75 kHz	$U_{rel}=0.022\%$		
		Frequency Response of The IQ Path		-1 dB~1 dB, 0 dBm ,120 MHz~136 MHz	$U=0.097$ dB		
		RF Frequency Characteristic s		-5 dB~5 dB,5 MHz~ 49.999 MHz	$U=0.057$ dB		
				-5 dB~5 dB, 50 MHz	$U=0.035$ dB		
				-5 dB ~ 5 dB,50 MHz ~ 100 MHz	$U=0.059$ dB		
				-5 dB ~ 5 dB, 100.1 MHz ~3 GHz	$U=0.060$ dB		
				-5 dB ~ 5 dB, 3.01 GHz~6 GHz	$U=0.070$ dB		
				-5 dB ~ 5 dB, 6.01 GHz~ 12 GHz	$U=0.11$ dB		
				-5 dB ~ 5 dB, 12.01 GHz ~ 16 GHz	$U=0.13$ dB		
			-5 dB ~ 5 dB,16.01 GHz ~ 26.5 GHz	$U=0.18$ dB			




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pulse Digitization Uncertainty	ilac-M	-5 dB~5 dB, RBW1 MHz,log scale	U=0.04 dB		
				-5 dB~5 dB, RBW 2 MHz ,log scale	U=0.09 dB		
				-15%~15%,RBW 1 MHz , linear scale	U=0.31%		
				-15%~15%,RBW 2 MHz , linear scale	U=0.5%		
		Delayed Sweep		1000 μs~65000 μs	U=0.0035 μs		
		Gate Delay And Gate Length		Min Gate Delay 2 s~6 s	U=0.036 μs		
				MAX Gate Delay 2 s~6 s	U=0.036 μs		
				Gate Length 1 μs	U=0.036 μs		
				Gate Length 65 ms	U=0.0067 ms		
		Fast Sweep Time		300 MHz	U=0.00001 MHz		
		First LO Output Amplitude		0 dBm~20 dBm,3 GHz~6.8 GHz	U=0.15 dB		
7	*Digital Mobile Test Set	RF GEN Level	Digital Mobile Test Set Test Procedure CAL-CHA004-01	-15 dBm~-39 dBm, 850 MHz~2650 MHz	U=0.19 dB		
				-39 dBm~-54 dBm, 850 MHz~2650 MHz	U=0.20 dB		
				-54 dBm~-69 dBm, 850 MHz~2650 MHz	U=0.21 dB		
				-69 dBm~-89 dBm, 850 MHz~2650 MHz	U=0.22 dB		




在线扫码获取验证

No. CNAS L0640

第 70 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-89 dBm~-110 dBm, 850 MHz~2650 MHz	U=0.23 dB		
				-20 dBc~-90 dBc ,810 MHz~960 MHz	U=0.6 dB		
		RF GEN Harmonics Distortion		-20 dBc~-90 dBc ,960 MHz~1700 MHz	U=0.9 dB		
				-20 dBc~-90 dBc ,1700 MHz~1750 MHz	U=1.1 dB		
				-20 dBc~-90 dBc ,1750 MHz~1990 MHz	U=1.9 dB		
		RF GEN Spurious		-90 dBc~-105 dBc ,996 MHz~1990MHz	U=1.2 dB		
		RF Generator Residual FM		0.1 Hz~7 Hz,Carrier Frequency: 800 MHz~2000 MHz,Power: -20 dBm	U=1.6 Hz		
		RF Generator FM		0.5 kHz~20 kHz,Carrier 800 MHz~2000 MHz,Mod Rate 1 kHz	U _{rel} =1.0%		
		RF Generator FM Distortion		0.01%~0.5%,Deviation 0.5 kHz~20 kHz, Carrier Frequency: 800 MHz~2000 MHz	U=0.06%		
		TDMA GEN Digital MOD EVM		0.01%~3%,Carrier Frequency: 790 MHz~1667 MHz	U=0.57%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TDMA GEN Digital Modulation Magnitude Error		0.01%~2%, Carrier Frequency: 790 MHz~ 1667 MHz	U=0.38%		
		TDMA GEN Digital Modulation Phase Error		0.01° ~2°, Carrier Frequency: 790 MHz~ 1667 MHz	U=0.3°		
		TDMA GEN Digital Modulation Frequency Error		0.01 Hz~4 Hz , Carrier Frequency: 790 MHz~ 1667 MHz	U=0.05 Hz		
		TDMA GEN MOD IQ Origin Offset		-40 dB~-30 dB, Carrier Frequency: 790 MHz~ 1667 MHz	U=0.5 dB		
		RF Analyzer Level		-14 dBm~8 dBm ,835 MHz~1880 MHz	U=0.10 dB		
		TDMA Analyzer Digital MOD EVM		-1% ~1% , Carrier Frequency: 900 MHz~ 1900 MHz	U=0.57%		
		TDMA Analyzer Digital MOD Magnitude Error		-7% ~7% , Carrier Frequency: 900 MHz~ 1900 MHz	U=0.38%		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TDMA Analyzer Digital MOD Phase Error	ilac-MRA CHINA NATIONAL ACCREDITATION FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	-4 ° ~ 4 ° , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.3 °		
		TDMA Analyzer Digital MOD Frequency Error		10 Hz ~ 10 Hz , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.1 Hz		
		TDMA Analyzer Digital MOD Origin Offset		-5 dB ~ 5 dB , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.5 dB		
		TDMA Analyzer Digital MOD Residual EVM		0.01% ~ 1% , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.48%		
		TDMA Analyzer Digital MOD Residual Magnitude Error		0.01% ~ 0.7% , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.33%		
		TDMA Analyzer Digital MOD Residual Phase Error		0.01 ° ~ 0.4 ° , Carrier Frequency: 900 MHz ~ 1900 MHz	U=0.24 °		



No. CNAS L0640

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TDMA Analyzer Digital MOD IQ Residual Origin Offset		-60 dB~-50 dB , Carrier Frequency: 900 MHz~1900 MHz	U=0.31 dB		
		TDMA IQ Tuning		-1.0 dB~1.0 dB,Carrier 850 MHz~2000 MHz	U=0.27dB		
		TDMA Analyzer Adjacent Power Accuracy		-1.2 dB~1.2 dB,Carrier 900 MHz~1900 MHz	U=0.27 dB		
		GSM Analyzer IQ Tuning		-1.0 dB~1.0 dB,Carrier 825 MHz~1950 MHz	U=0.27 dB		
		GSM GEN Modulation Accuracy Amplitude		-0.3 dB~0.3 dB ,Carrier 940 MHz~1840 MHz	U=0.12 dB		
		GSM GEN MOD Accuracy Peak Phase Error		-4.0 ° ~4.0 ° , Carrier 940 MHz~1840 MHz	U=1.4 °		
		GSM GEN MOD Accuracy RMS Phase Error		-1.0° ~1.0° , Carrier 940 MHz~1840 MHz	U=0.30°		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		GSM GEN MOD Accuracy Frequency Error		-9.4 Hz~9.4 Hz,Carrier 940 MHz~1840 MHz	U=1 Hz		
		GSM Output RF Spectrum Accuracy		-1.5 dB~1.5 dB,Carrier 900 MHz~1800 MHz	U=0.22 dB		
		GSM AN MOD Accuracy Peak Phase Error		-4.0 ° ~4.0 ° , Carrier 835 MHz~1880 MHz	U=1.4 °		
		GSM AN MOD Accuracy RMS Phase Error		-1.0 ° ~1.0 ° , Carrier 835 MHz~ 1880 MHz	U=0.30 °		
		GSM AN MOD Accuracy Frequency Error		-12 Hz~12 Hz,Carrier 835 MHz~1880 MHz	U=4.8 Hz		
		CDMA/1xEV GEN Digital Power		-26.94 dBm~-16.94 dBm,Carrier 870 MHz~1930 MHz	U=0.16 dB		
		CDMA GEN Modulation Accuracy RHO		0.9~1,Carrier 800 MHz~2150 MHz	U=0.0019		



No. CNAS L0640

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		CDMA GEN Modulation Accuracy EVM		0~10% ,Carrier 800 MHz~2150 MHz	U=1.0%		
		CDMA AN Modulation Accuracy Frequency Error		-15 Hz~15 Hz,Carrier 900 MHz~1900 MHz	U=9.5 Hz		
		CDMA AN Average Power		-9.0 dBm~3.0 dBm , Carrier 850 MHz~1850 MHz	U=0.1 dB		
		CDMA AN Tuned Power		-56 dBm~-1.0 dBm , Carrier 834 MHz~1960 MHz	U=0.14 dB		
		CDMA AN Modulation Accuracy RHO		0.9~1,Carrier 900 MHz~1900 MHz	U=0.0003		
		CDMA AN Modulation Residual RHO		0.9~1,Carrier 900 MHz~1900 MHz	U=0.0003		
		1xEV GEN Modulation Accuracy RHO		0.9~1,Carrier 870 MHz~2150 MHz	U=0.0019		
		1xEV GEN Modulation Accuracy EVM		0.1%~10%, Carrier 870 MHz~2150 MHz	U=1.0%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		CDMA/1xEV GEN Residual Frequency Error		-15 Hz~15 Hz,Carrier 900 MHz~1900 MHz	U=0.59 Hz		
		1xEV AN Modulation Accuracy Residual RHO		0.9~1,Carrier 900 MHz~1900 MHz	U=0.0003		
		1xEV AN Modulation Accuracy Residual EVM		4% , Carrier 900 MHz~1900 MHz	U=0.38%		
		1xEV AN Tuned Power		-56.0 dBm~-1.0 dBm,Carrier 834 MHz~1960 MHz	U=0.14 dB		
		WCDMA GEN Digital Power		-26.94 dBm~-16.94 dBm,Carrier 870 MHz~2760 MHz	U=0.1 dB		
		WCDMA GEN Modulation Accuracy EVM		0.1%~10% , Carrier 800 MHz~2655 MHz	U=1.0%		
		WCDMA GEN Modulation Accuracy IQ Offset		-35 dBc~-20 dBc,Carrier 800 MHz~2655 MHz	U=0.5 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		WCDMA AN Channel Power		-56.0 dBm ~ -1.0 dBm, Carrier 834 MHz ~ 2530 MHz	U=0.14 dB		
		WCDMA AN Level		-10.0 dBm ~ 7.0 dBm, Carrier 850 MHz ~ 2530 MHz	U=0.07 dB		
		WCDMA AN Modulation Accuracy Residual EVM		0.01% ~ 2.20% , Carrier 900 MHz ~ 2530 MHz	U=0.48%		
		WCDMA ACLR		-0.8 dB ~ 0.8 dB, Carrier 814 MHz ~ 2530 MHz	U=0.15 dB		
		WCDMA AN MOD Residual Frequency Error		-5 Hz ~ 5 Hz, Carrier 900 MHz ~ 2530 MHz	U=0.69 Hz		
		AA GEN Modulation Accuracy RMS Phase Error		-1.0 ° ~ 1.0 ° , Carrier 478.2 MHz ~ 2687.6 MHz	U=0.30 °		
		AA GEN Modulation Accuracy Peak Phase Error		-4.0 ° ~ 4.0 ° , Carrier 478.2 MHz ~ 2687.6 MHz	U=1.4 °		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AA GEN Modulation Accuracy EVM		0.1%~4%, Carrier 478.2 MHz~2687.6 MHz	U=0.5%		
		AA GEN MOD Accuracy Origin Offset		-39 dB~-30 dB, Carrier 478.2 MHz~2687.6 MHz	U=1.6 dB		
		AA GEN Spurious		-80 dBc~-50 dBc, Carrier 955.884 MHz~2010 MHz	U=0.9 dB		
		AA GEN Level		-30 dBm~-80 dBm, Carrier 460.6 MHz~2687.6 MHz	U=0.29 dB		
		AA AN Frequency Error		-12 Hz~12 Hz, Carrier 450.6 MHz~2524.8 MHz	U=1.3 Hz		
		AA AN Level Accuracy		-15 dBm~8 dBm, Carrier 450.6 MHz~2140.8 MHz	U=0.25 dB		
		TD-SCDMA AN Channel Power		-56.0 dBm~-1.0 dBm, Carrier 1880.8 MHz~2024.2 MHz	U=0.14 dB		
		TD-SCDMA ACLR Accuracy		-0.8 dB~0.8 dB, Carrier 1880.8 MHz~2017 MHz	U=0.15 dB		
		TD-SCDMA GEN Modulation Accuracy EVM		0.1%~2%, Carrier 1880.8 MHz~2017 MHz	U=0.48%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TD-SCDMA AN Modulation Accuracy Frequency Error		-5 Hz~10 Hz,Carrier 1880.8 MHz~2017 MHz	U=0.69 Hz		
		TD-SCDMA GEN Digital Power		-26.94 dBm~-16.94 dBm,Carrier 1880 MHz~2024.2 MHz	U=0.06 dB		
		TD-SCDMA GEN Modulation Accuracy EVM		0.1%~10%, Carrier 411 MHz~1800 MHz	U=1.0%		
		TD-SCDMA GEN Modulation Accuracy IQ Offset		-30 dBc~-20 dBc,Carrier 411 MHz~1800 MHz	U=0.5 dB		
		EGPRS GEN Modulation Accuracy EVM		0.1%~4.0%,Carrier 460 MHz~1880 MHz	U=0.48%		
		EGPRS GEN Modulation Accuracy Frequency Error		-9 Hz~9 Hz,Carrier 460 MHz~1880 MHz	U=1 Hz		
		EGPRS GEN Modulation Accuracy Origin Offset		0 dB~50 dB,Carrier 460 MHz~1880 MHz	U=0.48 dB		




在线扫码获取验证

No. CNAS L0640

第 80 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		EGPRS AN Modulation Accuracy RMS EVM		0.1% ~ 1.0%, Carrier 450 MHz ~ 1990 MHz	U=0.48 %		
		EGPRS AN Modulation Accuracy PEAK EVM		4.0% ~ 4.0%, Carrier 450 MHz ~ 1990 MHz	U=0.48%		
		EGPRS AN Modulation Accuracy Frequency Error		-10 Hz ~ 10 Hz, Carrier 450 MHz ~ 1990 MHz	U=1 Hz		
		EGPRS AN Modulation Accuracy Origin Offset		-1.5 dB ~ 1.5 dB, Carrier 450 MHz ~ 1990 MHz	U=0.5 dB		
		AF GEN Level		100 mV ~ 6000 mV, 100 Hz ~ 5 kHz	U _{rel} =0.1%		
		AF GEN Distortion		0.001% ~ 0.1%, 200 mV ~ 6 V, 100 Hz ~ 5 kHz	U=0.012%		
		AF AN Level		0.5 V ~ 5 V, 200 Hz ~ 500 Hz	U _{rel} =0.02%		
				0.5 V ~ 5 V, 0.5 kHz ~ 10 kHz	U _{rel} =0.03%		
				50 mV, 200 Hz ~ 500 Hz	U _{rel} =0.03%		
		AF AN Distortion		50 mV, 0.5 kHz ~ 10 kHz	U _{rel} =0.04%		
				4% ~ 11%, 2 kHz ~ 3 kHz , 50 mV	U=0.11%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				4%~11%,2 kHz~3 kHz 100 mV	$U=0.19\%$		
		AF AN SINAD		19 dB~27 dB,50 mV~ 100 mV,2 kHz~3 kHz	$U=0.5\text{ dB}$		
		Time-base Frequency		10 MHz	$U_{rel}=1\times 10^{-11}$		
8	*Wireless Communications Test Set	TX/RX Port Isolation	Wireless Communications Test Set Test Procedure CAL- CHA004-02	-140 dB~+40 dB,300 MHz~6000 MHz	$U=0.6\text{ dB}$		
		Reference Frequency Amplitude		1.4 V~20 V,1 MHz~10 MHz	$U=0.25\text{ V}$		
		Reference Frequency Accuracy		10 MHz	$U=0.0016\text{ Hz}$		
		Audio Generator Distortion		0.001%~0.05%, 500 mV~5000 mV,20 Hz~ 10 kHz	$U=0.016\%$		
		Audio Generator Level		0.1 V ~5 V,1 kHz ~10 kHz	$U=0.1\text{ mV}$		
		Audio Generator Output Impedance		50 Ω ~75 Ω ,1 kHz ~ 10 kHz	$U=0.06\ \Omega$		
		Audio Generator Max Output Current		5 mA~20 mA, 1kHz~ 10kHz	$U=0.14\text{ mA}$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Audio Analyzer Frequency Response		1.96V~2V,50 Hz ~10 kHz	U=2.0 mV		
		Audio Analyzer Inherent Distortion		0.01%~0.05%,0.5 V ~5 V,100 Hz~10 kHz	U=0.012%		
		Audio Analyzer Linearity		0.01%~2%,0.02 V ~5 V,20 kHz	U=0.05%		
		RX SSB Phase Noise		-150 dBc/Hz~-50 dBc/Hz,50 MHz ~ 6000 MHz	U=0.1 dB		
		RX AVG Noise Level		-170 dBm~-50 dBm,50 MHz ~ 6000 MHz	U=0.53 dB		
		RX Dynamic Range		-150 dBc~-50 dBc,50 MHz ~ 6000 MHz	U=0.5 dB		
		RX Power Level		-110 dBm~-20 dBm,300 MHz~6000 MHz	U=0.073 dB		
				-20 dBm ~10 dBm,300 MHz~6000 MHz	U=0.046 dB		
		RX Linearity		-2 dB~2 dB,-60 dBm~0 dBm,50 MHz ~ 6000 MHz	U=0.05 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX Frequency Accuracy (Without Reference Connected)	ilac-MRA CNAS INTERNATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	10 MHz~6000 MHz	$U_{rel}=1 \times 10^{-8}$		
		RX Frequency Accuracy (With Reference Connected)		10 MHz~6000 MHz	$U=0.0048$ Hz		
		RX Frequency Response		-5 dB~5 dB,-60 dBm~0 dBm,200 MHz~6000 MHz	$U=0.12$ dB		
		RX Residual AM		0.001%~1%,-40 dBm~0 dBm,50 MHz ~ 6000 MHz	$U=0.001\%$		
		RX Residual FM		0.1 Hz~6 Hz,-20 dBm,50 MHz ~ 6000 MHz	$U=1.8$ Hz		
		RX Residual Response		-100 dBm~-60 dBm,50 MHz ~ 6000 MHz	$U=0.25$ dB		
		RX Harmonics		-20 dBc~-80 dBc,50 MHz ~ 6000 MHz	$U=0.6$ dB		
		RX Spurious / Image Response		-20 dBc~-80 dBc,,50 MHz ~ 6000 MHz	$U=0.1$ dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX IF Frequency Response		0 dB~3 dB, Carrire frequency ,50 MHz ~ 6000 MHz, IF band 120 MHz	U=0.047 dB		
		RX VSWR		1~4, 50 MHz~6000 MHz	U=0.01		
		TX Level Settling Time		0.1 μs~4 μs, 50 MHz ~ 6000 MHz, -50 dBm ~ -10 dBm	U=0.1 μs		
		TX Frequency		50 MHz~6000 MHz	U _{rel} =1x10 ⁻¹⁰		
		TX VSWR		1~4, 50 MHz~6000 MHz	U=0.01		
		TX MAX Output Power		-15 dBm~20 dBm, 50 MHz~6000 MHz	U=0.2 dB		
		RX RBW		1 Hz~3 Hz	U=0.02 Hz		
				10 Hz~30 Hz	U=0.1 Hz		
				100 Hz~300 Hz	U=1 Hz		
				1 kHz~3 kHz	U=0.01 kHz		
				10 kHz~30 kHz	U=0.1 kHz		
				100 kHz~300 kHz	U=1 kHz		
				1 MHz~3 MHz	U=0.01 MHz		
		TX SNR		60 dB~100 dB, 50 MHz~6000 MHz, Resolution Bandwidth 1 kHz	U=1 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TX RFIO Port Balance		-10 dBm,50 MHz~6000 MHz	U=0.05 dB		
		TX Output Power		10 dBm~20 dBm,90 MHz~6000 MHz	U=0.14 dB		
				-20 dBm~10 dBm,90 MHz~6000 MHz	U=0.070 dB		
				-110 dBm~-20 dBm,90 MHz~6000 MHz	U=0.13 dB		
				-120 dBm~-110 dBm,90 MHz~6000 MHz	U=0.14 dB		
		TX Output Linearity		0 dB ~-40 dB,380 MHz~6000 MHz	U=0.07 dB		
		TX Return Loss		0 dB ~25 dB,380 MHz~6000 MHz	U=0.1 dB		
		TX Frequency Settling Time		1 μs~400 μs,50 MHz~6000 MHz, -50 dBm~ -10 dBm	U=1 μs		
		TX Residual FM		0.1 Hz~15 kHz,50 MHz~6000 MHz	U=1.8 Hz		
		TX SSB Phase Noise		-150 dBc/Hz~-50 dBc/Hz,50 MHz~6000 MHz	U=0.21 dB		
		TX Harmonics		-70 dBc~-30 dBc,50 MHz~6000 MHz	U=0.46 dB		
		TX IN Band Spurious		-70 dBc~-30 dBc,50 MHz~6000 MHz	U=0.5 dB		
		1XEV-DO TX RHO		0.9 ~1,870 MHz~2150 MHz	U=0.0019		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		1XEV-DO RX ACP		-90 dB~-10 dB,870 MHz~2150 MHz,-10 dBm~0 dBm	U=0.8 dB		
		1XEV-DO RX EVM		0.01%~4%,870 MHz~ 2150 MHz	U=0.38%		
		1XEV-DO RX Frequency Error		±15 Hz~15 Hz,870 MHz~2150 MHz	U=0.59 Hz		
		TX Frequency Accuracy (Bluetooth)		2402 MHz ~ 2495 MHz	U=6.7 Hz		
		TX Bluetooth Frequency Deviation		-500 kHz ~ 500 kHz,2402 MHz~2495 MHz	U=3 kHz		
		RX Bluetooth Level Accuracy		- 105 dBm ~ 5 dBm,2402 MHz~2495 MHz	U=0.085 dB		
		TX Carrier Suppression		- 105 dBm ~ 5 dBm,50 MHz ~ 6000 MHz	U=0.2 dB		
		RX Bluetooth Frequency Offset		0.1 kHz~1 MHz,, Carrier frequency 2402 MHz ~ 2495 MHz	U=0.1 kHz		
		RX Bluetooth Level Accuracy		- 105 dBm ~ 5 dBm,2402 MHz~2495 MHz	U=0.3 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX Bluetooth Leakage Power		-100 dB ~ -55 dB, 2402 MHz ~ 2495 MHz	U=0.1 dB		
		C2K TX ACP		-90 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.1 dB		
		C2K TX EVM		0.01% ~ 25%, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.48%		
		C2K TX RHO		0.9 ~ 1.0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.001		
		RX SEQAN Format Factor		-1 dB ~ 1 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.002 dB		
		C2K RX ACP		-90 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.8 dB		
		C2K RX EVM		0.01% ~ 25%, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.38%		
		C2K RX RHO		0.9 ~ 1.0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.0012		
		C2K RX Frequency Error		0.01 Hz ~ 10 kHz, 380 MHz ~ 2800 MHz	U=0.01 Hz		
		EDGE RX Origin Offset		-50 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	U=0.48 dB		




在线扫码获取验证

No. CNAS L0640

第 88 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		EDGE RX EVM		0.01%~10% ,0 dBm~- 20 dBm,380 MHz~2800 MHz	U=0.48%		
		EDGE RX ORFS		-90 dB ~ -10 dB,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.8 dB		
		EDGE RX Frequency Error		1 Hz ~10 kHz,380 MHz~2800 MHz	U=1 Hz		
		EDGE TX EVM		0.01%~10% ,0 dBm~- 20 dBm,380 MHz~2800 MHz	U=0.48%		
		TX FM Deviation		8 kHz,Frequency300 Hz ~ 15.999 kHz	U _{rel} =1.2%		
		TX FM Distortion		0.1%~ 3%,Frequency:300 Hz ~ 15.999 kHz ,Mod Rate: 8 kHz	U=0.3%		
		TX FIX Spurious		-90 dB ~ -10 dB,Frequency300 Hz ~ 15.999 kHz,Mod Rate: 8 kHz	U=0.5 dB		
		RX GPS Bias-Tee		10 mA~100 mA,Voltage Setting: 3.3V~5V	U=2.0 mA		
		RX GPS Frequency		1227.6 MHz~1575.42 MHz	U=11 Hz		
		TX GSM Frequency Error		-80 kHz ~80 kHz,710 MHz~1990 MHz,0 dBm~-20 dBm	U=6.1 Hz		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TX GSM Phase Error (Peak, RMS)		-20 ° ~ 20 ° ,380 MHz~710 MHz	U=0.29 °		
				-20 ° ~ 20 ° ,710 MHz~1990 MHz	U=0.078 °		
		RX GSM Origin Offset		-50 dB~-10 dB,0 dBm~-20 dBm,710 MHz~1990 MHz	U=3.1 dB		
		RX GSM ORFS		-90 dB ~ -10 dB,0 dBm~-20 dBm,710 MHz~1990 MHz	U=0.31 dB		
		RX GSM Frequency Error		0.01 Hz~10 kHz,710 MHz~1990 MHz	U=0.01 Hz		
		TX GSM Phase Error (Peak)		-4.0 ° ~ 4.0 ° ,710 MHz~1990 MHz	U=0.079 °		
		RX GSM Phase Error (RMS)		-1.0 ° ~ 1.0 ° ,450 MHz~710 MHz	U=0.29 °		
				-1.0 ° ~ 1.0 ° ,710 MHz~1990 MHz	U=0.079 °		
		TX LTE EVM		0.01%~8% ,0 dBm~-20 dBm,380 MHz~3800 MHz	U=0.29%		
				0.01%~8% ,0 dBm~-20 dBm,3800 MHz~6000 MHz	U=0.48%		
		TX LTE ACP		-80 dB ~ -10 dB,0 dBm~-20 dBm,380 MHz~3800 MHz	U=0.63 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX LTE CHP		0 dBm~-20 dBm,380 MHz~3800 MHz	U=0.05 dB		
		RX LTE EVM		0.01%~8%,0 dBm~-20 dBm,380 MHz~3800 MHz	U=0.23%		
		RX LTE I/Q Offset		-50 dB~-10 dB,0 dBm~-20 dBm,380 MHz~3800 MHz	U=0.48 dB		
		RX LTE Frequency Error		0.01 Hz~10 kHz,380 MHz~3800 MHz	U=0.07 Hz		
		RX Seqaq Trans Power		0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.1 dB		
		TX TDSCDMA Modulation (EVM)		0.01% ~ 18%,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.95%		
		TX TD-SCDMA Modulation (Rho)		0.9~1,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.0019		
		RX TD-SCDMA ACP		-80 dB~-10 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.8 dB		
		RX TD-SCDMA EVM		0.01% ~ 18%,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.48%		
		RX TDSCDMA I/Q Imbalance		-1 dB~1 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.5 dB		




No. CNAS L0640

第 91 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX TDSCDMA I/Q Origin Offset		-50 dB~-10 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.4 dB		
		RX TDSCDMA Frequency Error		-20 Hz~20 Hz,380 MHz~2800 MHz,0 dBm~-20 dBm	U=0.75 Hz		
		TX WCDMA HSDPA EVM		0.01% ~ 25%,0 dBm~-20 dBm,380 MHz~6000 MHz	U=1.1%		
		TX WCDMA EVM		0.01% ~ 25%,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.48%		
		TX WCDMA Level Error		0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.17 dB		
		RX TDSCDMA I/Q Origin Offset		-50 dB~-10 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.4 dB		
		RX WCDMA Peak Code Domain Error		-1 dB~1 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.1 dB		
		RX WCDMA ACP		-80 dB~-10 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.8 dB		
		RX WCDMA EVM		0.01% ~ 25%,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.48%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX WCDMA SEQAN Format Factor		0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.17 dB		
		RX WCDMA Frequency Error		-20 Hz~20 Hz,380 MHz~2800 MHz,0 dBm~-20 dBm	U=0.01 Hz		
		TX WLAN 802.11a EVM		-53 dB~-22 dB,5.8GHz,0 dBm~-20 dBm	U=0.29 dB		
		TX WLAN 802.11ac EVM		-53 dB~-22 dB,5.5 GHz~ 5.8 GHz,0 dBm~-20 dBm,Bandwidth 80 MHz	U=0.35 dB		
		TX WLAN 802.11ac EVM		-53 dB~-22 dB,5.5 GHz~ 5.8 GHz,0 dBm~-20 dBm,Bandwidth 160 MHz	U=0.36 dB		
		TX WLAN 802.11b EVM		0.01%~ 8%,2.4 GHz,0 dBm~-20 dBm	U=0.24%		
		TX WLAN 802.11g EVM		-53 dB~-22 dB,2.4 GHz,0 dBm~-20 dBm	U=0.31 dB		
		TX WLAN 802.11n EVM		-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm, Bandwidth 20 MHz	U=2.3 dB		
				-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm, Bandwidth 40 MHz	U=0.29 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TX WLAN 802.11af EVM		-58 dB~-22 dB,700 MHz,-20 dBm~0 dBm,Bandwidth 8 MHz	U=0.35 dB		
		TX WLAN 802.11ah EVM		-58 dB~-22 dB,900 MHz,-20 dBm~0 dBm,Bandwidth 16 MHz	U=0.39 dB		
		TX WLAN 802.11ax EVM		-53 dB~-22 dB,5.8 GHz,-20 dBm~0 dBm,Bandwidth 80 MHz	U=0.58 dB		
		TX WLAN Frequency		-53 dB~-22 dB,5.8 GHz,-20 dBm~0 dBm,Bandwidth 16 MHz	U=0.80 dB		
		TX WLAN Power		2000 MHz~6000 MHz	U=6.8 Hz		
		TX Modulation Depth		-60 dBm~-40 dBm,2000 MHz~6000 MHz	U=0.12 dB		
		RX WLAN 802.11a EVM		-40 dBm~-10 dBm,2000 MHz~6000 MHz	U=0.08 dB		
		RX WLAN 802.11ac EVM		-500 Hz~500 Hz,Mod index 0.28~0.35, 2.4 GHz ~ 2.48 GHz, -90 dBm ~ -20 dBm, Bluetooth GFSK	U=9.7 Hz		
				-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm	U=0.3 dB		
				-53 dB~-22 dB,5.5 GHz~ 5.8 GHz, 0 dBm~-20 dBm,80 MHz Bandwidth	U=0.39 dB		




在线扫码获取验证

No. CNAS L0640

第 94 页 共 209 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-53 dB~-22 dB,5.5 GHz~ 5.8 GHz, 0 dBm~-20 dBm,160 MHz Bandwidth	U=0.38 dB		
		RX WLAN 802.11b EVM		0.01%~ 8%,2.4 GHz,0 dBm~-20 dBm	U=0.29%		
		RX WLAN 802.11g EVM		-53 dB~-22 dB,2.4 GHz,0 dBm~-20 dBm	U=0.31 dB		
		RX WLAN 802.11n EVM		-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm,20 MHz Bandwidth	U=1.6 dB		
		RX WLAN 802.11n EVM		-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm,40 MHz Bandwidth	U=0.29 dB		
		RX WLAN 802.11af EVM		-53 dB~-22 dB,700 MHz,0 dBm~-20 dBm,8 MHz Bandwidth	U=0.68 dB		
		RX WLAN 802.11ah EVM		-53 dB~-22 dB,900 MHz,0 dBm~-20 dBm,16 MHz Bandwidth	U=0.63 dB		
		RX WLAN 802.11ax EVM		-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm,80 MHz Bandwidth	U=0.54 dB		
		RX WLAN 802.11ax EVM		-53 dB~-22 dB,5.8 GHz,0 dBm~-20 dBm,160 MHz Bandwidth	U=0.48 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX Frequency Deviation Range Accuracy	ilac-MRA	0.01 Hz~400 kHz,2000 MHz~6000 MHz	$U=0.12$ Hz		
		RX UL MOD Level		-10 dBm~-50 dBm,800 MHz~6000 MHz	$U=0.13$ dB		
		RX UL EVM		0.01%~25%,Bandwidth 5 MHz~20 MHz,-10 dBm~-50dBm,500 MHz~6000 MHz	$U=0.57\%$		
		TX DL EVM		0.01%~25%,Bandwidth 5 MHz~20 MHz,-10 dBm~-50dBm,500 MHz~6000 MHz	$U=0.5\%$		
9	*LF/MW Network Analyzer	RF Output Level Linearity	LF/MW Network Analyzer Test Procedure CAL-CHA005-01	-25 dBm~15 dBm,9 kHz~100 kHz	$U=0.01$ dB		
				-25 dBm~15 dBm,100 kHz ~1 MHz	$U=0.045$ dB		
				-25 dBm~15 dBm,1 MHz~50 MHz	$U=0.10$ dB		
				-25 dBm~15 dBm,50 MHz~500 MHz	$U=0.02$ dB		
				-25 dBm~15 dBm,500 MHz~20 GHz	$U=0.01$ dB		
				Frequency	9 kHz~100 kHz		
		100 kHz~20 GHz			$U_{rel}=7.3 \times 10^{-10}$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RF Output Level		15 dBm~20 dBm ,9 kHz~100 kHz	U=0.10 dB		
				15 dBm~20 dBm ,100 kHz~8 GHz	U=0.053 dB		
				15 dBm~20 dBm ,8 GHz~20 GHz	U=0.081 dB		
				10 dBm~15 dBm ,9 kHz~100 kHz	U=0.097 dB		
				10 dBm~15 dBm ,100 kHz~8 GHz	U=0.051 dB		
				10 dBm~15 dBm ,8 GHz~20 GHz	U=0.079 dB		
				3 dBm~10 dBm ,9 kHz~100 kHz	U=0.058 dB		
				3 dBm~10 dBm ,100 kHz~8 GHz	U=0.043 dB		
				3 dBm~10 dBm ,8 GHz~20 GHz	U=0.075 dB		
				-3 dBm~3 dBm ,9 kHz~12.4 GHz	U=0.030 dB		
				-3 dBm~3 dBm ,12.4 GHz~18 GHz	U=0.034 dB		
				-3 dBm~3 dBm ,18 GHz~20 GHz	U=0.073 dB		
				-20 dBm~-3 dBm ,9 kHz~100 kHz	U=0.13 dB		
				-20 dBm~-3 dBm ,100 kHz~8 GHz	U=0.047 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~-3 dBm ,8 GHz~12.4 GHz	U=0.061 dB		
				-20 dBm~-3 dBm ,12.4 GHz~14 GHz	U=0.070 dB		
				-20 dBm~-3 dBm ,14 GHz~18 GHz	U=0.074 dB		
				-20 dBm~-3 dBm ,18 GHz~20 GHz	U=0.077 dB		
		RF Output Flatness		-17 dBm~10 dBm ,9 kHz~300 kHz	U=0.062 dB		
				-17 dBm~10 dBm ,0.3 MHz~45 MHz	U=0.08 dB		
				-17 dBm~10 dBm ,45 MHz~2 GHz	U=0.044 dB		
				-17 dBm~10 dBm ,2 GHz~20 GHz	U=0.10 dB		
		Noise Floor		-145 dBm~-80 dBm,9 kHz~10 MHz	U=0.71 dB		
				-145 dBm~-80 dBm,10 MHz~20 GHz	U=1.1 dB		
		Directivity Calibration Coefficients		-60 dB~-10 dB ,9 kHz~300 kHz	U=0.17 dB		
				-62 dB~-10 dB,300 kHz~8.5 GHz	U=0.17 dB		
				-40 dB~-10 dB,8.5 GHz~20 GHz	U=0.66 dB		
		Source Match Calibration		-60 dB~-10 dB,9 kHz ~ 8.5 GHz	U=0.42 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Coefficients		-60 dB~-10 dB,8.5GHz ~ 20 GHz	U=2.3 dB		
		Load Match Calibration Coefficients		-60 dB~-10 dB,9 kHz ~ 300 kHz	U=0.11 dB		
			-60 dB~-10 dB,300 kHz ~ 8.5 GHz	U=0.13 dB			
			-60 dB~-10 dB,8.5GHz ~ 20 GHz	U=0.27 dB			
		Transmission Tracking Coefficients		-2 dB~2 dB,9 kHz ~ 300 kHz	U=0.051 dB		
			-2 dB~2 dB,300 kHz ~ 8.5 GHz	U=0.04 dB			
			-2 dB~2 dB,8.5GHz ~ 20 GHz	U=0.27 dB			
		Reflection Tracking Coefficients		-2 dB~2 dB,9 kHz ~ 8.5 GHz	U=0.021 dB		
			-2 dB~2 dB,8.5GHz ~ 20 GHz	U=0.077 dB			
		Compression Magnitude		-3 dB~3 dB ,9 kHz~8.5 GHz	U=0.023 dB		
			-3 dB~3 dB ,8.5 GHz~ 20 GHz	U=0.034 dB			
		Compression Phase		-30 ° ~ 30 ° ,9 kHz~ 8.5 GHz	U=0.058 °		
			-30 ° ~30 ° ,8.5 GHz~20 GHz	U=0.46 °			
		Dynamic Accuracy		-100 dBm~-80 dBm,10MHz~1.195 GHz	U=0.044 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-80 dBm~-60 dBm,10 MHz~1.195 GHz	U=0.019 dB		
				-60 dBm~-50 dBm,10 MHz~1.195 GHz	U=0.011 dB		
				-50 dBm~-35 dBm,10 MHz~1.195 GHz	U=0.007 dB		
				-35 dBm~0 dBm,10 MHz~1.195 GHz	U=0.00014 dB		
				0 dBm~10 dBm,10 MHz~1.195 GHz	U=0.0052 dB		
				-110 dBm~-100 dBm,100 kHz~1 MHz	U=0.16 dB		
				-100 dBm~-80 dBm,100 kHz~1 MHz	U=0.016 dB		
				-80 dBm~-60 dBm,100 kHz~1 MHz	U=0.012 dB		
				-60 dBm~-50 dBm,100 kHz~1 MHz	U=0.018 dB		
				-50 dBm~-30 dBm,100 kHz~1 MHz	U=0.0031 dB		
				-30 dBm~0 dBm,100 kHz~1 MHz	U=0.0088 dB		
		Dynamic Accuracy Phase		-20° ~ 20°, -115 dBm~-100 dBm,100 kHz~1 MHz	U=1.1°		
				-20° ~ 20°, -100 dBm~-80 dBm,100 kHz~1 MHz	U=0.12°		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 ° ~ 20 ° , -80 dBm ~ -60 dBm, 100 kHz ~ 1 MHz	U=0.036 °		
				-20 ° ~ 20 ° , -60 dBm ~ -50 dBm , 100 kHz ~ 1 MHz	U=0.032 °		
				-20 ° ~ 20 ° , -50 dBm ~ -30 dBm, 100 kHz ~ 1 MHz	U=0.02 °		
				-20 ° ~ 20 ° , -30 dBm ~ 0 dBm, 100 kHz ~ 1 MHz	U=0.014 °		
		Crosstalk		-150 dB ~ -90 dB , 9 kHz ~ 8.5 GHz	U=1.2 dB		
				-150 dB ~ -90 dB , 8.5 GHz ~ 20 GHz	U=2.4 dB		
		Trace Noise Magnitude		0.00001 dB ~ 0.02 dB , 9 kHz ~ 10 MHz	U=0.00012 dB		
				0.00001 dB ~ 0.02 dB , 10 MHz ~ 4.38 GHz	U=0.00004 dB		
				0.00001 dB ~ 0.02 dB , 4.38 GHz ~ 8.5 GHz	U=0.00013 dB		
				0.00001 dB ~ 0.02 dB , 8.5 GHz ~ 20 GHz	U=0.00026 dB		
		Trace Noise Phase		0.001 ° ~ 0.20 ° , 9 kHz ~ 10 MHz	U=0.0004 °		
				0.001 ° ~ 0.20 ° , 10 MHz ~ 4.38 GHz	U=0.0011 °		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~-3 dBm ,6 GHz~8 GHz	U=0.061 dB		
				-3 dBm~20 dBm ,6 GHz~8 GHz	U=0.078 dB		
				-20 dBm~-3 dBm ,8 GHz~14 GHz	U=0.070 dB		
				-3 dBm~20 dBm ,8 GHz~14 GHz	U=0.078 dB		
				-20 dBm~-3 dBm ,14 GHz~18 GHz	U=0.074 dB		
				-3 dBm~20 dBm ,14 GHz~18 GHz	U=0.078 dB		
				-20 dBm~-3 dBm ,18 GHz~24 GHz	U=0.077 dB		
				-3 dBm~20 dBm ,18 GHz~24 GHz	U=0.081 dB		
				-20 dBm~-3 dBm ,24 GHz~26.5 GHz	U=0.092 dB		
				-3 dBm~20 dBm ,24 GHz~26.5 GHz	U=0.095 dB		
				-20 dBm~20 dBm ,26.5 GHz~67 GHz	U=0.12 dB		
		Noise Floor		-140 dBm~-60 dBm,900 Hz~300 kHz	U=0.91 dB		
		Noise Floor		-140 dBm~-60 dBm,300 kHz~10 MHz	U=0.26 dB		
		Noise Floor		-140 dBm~-60 dBm,10 MHz~50 GHz	U=0.20 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Directivity Calibration Coefficients		-140 dBm~-60 dBm,50 GHz~67 GHz	U=0.91 dB		
				-30 dB~2 dB,900 Hz~ 50 MHz	U=0.15 dB		
				-39 dB~-5 dB,50 MHz~ 200 MHz	U=0.12 dB		
				-48 dB~-23 dB,200 MHz~500 MHz	U=0.34 dB		
				-46 dB~-20 dB,500 MHz~10 GHz	U=0.30 dB		
				-44 dB~-16 dB,10 GHz~20 GHz	U=0.24 dB		
				-43 dB~-14 dB,20 GHz~26.5 GHz	U=0.22 dB		
				-39 dB~-13 dB,26.5 GHz~50 GHz	U=0.30 dB		
				-38 dB~-10 dB,50 GHz~67 GHz	U=0.34 dB		
		Source Match Calibration Coefficients		-22 dB~-5 dB,900 Hz~ 50 MHz	U=0.097 dB		
				-43 dB~-6 dB,50 MHz~ 200 MHz	U=0.086 dB		
				-41 dB~-6 dB,200 MHz~500 MHz	U=0.11 dB		
				-46 dB~-6 dB,500 MHz~3.2 GHz	U=0.086 dB		
				-39 dB~-7 dB,3.2 GHz~13.5 GHz	U=0.15 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-36 dB~-7 dB,13.5 GHz~20 GHz	U=0.22 dB		
				-32 dB~-7 dB,20 GHz~26.5 GHz	U=0.34 dB		
				-30 dB~-7 dB,26.5 GHz~50 GHz	U=0.43 dB		
				-24 dB~-5 dB,50 GHz~67 GHz	U=0.66 dB		
		Load Match Calibration Coefficients		-29 dB~-7 dB,900 Hz~50 MHz	U=0.12 dB		
				-37 dB~-7 dB,50 MHz~200 MHz	U=0.18 dB		
				-44 dB~-7 dB,200 MHz~500 MHz	U=0.084 dB		
				-46 dB~-6 dB,500 MHz~10 GHz	U=0.10 dB		
				-40 dB~-6 dB,10 GHz~20 GHz	U=0.12 dB		
				-39 dB~-7 dB,20 GHz~26.5 GHz	U=0.15 dB		
				-35 dB~-7 dB,26.5 GHz~50 GHz	U=0.21 dB		
				-34 dB~-7 dB,50 GHz~60 GHz	U=0.25 dB		
				-32 dB~-6 dB,60 GHz~67 GHz	U=0.30 dB		
			Compression Magnitude (including		-1 dB~1 dB ,900 Hz~100 MHz	U=0.016 dB	



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Noise Receiver)		-1 dB~1 dB,100 MHz~50 GHz	U=0.024 dB		
				-1 dB~1 dB,50 GHz~67 GHz	U=0.034 dB		
		Compression Phase		-30 ° ~30 °,900 Hz~100 MHz	U=0.11 °		
				-30 ° ~30 °,100 MHz~50 GHz	U=0.08 °		
		Dynamic Accuracy (including Noise Receiver)		-30 ° ~30 °,50 GHz~67 GHz	U=0.11 °		
				-40 dB~-35 dB,900 kHz~2 GHz	U=0.0065 dB		
				-35 dB~-30 dB,900 kHz~2 GHz	U=0.0059 dB		
				-30 dB~-25 dB,900 kHz~2 GHz	U=0.0054 dB		
				-25 dB~-20 dB,900 kHz~2 GHz	U=0.0048 dB		
				-20 dB~-15 dB,900 kHz~2 GHz	U=0.0039 dB		
				-15 dB~-10 dB,900 kHz~2 GHz	U=0.0032 dB		
				-10 dB~-5 dB,900 kHz~2 GHz	U=0.0024 dB		
				-5 dB~0 dB,900 kHz~2 GHz	U=0.0012 dB		
			0 dB~5 dB,900 kHz~2 GHz	U=0.0010 dB			



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				5 dB~10 dB,900 kHz~2 GHz	U=0.0023 dB		
				10 dB~15 dB,900 kHz~2 GHz	U=0.0033 dB		
				15 dB~20 dB,900 kHz~2 GHz	U=0.0039 dB		
				20 dB~25 dB,900 kHz~2 GHz	U=0.0047 dB		
		Trace Noise Magnitude		0 dB~0.02 dB ,900 Hz~300 kHz	U=0.00044 dB		
				0 dB~0.02 dB ,300 kHz~45 MHz	U=0.00002 dB		
				0 dB~0.02 dB ,45 MHz~50 GHz	U=0.000017 dB		
				0 dB~0.02 dB ,50 GHz~67 GHz	U=0.000025 dB		
		Trace Noise Phase		0.001° ~0.20° ,900 Hz~300 kHz	U=0.00033 °		
				0.001° ~0.20° ,300 kHz~45 MHz	U=0.00010 °		
				0.001° ~0.20° ,45 MHz~50 GHz	U=0.000068 °		
				0.001° ~0.20° ,50 GHz~67 GHz	U=0.00055 °		
		Noise Jitter		0 dB~0.10 dB,45 MHz~50 GHz	U=0.0057 dB		
		Receiver Noise Figure		10 dB~21 dB ,10 MHz~50 GHz	U=0.12 dB		



No. CNAS L0640


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Receiver Noise Figure (26.5 GHz and below)	ilac-M	10 dB~21 dB ,10 MHz~18 GHz	U=0.41 dB		
				10 dB~21 dB ,18 GHz~25 GHz	U=0.50 dB		
				10 dB~21 dB ,25 GHz~26.5 GHz	U=0.78 dB		
		VSWR		1 ~4 ,300 kHz~67 GHz	U=0.03		
		SA Detector Accuracy		-0.2 dB~0.2 dB ,300 kHz~67 GHz	U=0.005 dB		
		Display Average Noise Level		-150 dBm~-80 dBm,900 Hz~67 GHz, RBW 1 Hz	U=0.15 dB		
		Dynamic Range		-150 dB~-90 dB ,900 Hz~67 GHz	U=0.92 dB		
11	*Portable Network Analyzer	Display Linearity	Portable Network Analyzer Test Procedure CAL-CHA005-03	0 dBm~-50 dBm,300 kHz ~14 GHz	U=0.02 dB		
		Output Frequency		2 GHz	$U_{rel}=2.8 \times 10^{-8}$		
		Output Power Linearity		0 dBm~-50 dBm,300 kHz ~14 GHz	U=0.23 dB		
		Output Power Level		-9 dBm~-11 dBm,300 kHz ~14 GHz	U=0.11 dB		
		Absolute Level Measurement		-0.3 dB~0.3 dB,65.833 MHz	U=0.12 dB		
		Frequency Response		-1.2 dB~1.2 dB,300 kHz ~14 GHz	U=0.15 dB		
		Attenuator		0 dB~30 dB,65.833 MHz	U=0.11 dB		



在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Trace Noise Magnitude	ilac-M	0 dB~0.02 dB ,300 kHz~14 GHz	U=0.001 dB		
		Trace Noise Phase		0.001 ° ~0.20 ° ,300 kHz~14 GHz	U=0.01 °		
		Phase Noise		-130 dB/Hz~-70 dB/Hz ,Frequency: 500 MHz,Deviation: 10 kHz~1 MHz	U=2.2 dB		
		Noise Display		-150 dBm~-70 dBm,300 kHz~14 GHz ,RBW 1 Hz	U=0.4 dB		
		Dynamic Range		-150 dB~-90 dB ,300 kHz~13 GHz	U=0.2 dB		
		Third Order Intermodulation		5 dBm~20 dBm ,30 MHz~14 GHz	U=0.5 dB		
12	*General-Purpose Network Analyzer	Test Port Output Linearity	General-purpose Network Analyzer Test Procedure CAL-CHA005-04	2.5 dBm~13 dBm,300 kHz~6 GHz	U=0.025 dB		
		Output Frequency		300 kHz~6.75 GHz ,	U _{rel} =2.5x10 ⁻⁷		
		Minimum R Level		-30 dBm~-50 dBm ,300 kHz~6 GHz	U=1.5 dB		
		Test Port Output Power		-17 dBm~10 dBm ,0.3 MHz~6 GHz	U=0.19 dB		
		Receiver Magnitude Frequency Response		-2 dB~2 dB ,0.3 MHz~6 GHz	U=0.050 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Receiver Phase		-10° ~ 10° ,0.3	U=0.6°		
		Frequency Response		MHz~6 GHz,-17dBm~10 dBm			
		Test Port Input Noise Floor Level		-80 dBm~-140 dBm ,300 kHz~6 GHz	U=1.2 dB		
		Crosstalk		-150 dB~-90 dB ,300 kHz~6 GHz	U=2.2 dB		
		Test Port Return loss		-50 dB~-10 dB ,300 kHz ~6 GHz	U=0.38 dB		
		Directivity Calibration Coefficients		-40 dB~-10 dB ,300 kHz ~6 GHz	U=0.99 dB		
		Source Match Calibration Coefficients		-60 dB~-10 dB,300 kHz ~6 GHz	U=0.26 dB		
		Load Match Calibration Coefficients		-60 dB~-10 dB,300 kHz ~6 GHz	U=0.22 dB		
		Calibration Coefficients - Transmission Tracking		-1 dB~1 dB,300 kHz ~6 GHz	U=0.020 dB		
		Calibration Coefficients - Reflection Tracking		-1 dB~1 dB,300 kHz ~6 GHz	U=0.020 dB		
		Receiver Compression - Magnitude		-1 dB~1 dB ,300 kHz ~6 GHz	U=0.048 dB		




在线扫码获取验证

No. CNAS L0640

第 110 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Receiver Compression - Phase		30 ° ~ 30 ° , 300 kHz ~ 6 GHz	U=0.033 °		
		Receiver Dynamic		-110 dBm ~ 0 dBm, 300 kHz ~ 6 GHz	U=0.006 dB		
		Trace Noise Magnitude		0 dB ~ 0.02 dB , 300 kHz ~ 6 GHz	U=0.00016 dB		
		Trace Noise Phase		0.001 ° ~ 0.20 ° , 300 kHz ~ 6 GHz	U=0.00033 °		
		Source Harmonics		-60 dBc ~ -10 dBc , 300 kHz ~ 6 GHz	U=1.8 dB		
		Source Mixer Spurs		-80 dB ~ -50 dB , 300 kHz ~ 6 GHz, Offset 4 kHz ~ 100 kHz	U=1.6 dB		
		Phase Noise		-100 dBc/Hz ~ -50 dBc/Hz , Frequency 136 MHz, offset 100 Hz ~ 100 kHz	U=1.6 dB		
		Source & Receiver Harmonics		-60 dBc ~ -10 dBc , 300 kHz ~ 6 GHz	U=0.24 dB		
		Harmonic Measurement		-60 dBc ~ -10 dBc , 300 kHz ~ 6 GHz	U=0.10 dB		
13	*DC Power Supply	Transient Recovery Time	DC Power Supply Test Procedure CAL-CHA006-01	1 μs ~ 6 ms, 10 mV ~ 300 V	U=0.053 μs		
		Transient Recovery Voltage		10 mV ~ 300 V, 1 μs ~ 6 ms	U _{rel} =0.02%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		CV PARD PK-PK		0.1 mV~300 mV,20 Hz~20 MHz	$U_{rel}=3.7\%$		
		CV PARD RMS		0.01 mV~150 mV,20 Hz~20 MHz	$U_{rel}=1.0\%$		
		CV Voltage		10 mV~1V	$U_{rel}=0.054\%$		
				1 V~10V	$U_{rel}=0.0039\%$		
				10 V~100V	$U_{rel}=0.0012\%$		
		CV Voltage Readback		100 V~1000 V	$U_{rel}=0.0013\%$		
				10 mV~1 V	$U_{rel}=0.054\%$		
				1 V~10V	$U_{rel}=0.0015\%$		
		CV Load Effect		10 V~100V	$U_{rel}=0.0013\%$		
				100 V~1000 V	$U_{rel}=0.0013\%$		
				10 mV~1 V	$U_{rel}=0.0094\%$		
		CV Source Effect		1 V~10 V	$U_{rel}=0.0017\%$		
				10 V~100 V	$U_{rel}=0.00011\%$		
				100 V~600 V	$U_{rel}=0.0003\%$		
		CV Source Effect		10 mV~1 V	$U_{rel}=0.0094\%$		
				1 V~10 V	$U_{rel}=0.0013\%$		
				10 V~100 V	$U_{rel}=0.00011\%$		
				100 V~600 V	$U_{rel}=0.0003\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		CC Current	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	1 mA~500 mA	$U_{rel}=0.011\%$		
				0.5 A~3 A	$U_{rel}=0.0075\%$		
				3 A~5 A	$U_{rel}=0.012\%$		
				5 A~15 A	$U_{rel}=0.011\%$		
				15 A~100 A	$U_{rel}=0.0069\%$		
				100 A~300 A	$U_{rel}=0.012\%$		
		CC Current Readback		1 mA~500 mA	$U_{rel}=0.013\%$		
				0.5 A~3 A	$U_{rel}=0.008\%$		
				3 A~5 A	$U_{rel}=0.012\%$		
				5 A~15 A	$U_{rel}=0.011\%$		
				15 A~100 A	$U_{rel}=0.0069\%$		
				100 A~300 A	$U_{rel}=0.012\%$		
		Current Sink		0.02 A~12 A	$U_{rel}=0.010\%$		
		CC Load Effect		0.5 A~7 A	$U_{rel}=0.0004\%$		
				7 A~15 A	$U_{rel}=0.00067\%$		
				15 A~100 A	$U_{rel}=0.00033\%$		
				100 A~300 A	$U_{rel}=0.003\%$		
		CC Source Effect		0.5 A~7 A	$U_{rel}=0.0004\%$		
				7 A~15 A	$U_{rel}=0.00067\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA	15 A~100 A	$U_{rel}=0.00033\%$		
				100 A~300 A	$U_{rel}=0.003\%$		
		CC PARD RMS		0.01 mA~200 mA,20 Hz~20 MHz	$U_{rel}=3.7\%$		
		Resistance		1 Ω	$U_{rel}=0.22\%$		
14	*Oscilloscope	Delay Time	Oscilloscope Test Procedure CAL-CHA007-01	2.5 ns~800 ns	$U=0.002$ ns		
				800 ns~8 μ s	$U=0.012$ ns		
				8 μ s~1 ms	$U=1.3$ ns		
		Horizontal Delta Time		1.9 ns~2 ns	$U=0.12$ ps		
				2 ns~600 ns	$U=0.002$ ns		
				600 ns~800 ns	$U=0.04$ ns		
				800 ns~1 μ s	$U=0.1$ ns		
				1 μ s~2 μ s	$U=0.2$ ns		
		Timescale		2 μ s~100 μ s	$U=0.02$ μ s		
				25 ns~100 ns	$U_{rel}=1.9 \times 10^{-9}$		
		DC Voltage		10 mV~14.9 mV	$U_{rel}=0.013\%$		
				14.9 mV~143 mV	$U_{rel}=0.0014\%$		
				143 mV~715 mV	$U_{rel}=0.0015\%$		
				0.715 V~1.43 V	$U_{rel}=0.0016\%$		
1.43 V~7.15 V	$U_{rel}=0.0010\%$						




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Gain Accuracy		7.15 V~35 V	$U_{rel}=0.0013\%$		
				35 V~160 V	$U_{rel}=0.0014\%$		
				-2%~2%,6 mV~9 mV,1 mV/div	$U=0.36\%$		
				-3%~3%,12 mV~18 mV,2 mV/div	$U=0.33\%$		
				-3%~3%,30 mV~45 mV,4.98 mV/div~6.25 mV/div	$U=0.12\%$		
				-3%~3%,60 mV~400 mV,10 mV/div	$U=0.013\%$		
				-3%~3%,120 mV~400 mV,20 mV/div	$U=0.016\%$		
				-3%~3%,300 mV~700 mV,49.8 mV/div~50 mV/div	$U=0.025\%$		
				-3%~3%,600 mV~1200 mV,100 mV/div~150 mV/div	$U=0.013\%$		
				-3%~3%,1.2 V~2.2 V,200 mV/div	$U=0.02\%$		
				-3%~3%,1.2 V~4.5 V,400 mV/div~500 mV/div	$U=0.025\%$		
				-3%~3%,1.2 V~9 V,1 V/div	$U=0.016\%$		
				-2%~2%,12 V~16 V,2 V/div	$U=0.02\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-2%~2%,30V,5 V/div	U=0.02%		
				-2%~2%,60V,10 V/div	U=0.8%		
		Bandwidth		-3 dB~3 dB,50 MHz~2300 MHz	U=0.23 dB		
				-3 dB~3 dB,2.3 GHz~11 GHz	U=0.30 dB		
				-3 dB~3 dB,11 GHz~18 GHz	U=0.60 dB		
				-3 dB~3 dB,18 GHz~26.5 GHz	U=0.69 dB		
				-3 dB~3 dB,26.5 GHz~33 GHz	U=0.70 dB		
				-3 dB~3 dB,33 GHz~50 GHz	U=0.81 dB		
			Threshold Voltage		± (10 mV~5 V)	U _{rel} =0.02%	
		DC Calibrator		0.25 V~5 V	U _{rel} =0.004%		
		Aux Trigger Out		Low Level ± (10 mV~0.7 V)	U _{rel} =1.7%		
				High Level 0.5 V~4 V	U _{rel} =2%		
				Peak to Peak 50 mV~1 V	U _{rel} =4.8%		
		Delta Time Measurement		0.1 ps~5.7 ps, 501 MHz	U=0.017 ps		
				5.7 ps~12 ps,20 MHz	U=0.048 ps		
				45 ps~940 ps,24 MHz~240 MHz	U=0.47 ps		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Impedance Test		2 ns~940 ns,24 kHz~2.4 MHz	$U=0.047$ ns		
				$2.0 \mu s \sim 9.4 \mu s, 2.4$ kHz	$U=0.047 \mu s$		
				50 Ω	$U_{rel}=0.0036\%$		
				1 M Ω	$U_{rel}=0.013\%$		
				250 k Ω	$U_{rel}=0.2\%$		
15	*Digital Communications Analyzer Oscilloscope	Time-base	Digital Communications Analyzer Oscilloscope Test Procedure CAL-CHA007-02	40 ps~60 ps, Delay: 24 ns Frequency: 19.98 GHz	$U=0.37$ ps		
				40 ps~60 ps, Delay: 27.95 ns Frequency: 19.98 GHz	$U=0.52$ ps		
				40 ps~60 ps, Delay: 35.95 ns Frequency: 19.98 GHz	$U=0.55$ ps		
				40 ps~60 ps, Delay: 59.95 ns Frequency: 19.98 GHz	$U=0.60$ ps		
				90 ps~110 ps, Delay: 24 ns Frequency: 10 GHz	$U=1.1$ ps		
				190 ps~210 ps, Delay: 24 ns Frequency: 5 GHz	$U=0.88$ ps		
				490 ps~510 ps, Delay: 24 ns Frequency: 2 GHz	$U=1.3$ ps		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				9980 ps~10020 ps, Delay: 24 ns Frequency: 1 GHz	U=1.6 ps		
				1990 ps~2010 ps, Delay: 24 ns Frequency: 500 MHz	U=1.2 ps		
				4987 ps~5013 ps, Delay: 24 ns Frequency: 200 MHz	U=2.1 ps		
				980 ps~1020 ps, Delay: 24 ns Frequency: 100 MHz	U=4.4 ps		
				19.97 ns~20.03 ns, Delay: 24 ns Frequency: 50 MHz	U=4.2 ps		
				49.94 ns~50.06 ns, Delay: 24 ns Frequency: 20 MHz	U=10 ps		
		Fine Time-base		24 ns	U=0.86 ps		
				28 ns	U=0.75 ps		
				40 ns	U=0.53 ps		
				64 ns	U=0.46 ps		
		Trigger Jitter Test		0.03 ps~ 3 ps, Delay 24 ns, 2 GHz	U=0.13 ps		
				0.03 ps~ 3 ps, Delay 24 ns, 2.5 GHz	U=0.14 ps		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.03 ps~ 3 ps, Delay24 ns,2.75 GHz	U=0.083 ps		
				0.03 ps~ 3 ps, Delay24 ns,3 GHz	U=0.15 ps		
				0.03 ps~ 3 ps, Delay24 ns,3.2 GHz	U=0.14 ps		
				0.03 ps~ 3 ps, Delay24 ns,5 GHz	U=0.11 ps		
				0.03 ps~ 3 ps, Delay24 ns,8 GHz	U=0.06 ps		
				0.03 ps~ 3 ps, Delay24 ns,10 GHz	U=0.058 ps		
				0.03 ps~ 3 ps, Delay24 ns,12 GHz	U=0.1 ps		
				0.03 ps~ 3 ps, Delay24 ns,13 GHz	U=0.059 ps		
				0.03 ps~ 3 ps, Delay24 ns,14 GHz	U=0.1 ps		
				0.03 ps~ 3 ps, Delay24 ns,15 GHz	U=0.059 ps		
				0.03 ps~ 3 ps, Delay41.1 ns,3 GHz	U=0.14 ps		
				0.03 ps~ 3 ps, Delay100 ns,12 GHz	U=0.058 ps		
				0.03 ps~ 15 ps, Delay200 ns,2.5 GHz	U=0.087 ps		
		Trigger Level		-0.4 V	U=0.0035 V		
				-0.32 V	U=0.0027 V		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-0.24 V	U=0.0031 V		
				-0.16 V	U=0.0018 V		
				-0.08 V	U=0.0025 V		
				0 V	U=0.0013 V		
				0.08 V	U=0.0018 V		
				0.16 V~0.24 V	U=0.0027 V		
				0.32 V	U=0.0032 V		
				0.4 V	U=0.0036V		
16	*Oscilloscope Module	Bandwidth	Oscilloscope Module Test Procedure CAL-CHA007-03	0 dB~3 dB,12.4 GHz	U=0.23 dB		
				0 dB~3 dB,18 GHz	U=0.25dB		
				0 dB~3 dB,19.98 GHz	U=0.29 dB		
				0 dB~3 dB,26.5 GHz	U=0.24 dB		
				0 dB~3 dB,50 GHz	U=0.44 dB		
		DC Voltage Accuracy		0.5 V	U=0.00045 V		
				0.25 V	U=0.00016 V		
				-0.25 V	U=0.00021 V		
				-0.5 V	U=0.00044 V		
		Input Reflection		-5%~5%,Input Port,30 ps POS/NEG Rise time	U=2.3%		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
17	*RF/MW Power Meter	Instrumentation Measurement	RF/MW Power Meter Test Procedure CAL-CHA008-01	-10%~10%, Trigger Port, 100 ps~200 ps POS/NEG Rise time	U=2.3%		
				Low Level: 0 mV	U=0.055 mV		
				High Level: 200 mV	U=0.062 mV		
				-1%~1%, 1 ns~100 ns	U=0.11%		
				-5%~5%, 45 ps~1 ns	U=0.18%		
				0.1 ps~45 ps	U=2.8 ps		
				3.1623 μW	U=18 nW		
				10 μW	U=0.020 μW		
				31.6228 μW	U=18 nW		
				100 μW	U=18 nW		
316.2278 μW	U=60 nW						
1 mW	U=61 nW						
3.1623 mW	U=0.58 μW						
10 mW	U=0.58 μW						
31.6228 mW	U=5.8 μW						
100 mW	U=5.8 μW						
Power Reference	1 mW, 50 MHz	U _{rel} =0.34%					




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
		Output VSWR		1~1.5,50 MHz	$U=0.028$			
		Timebase		10 MHz	$U=0.092$ kHz			
		Power Measurement (With Sensor)		-30 dBm,50 MHz	$U=0.18$ dB			
				0 dBm,50 MHz	$U=0.022$ dB			
		Normal Instrumentation Measurement		1 μ W~1 mW, 50 MHz	$U_{rel}=0.38\%$			
18	*Peak-Average Power Meter	Time-base Frequency	Peak-Average Power Meter Test Procedure CAL-CHA008-02	10 MHz	$U_{rel}=1 \times 10^{-9}$			
		Power Reference Level		1 mW, 50 MHz	$U_{rel}=0.23\%$			
		Absolute Power		63 μ W~100 μ W	$U_{rel}=0.11\%$			
				100 μ W~320 μ W	$U_{rel}=0.064\%$			
				320 μ W~3.2 mW	$U_{rel}=0.03\%$			
				3.2 mW~100 mW	$U_{rel}=0.01\%$			
		Rise/Fall Time			0.8 ns~13 ns			$U=0.25$ ns
		Power Reference Output VSWR			1~1.5,50 MHz			$U=0.019$
		Instrumentation Linearity			-1%~1%, -10 dBm~-6 dBm			$U=0.064\%$



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-1%~1%, -6 dBm~8 dBm	U=0.042%		
				-1%~1%, 8 dBm~20 dBm	U=0.028%		
19	*Handheld RF/MW Power Meter	Noise	Handheld RF/MW Power Meter Test Procedure CAL-CHA008-03	-70 dBm~-60 dBm, 10 MHz~6 GHz	U=0.53 dB		
		SWR		1~1.5, 10 MHz~1.5 GHz	U=0.007		
				1~1.5, 1.5 GHz~6 GHz	U=0.014		
		Linearity		-40 dBm~-39 dBm, 50 MHz	U=0.039 dB		
				-39 dBm~-14 dBm, 50MHz	U=0.035 dB		
				-14 dBm~6 dBm, 50MHz	U=0.028 dB		
		Frequency Response		-0.3 dB~0.3 dB, 0 dBm, 10 MHz~6 GHz	U=0.068 dB		
Power	0dBm, 50MHz	U=0.016dB					
20	*Termination RF/MW Power Meter (Indicator)	Power Reference	Termination RF/MW Power Meter Test Procedure CAL-CHA008-04	1 mW, 50 MHz、1GHz	U=0.006 mW		
		Instrument Plus Power Sensor Linearity		9.02 mW ~ 35.88 mW, 50 MHz	U _{rel} =0.053%		
				35.88 mW ~ 45.17 mW, 50MHz	U _{rel} =0.038%		
				45.17 mW ~ 56.89 mW, 50MHz	U _{rel} =0.071%		
				56.89 mW ~ 71.58 mW, 50MHz	U _{rel} =0.036%		




No. CNAS L0640

第 123 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				71.58 mW ~ 90.16 mW, 50MHz	$U_{rel}=0.045\%$		
21	*Graphics Power Meter (Indicator)	Instrumentation Measurement	Graphics Power Meter Test Procedure CAL-CHA008-05	6.991 dBm, Range - 11.822 dBm ~ 6.996 dBm	$U=0.001$ dB		
				-11.839 dBm, Range - 11.822 dBm ~ 6.996 dBm	$U=0.005$ dB		
				-11.843 dBm, Range - 25.771 dBm ~ -11.822 dBm	$U=0.002$ dB		
				-25.628 dBm, Range - 25.771 dBm ~ -11.822 dBm	$U=0.005$ dB		
				-25.868 dBm, Range - 41.808 dBm ~ -25.864 dBm	$U=0.003$ dB		
				-41.907 dBm, Range - 41.808 dBm ~ -25.864 dBm	$U=0.005$ dB		
				-41.915 dBm, Range - 57.814 dBm ~ -41.806 dBm	$U=0.005$ dB		
				-57.956 dBm, Range - 57.814 dBm ~ -41.806 dBm	$U=0.008$ dB		
				-57.952 dBm, Range - 61.727 dBm ~ -57.805 dBm	$U=0.009$ dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Linearity		-61.809 dBm, Range -61.727 dBm ~ -57.805 dBm	U=0.012 dB		
				-0.028 dB~0.028 dB, Range -11.822 dBm ~ 6.996 dBm	U=0.005 dB		
				-0.028 dB~0.028 dB, Range -11.822 dBm ~ 6.996 dBm to Range -25.771 dBm ~ -11.822 dBm	U=0.005 dB		
				-0.028 dB~0.028 dB, Range -25.771 dBm ~ -11.822 dBm	U=0.004 dB		
				-0.028 dB~0.028 dB, Range -25.771 dBm ~ -11.822 dBm to Range -41.808 dBm ~ -25.864 dBm	U=0.004 dB		
				-0.028 dB~0.028 dB, Range -41.808 dBm ~ -25.864 dBm	U=0.004 dB		
				-0.028 dB~0.028 dB, Range -41.808 dBm ~ -25.864 dBm to Range -57.814 dBm ~ -41.806 dBm	U=0.005 dB		
				-0.028 dB~0.028 dB, Range -57.814 dBm ~ -41.806 dBm	U=0.005 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-0.075 dB~0.075 dB, Range -57.814 dBm ~ -41.806 dBm to Range -61.727 dBm ~ -57.805 dBm	U=0.010 dB		
				-0.23 dB~0.23 dB, Range -61.727 dBm ~ -57.805 dBm	U=0.007 dB		
				50 MHz	U=98 Hz		
				1 mW, 50 MHz	U=0.0034 mW		
				1~1.5,50 MHz	U=0.029		
22	*Thermistor Power Meter (Indicator)	Recorder Output Voltage	Thermistor Power Meter Test Procedure CAL-CHA008-06	1000 mV, 1 mW Range	U=1.6 mV		
		948.8 mV, 1 mW Range		U=0.82 mV			
		1000 mV, 0.01 mW		U=6.5 mV			
		948.8 mV, 0.03 mW		U=7.2 mV			
		1000 mV, 0.1 mW		U=3.8 mV			
		948.8 mV, 0.3 mW		U=3.8 mV			
		1000 mV, 1 mW		U=3.8 mV			
		948.8 mV, 3 mW		U=3.4 mV			
		1000 mV, 10 mW		U=3.7 mV			
		Meter Accuracy		-0.5 div~0.5 div, Range: 0.01 mW	U=0.33 div		
Meter Scale							




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-0.5 div~0.5 div, Range: 0.03 mW	U=0.49 div		
				-0.5 div~0.5 div, Range: 0.1 mW	U=0.19 div		
				-0.5 div~0.5 div, Range: 0.3 mW	U=0.22 div		
				-0.5 div~0.5 div, Range: 1 mW	U=0.19 div		
				-0.5 div~0.5 div, Range: 3 mW	U=0.22 div		
				-0.5 div~0.5 div, Range: 10 mW	U=0.19 div		
		Calibration Factor		989 mV, Cal factor: 89%	U=2.3 mV		
				978 mV, Cal factor: 90%	U=2.5 mV		
				967 mV, Cal factor: 91%	U=2.6 mV		
				957 mV, Cal factor: 92%	U=2.3 mV		
				946 mV, Cal factor: 93%	U=2.3 mV		
				935 mV, Cal factor: 94%	U=2.3 mV		
				926 mV, Cal factor: 95%	U=2.3 mV		
				916 mV, Cal factor: 96%	U=2.5 mV		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				907 mV, Cal factor: 97%	U=6.6 mV		
				897 mV, Cal factor: 98%	U=2.8 mV		
				889 mV, Calfactor: 99%	U=3.0 mV		
				880 mV, Calfactor: 100%	U=2.8 mV		
		Meter Linearity		-0.5 div~0.5 div, Range: 3 mW, Power Level: 1 mW	U=0.16 div		
				-0.5 div~0.5 div, Range: 3 mW, Power Level: 2 mW	U=0.18 div		
				-0.5 div~0.5 div, Range: 3 mW, Power Level: 3 mW	U=0.21 div		
		Zero Carryover		-5 mV~5 mV, Range: 0.01 mW	U=0.80 mV		
				-5 mV~5 mV, Range: 0.03 mW	U=0.64 mV		
				-5 mV~5 mV, Range: 0.1 mW	U=0.24 mV		
				-5 mV~5 mV, Range: 0.3 mW	U=0.13 mV		
				-5 mV~5 mV, Range: 1 mW	U=0.13 mV		
				-5 mV~5 mV, Range: 3 mW	U=0.06 mV		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-M	-5 mV~5 mV, Range: 10 mW	$U=0.06$ mV		
		Power Reference Level		1 mW, 50 MHz	$U_{rel}=0.23\%$		
		Power Reference Output		1~1.5,50 MHz	$U=0.019$		
		Absolute Power		-30 dBm~-27 dBm	$U=0.29$ dB		
				-27 dBm~-23 dBm	$U=0.10$ dB		
				-23 dBm~-20 dBm	$U=0.60$ dB		
				-20 dBm~-14 dBm	$U=0.12$ dB		
				-14 dBm~-8 dBm	$U=0.033$ dB		
				-8 dBm~-3 dBm	$U=0.014$ dB		
				-3 dBm~10 dBm	$U=0.00065$ dB		
23	*Voltage/Power Meter Indicator	Power REF Level	Voltage/Power Meter Test Procedure CAL-CHA008-07	1 mW, 50 MHz	$U=0.0057$ mW		
		Analog Outputs		10 mV~3 V	$U=0.001$ V		
		DC Measuring		10 mV	$U=0.74$ μV		
				1 V	$U=0.00050$ V		
				-1 V	$U=0.00050$ V		
				10 V	$U=0.00050$ V		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Measuring		100 V	$U=0.005 V$		
				400 V	$U=0.05 V$		
				1 mV, Frequency: 100 kHz	$U=0.0051 mV$		
				10 mV, Frequency: 100 kHz	$U=0.010 mV$		
				100 mV, Frequency: 100 kHz	$U=0.063 mV$		
				1 V, Frequency: 100 kHz	$U=0.00052 V$		
				10 V, Frequency: 100 kHz	$U=0.0013 V$		
				24	*Power Sensor	Calibration Factor	Power Sensor Test Procedure CAL-CHA009-01
1 $\mu W \sim 100 mW$, 5 MHz $\sim 8 GHz$	$U=0.54%$						
1 $\mu W \sim 100 mW$, 8 GHz $\sim 11 GHz$	$U=0.57%$						
1 $\mu W \sim 100 mW$, 11 GHz $\sim 13 GHz$	$U=0.63%$						
1 $\mu W \sim 100 mW$, 13 GHz $\sim 16 GHz$	$U=0.79%$						
1 $\mu W \sim 100 mW$, 16 GHz $\sim 17 GHz$	$U=0.84%$						
1 $\mu W \sim 100 mW$, 17 GHz $\sim 18 GHz$	$U=1.0%$						
1 $\mu W \sim 100 mW$, 18 GHz $\sim 26.5 GHz$	$U=1.6%$						




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		VSWR		1 μW~100 mW, 26.5 GHz~33 GHz	U=1.9%		
				1 μW~100 mW, 33 GHz~35 GHz	U=2.2%		
				1 μW~100 mW, 35 GHz~38 GHz	U=2.3%		
				1 μW~100 mW, 38 GHz~39 GHz	U=2.5%		
				1 μW~100 mW, 39 GHz~41 GHz	U=2.6%		
				1 μW~100 mW, 41 GHz~44 GHz	U=2.7%		
				1 μW~100 mW, 44 GHz~50 GHz	U=2.9%		
				1~2, Frequency: 9 kHz~5 MHz	U=0.0061		
				1~2, Frequency: 5 MHz~100 MHz	U=0.0054		
				1~2, Frequency: 100 MHz~2 GHz	U=0.0035		
				1~2, Frequency: 2 GHz~10 GHz	U=0.0053		
				1~2, Frequency: 10 GHz~20 GHz	U=0.0074		
				1~2, Frequency: 20 GHz~33 GHz	U=0.0098		
				1~2, Frequency: 33 GHz~40 GHz	U=0.013		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Linearity		1~2, Frequency: 40 GHz ~ 50 GHz	U=0.022		
				-3%~3%, 24 dBm~27 dBm, Frequency: 50 MHz	U=0.55%		
				-3%~3%, 11 dBm~24 dBm, Frequency: 50 MHz	U=0.13%		
				-3%~3%, 8 dBm ~ 11 dBm, Frequency: 50 MHz	U=0.09%		
				-3%~3%, 5 dBm~8 dBm, Frequency: 50 MHz	U=0.07%		
				-3%~3%, -1 dBm~5 dBm, Frequency: 50 MHz	U=0.05%		
				-3%~3%, -10 dBm~-1 dBm, Frequency: 50 MHz	U=0.09%		
				-3%~3%, -21 dBm~-10 dBm, Frequency: 50 MHz	U=0.15%		
				-3%~3%, -30 dBm~-21 dBm, Frequency: 50 MHz	U=0.18%		
				-3%~3%, -36 dBm~-30dBm, Frequency: 50 MHz	U=0.40%		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-3%~3%,-37 dBm~-36 dBm, Frequency: 50 MHz	$U=0.48\%$		
		Internal Calibration Accuracy		1 mW, 50 MHz	$U_{rel}=0.24\%$		
		Sensor Rise and Fall Time		1 ns~13 ns, 50 MHz~40 GHz	$U=0.89\text{ ns}$		
		Power Accuracy		-20 dBm, 50 MHz	$U=0.033\text{ dB}$		
				-20 dBm, 100 MHz	$U=0.036\text{ dB}$		
				-20 dBm, 20 MHz, 30 MHz, 250 MHz	$U=0.037\text{ dB}$		
				-20 dBm, 15 MHz, 500 MHz	$U=0.038\text{ dB}$		
				-20 dBm, 750 MHz, 1 GHz, 1.25 GHz, 1.5 GHz, 1.75 GHz, 2 GHz, 2.25 GHz	$U=0.039\text{ dB}$		
				-20 dBm, 2.5 GHz, 2.75 GHz, 3 GHz, 3.25 GHz, 3.5 GHz, 3.75 GHz	$U=0.040\text{ dB}$		
				-20 dBm, 4 GHz, 4.25 GHz	$U=0.041\text{ dB}$		
				-20 dBm, 10 MHz, 4.5 GHz	$U=0.042\text{ dB}$		
				-20 dBm, 4.75 MHz	$U=0.043\text{ dB}$		
				-20 dBm, 5 GHz, 5.25 GHz, 5.5 GHz, 5.75 GHz, 6 GHz, 7.75 GHz, 8 GHz	$U=0.044\text{ dB}$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm, 6.25 GHz, 6.5 GHz, 6.75 GHz, 7 GHz, 7.25 GHz, 7.5 GHz, 8.25 GHz, 8.5 GHz	U=0.045 dB		
				-20 dBm, 8.75 GHz, 9 GHz, 9.5 GHz, 9.75 GHz, 13.75 GHz, 14.25 GHz, 14.75 GHz	U=0.046 dB		
				-20 dBm, 9.25 GHz, 10 GHz, 10.25 GHz, 10.5 GHz, 10.75 GHz, 11 GHz, 11.25 GHz, 11.5 GHz, 11.75 GHz, 12 GHz, 13 GHz, 14 GHz, 14.5 GHz, 15 GHz, 15.25 GHz	U=0.047 dB		
				-20 dBm, 12.25 GHz, 12.5 GHz, 12.75 GHz, 13.25 GHz, 15.5 GHz, 15.75 GHz	U=0.048 dB		
				-20 dBm, 16 GHz, 16.25 GHz	U=0.050 dB		
				-20 dBm, 16.5 GHz, 16.75 GHz, 17 GHz	U=0.052 dB		
				-20 dBm, 17.25 GHz	U=0.053 dB		
				-20 dBm, 17.5 GHz	U=0.054 dB		
				-20 dBm, 17.75 GHz	U=0.055 dB		
				-20 dBm, 18 GHz	U=0.056 dB		
				0 dBm, 50 MHz	U=0.032 dB		




在线扫码获取验证

No. CNAS L0640

第 134 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dBm, 20 MHz,30 MHz,100 MHz	U=0.036 dB		
				0 dBm, 15 MHz	U=0.037 dB		
				0 dBm, 250 MHz,500 MHz	U=0.038 dB		
				0 dBm, 750 MHz,1 GHz,1.25 GHz,1.75 GHz,2 GHz,2.25 GHz	U=0.039 dB		
				0 dBm,1.5 GHz,2.5 GHz,2.75 GHz,3 GHz,3.25 GHz,3.5 GHz,3.75 GHz	U=0.040 dB		
				0 dBm, 4 GHz,4.25 GHz	U=0.041 dB		
				0 dBm, 10 MHz,4.5 GHz	U=0.042 dB		
				0 dBm,4.75 GHz	U=0.043 dB		
				0 dBm,5 GHz,5.5 GHz,5.75 GHz,6 GHz,7.75 GHz,8 GHz	U=0.044 dB		
				0 dBm,5.25 GHz,6.25 GHz,6.5 GHz,6.75 GHz,7 GHz,7.25 GHz,7.5 GHz,8.25 GHz,8.5 GHz	U=0.045 dB		
				0 dBm,8.75 GHz,9 GHz,9.5 GHz,9.75 GHz	U=0.046 dB		
				0 dBm, 9.25 GHz,10 GHz	U=0.047 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-M	0 dBm,10.25 GHz,10.5 GHz,10.75 GHz,11 GHz,11.25 GHz,11.5 GHz,11.75 GHz,12 GHz,13 GHz,13.25 GHz,14 GHz,14.25 GHz,14.5 GHz,14.75 GHz,15 GHz	U=0.048 dB		
				0 dBm,12.25 GHz,12.5 GHz,12.75 GHz,13.5 GHz,15.25 GHz	U=0.049 dB		
				0 dBm, 15.5 GHz,15.75 GHz	U=0.050 dB		
				0 dBm, 16 GHz,16.25 GHz	U=0.051 dB		
				0 dBm, 16.75 GHz	U=0.053 dB		
				0 dBm, 16.5 GHz,17 GHz,17.25 GHz	U=0.054 dB		
				0 dBm, 17.5 GHz	U=0.055 dB		
				0 dBm, 17.75 GHz	U=0.056 dB		
				0 dBm, 18 GHz	U=0.057 dB		
				25	*Attenuator		
0 dB,300 kHz~50 MHz	U=0.041 dB						
0 dB,50 MHz~500 MHz	U=0.018 dB						
0 dB,500 MHz~1 GHz	U=0.021 dB						
0 dB,1 GHz~2 GHz	U=0.020 dB						

CHINA NATIONAL ACCREDITATION CENTER FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE




在线扫码获取验证

No. CNAS L0640

第 136 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB, 2 GHz~12 GHz	U=0.035dB		
				0 dB, 12 GHz~12.4 GHz	U=0.040 dB		
				0 dB, 12.4 GHz~16 GHz	U=0.041 dB		
				0 dB, 16 GHz~18 GHz	U=0.042 dB		
				0 dB, 18 GHz~20 GHz	U=0.045 dB		
				0 dB, 20 GHz~26.5 GHz	U=0.076 dB		
				0 dB, 26.5 GHz~40 GHz	U=0.084 dB		
				0 dB, 40 GHz~43 GHz	U=0.17 dB		
				0 dB, 43 GHz~50 GHz	U=0.16 dB		
				1 dB~12 dB, 1 kHz	U=0.002 dB		
				1 dB~12 dB, 300 kHz~50 MHz	U=0.066 dB		
				1 dB~12 dB, 50 MHz~100 MHz	U=0.026 dB		
				1 dB~12 dB, 100 MHz~500 MHz	U=0.020 dB		
				1 dB~12 dB, 500 MHz~1 GHz	U=0.022 dB		
				1 dB~12 dB, 1 GHz~2 GHz	U=0.023 dB		
				1 dB~12 dB, 2 GHz~8 GHz	U=0.034 dB		
				1 dB~12 dB, 8 GHz~12 GHz	U=0.040 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				1 dB~12 dB,12 GHz~12.4 GHz	U=0.060 dB		
				1 dB~12 dB,12.4 GHz~16 GHz	U=0.044 dB		
				1 dB~12 dB,16 GHz~20 GHz	U=0.045 dB		
				1 dB~12 dB,20 GHz~26.5 GHz	U=0.085 dB		
				1 dB~12 dB,26.5 GHz~40 GHz	U=0.094 dB		
				1 dB~12 dB,40 GHz~50 GHz	U=0.13 dB		
				13 dB~20 dB,1 kHz	U=0.003 dB		
				13 dB~20 dB,300 kHz~50 MHz	U=0.090 dB		
				13 dB~20 dB,50 MHz~100 MHz	U=0.035 dB		
				13 dB~20 dB,100 MHz~500 MHz	U=0.025 dB		
				13 dB~20 dB,500 MHz~2 GHz	U=0.029 dB		
				13 dB~20 dB,2 GHz~8 GHz	U=0.041 dB		
				13 dB~20 dB,8 GHz~12 GHz	U=0.047 dB		
				13 dB~20 dB,12 GHz~12.4 GHz	U=0.065 dB		
				13 dB~20 dB,12.4 GHz~18 GHz	U=0.051 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				13 dB~20 dB,18 GHz~20 GHz	U=0.053 dB		
				13 dB~20 dB,20 GHz~26.5 GHz	U=0.092 dB		
				13 dB~20 dB,26.5 GHz~40 GHz	U=0.12 dB		
				13 dB~20 dB,40 GHz~50 GHz	U=0.16 dB		
				21 dB~30 dB,1 kHz	U=0.003 dB		
				21 dB~30 dB,300 kHz~50 MHz	U=0.094 dB		
				21 dB~30 dB,50 MHz~100 MHz	U=0.045 dB		
				21 dB~30 dB,100 MHz~2 GHz	U=0.036 dB		
				21 dB~30 dB,2 GHz~8 GHz	U=0.062 dB		
				21 dB~30 dB,8 GHz~12 GHz	U=0.060 dB		
				21 dB~30 dB,12 GHz~12.4 GHz	U=0.067 dB		
				21 dB~30 dB,12.4 GHz~16 GHz	U=0.068 dB		
				21 dB~30 dB,16 GHz~18 GHz	U=0.071 dB		
				21 dB~30 dB,18 GHz~20 GHz	U=0.072 dB		
				21 dB~30 dB,20 GHz~26.5 GHz	U=0.12 dB		




No. CNAS L0640

第 139 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				21 dB~30 dB,26.5 GHz~40 GHz	U=0.13 dB		
				21 dB~30 dB,40 GHz~50 GHz	U=0.18 dB		
				31 dB~40 dB,1 kHz	U=0.008 dB		
				31 dB~40 dB,300 kHz~50 MHz	U=0.11 dB		
				31 dB~40 dB,50 MHz~100 MHz	U=0.045 dB		
				31 dB~40 dB,100 MHz~500 MHz	U=0.037 dB		
				31 dB~40 dB,500 MHz~1 GHz	U=0.040 dB		
				31 dB~40 dB,1 GHz~2 GHz	U=0.041 dB		
				31 dB~40 dB,2 GHz~8 GHz	U=0.063 dB		
				31 dB~40 dB,8 GHz~12 GHz	U=0.062 dB		
				31 dB~40 dB,12 GHz~12.4 GHz	U=0.069 dB		
				31 dB~40 dB,12.4 GHz~16 GHz	U=0.071 dB		
				31 dB~40 dB,16 GHz~18 GHz	U=0.070 dB		
				31 dB~40 dB,18 GHz~20 GHz	U=0.076 dB		
				31 dB~40 dB,20 GHz~26.5 GHz	U=0.12 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				31 dB~40 dB,26.5 GHz~40 GHz	U=0.17 dB		
				31 dB~40 dB,40 GHz~43 GHz	U=0.32 dB		
				31 dB~40 dB,43 GHz~50 GHz	U=0.31 dB		
				41 dB~50 dB,1 kHz	U=0.009 dB		
				41 dB~50 dB,300 kHz~50 MHz	U=0.10 dB		
				41 dB~50 dB,50 MHz~100 MHz	U=0.040 dB		
				41 dB~50 dB,100 MHz~500 MHz	U=0.030 dB		
				41 dB~50 dB,500 MHz~2 GHz	U=0.053 dB		
				41 dB~50 dB,2 GHz~8 GHz	U=0.087 dB		
				41 dB~50 dB,8 GHz~12 GHz	U=0.85 dB		
				41 dB~50 dB,12 GHz~12.4 GHz	U=0.095 dB		
				41 dB~50 dB,12.4 GHz~16 GHz	U=0.097 dB		
				41 dB~50 dB,16 GHz~18 GHz	U=0.098 dB		
				41 dB~50 dB,18 GHz~20 GHz	U=0.11 dB		
				41 dB~50 dB,20 GHz~26.5 GHz	U=0.17 dB		




No. CNAS L0640

第 141 页 共 209

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				41 dB~50 dB,26.5 GHz~40 GHz	U=0.24 dB		
				41 dB~50 dB,40 GHz~43 GHz	U=0.42 dB		
				41 dB~50 dB,43 GHz~50 GHz	U=0.40 dB		
				51 dB~60 dB,1 kHz	U=0.015 dB		
				51 dB~60 dB,300 kHz~50 MHz	U=0.12 dB		
				51 dB~60 dB,50 MHz~100 MHz	U=0.046 dB		
				51 dB~60 dB,100 MHz~500 MHz	U=0.032 dB		
				51 dB~60 dB,500 MHz~2 GHz	U=0.054 dB		
				51 dB~60 dB,2 GHz~8 GHz	U=0.088 dB		
				51 dB~60 dB,8 GHz~12 GHz	U=0.086 dB		
				51 dB~60 dB,12 GHz~12.4 GHz	U=0.097 dB		
				51 dB~60 dB,12.4 GHz~18 GHz	U=0.098 dB		
				51 dB~60 dB,18 GHz~20 GHz	U=0.11 dB		
				51 dB~60 dB,20 GHz~26.5 GHz	U=0.17 dB		
				51 dB~60 dB,26.5 GHz~40 GHz	U=0.24 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				51 dB~60 dB,40 GHz~43 GHz	U=0.57 dB		
				51 dB~60 dB,43 GHz~50 GHz	U=0.44 dB		
				61 dB~70 dB,300 kHz~50 MHz	U=0.15 dB		
				61 dB~70 dB,50 MHz~100 MHz	U=0.062 dB		
				61 dB~70 dB,100 MHz~500 MHz	U=0.050 dB		
				61 dB~70 dB,500 MHz~2 GHz	U=0.072 dB		
				61 dB~70 dB,2 GHz~12 GHz	U=0.12 dB		
				61 dB~70 dB,12 GHz~12.4 GHz	U=0.13 dB		
				61 dB~70 dB,12.4 GHz~18 GHz	U=0.14 dB		
				61 dB~70 dB,18 GHz~20 GHz	U=0.15 dB		
				61 dB~70 dB,20 GHz~26.5 GHz	U=0.24 dB		
				61 dB~70 dB,26.5 GHz~40 GHz	U=0.30 dB		
				61 dB~70 dB,40 GHz~43 GHz	U=0.44 dB		
				61 dB~70 dB,43 GHz~50 GHz	U=0.43 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				71 dB~80 dB,300 kHz~50 MHz	U=0.15 dB		
				71 dB~80 dB,50 MHz~100 MHz	U=0.065 dB		
				71 dB~80 dB,100 MHz~500 MHz	U=0.051 dB		
				71 dB~80 dB,500 MHz~2 GHz	U=0.081 dB		
				71 dB~80 dB,2 GHz~12 GHz	U=0.12 dB		
				71 dB~80 dB,12 GHz~18 GHz	U=0.14 dB		
				71 dB~80 dB,18 GHz~20 GHz	U=0.15 dB		
				71 dB~80 dB,20 GHz~26.5 GHz	U=0.24 dB		
				71 dB~80 dB,26.5 GHz~40 GHz	U=0.34 dB		
				71 dB~80 dB,40 GHz~43 GHz	U=0.58 dB		
				71 dB~80 dB,43 GHz~50 GHz	U=0.57 dB		
				81 dB~90 dB,300 kHz~50 MHz	U=0.18 dB		
				81 dB~90 dB,50 MHz~100 MHz	U=0.068 dB		
				81 dB~90 dB,100 MHz~500 MHz	U=0.069 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				81 dB~90 dB,500 MHz~1 GHz	U=0.096 dB		
				81 dB~90 dB,1 GHz~2 GHz	U=0.11 dB		
				81 dB~90 dB,2 GHz~12 GHz	U=0.16 dB		
				81 dB~90 dB,12 GHz~12.4 GHz	U=0.17 dB		
				81 dB~90 dB,12.4 GHz~18 GHz	U=0.18 dB		
				81 dB~90 dB,18 GHz~20 GHz	U=0.19 dB		
				81 dB~90 dB,20 GHz~26.5 GHz	U=0.31 dB		
				81 dB~90 dB,26.5 GHz~40 GHz	U=0.45 dB		
				81 dB~90 dB,40 GHz~43 GHz	U=0.68 dB		
				81 dB~90 dB,43 GHz~50 GHz	U=0.60 dB		
				91 dB~100 dB,300 kHz~50 MHz	U=0.18 dB		
				91 dB~100 dB,50 MHz~100 MHz	U=0.067 dB		
				91 dB~100 dB,100 MHz~500 MHz	U=0.068 dB		
				91 dB~100 dB,500 MHz~1 GHz	U=0.097 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				91 dB~100 dB,1 GHz~2 GHz	U=0.12 dB		
				91 dB~100 dB,2 GHz~8 GHz	U=0.18 dB		
				91 dB~100 dB,8 GHz~12 GHz	U=0.23 dB		
				91 dB~100 dB,12 GHz~12.4 GHz	U=0.30 dB		
				91 dB~100 dB,12.4 GHz~16 GHz	U=0.24 dB		
				91 dB~100 dB,16 GHz~20 GHz	U=0.27 dB		
				91 dB~100 dB,20 GHz~26.5 GHz	U=0.45 dB		
				91 dB~100 dB,26.5 GHz~40 GHz	U=0.47 dB		
				91 dB~100 dB,40 GHz~43 GHz	U=0.75 dB		
				91 dB~100 dB,43 GHz~50 GHz	U=0.74 dB		
				101 dB,300kHz~50 MHz	U=0.19 dB		
				101 dB,50 MHz~100 MHz	U=0.069 dB		
				101 dB,100 MHz~500 MHz	U=0.071 dB		
				101 dB,500 MHz~1 GHz	U=0.097 dB		
				101 dB,1 GHz~2 GHz	U=0.14 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				101 dB, 2 GHz~8 GHz	U=0.20 dB		
				101 dB, 8 GHz~12 GHz	U=0.23 dB		
				101 dB, 12 GHz~12.4 GHz	U=0.38 dB		
				101 dB, 12.4 GHz~16 GHz	U=0.24 dB		
				101 dB, 16 GHz~20 GHz	U=0.27 dB		
				101 dB, 20 GHz~26.5 GHz	U=0.45 dB		
				101 dB, 26.5 GHz~40 GHz	U=0.47 dB		
				101 dB, 40 GHz~43 GHz	U=0.76 dB		
				101 dB, 43 GHz~50 GHz	U=0.74 dB		
				102 dB~110 dB, 300 kHz~50 MHz	U=0.19 dB		
				102 dB~110 dB, 50 MHz~100 MHz	U=0.069 dB		
				102 dB~110 dB, 100 MHz~500 MHz	U=0.071 dB		
				102 dB~110 dB, 500 MHz~1 GHz	U=0.097 dB		
				102 dB~110 dB, 1 GHz~2 GHz	U=0.14 dB		
				102 dB~110 dB, 2 GHz~8 GHz	U=0.20 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				102 dB~110 dB,8 GHz~12 GHz	U=0.23 dB		
				102 dB~110 dB,12 GHz~12.4 GHz	U=0.38 dB		
				102 dB~110 dB,12.4 GHz~16 GHz	U=0.24 dB		
				102 dB~110 dB,16 GHz~18 GHz	U=0.27 dB		
				111 dB~120 dB,300 kHz~50 MHz	U=0.22 dB		
				111 dB~120 dB,50 MHz~100 MHz	U=0.070 dB		
				111 dB~120 dB,100 MHz~500 MHz	U=0.072 dB		
				111 dB~120 dB,500 MHz~1 GHz	U=0.098 dB		
				111 dB~120 dB,1 GHz~2 GHz	U=0.19 dB		
				111 dB~120 dB,2 GHz~8 GHz	U=0.27 dB		
				111 dB~120 dB,8 GHz~12 GHz	U=0.33 dB		
				111 dB~120 dB,12 GHz~16 GHz	U=0.41 dB		
				111 dB~120 dB,16 GHz~18 GHz	U=0.42 dB		
				30 dB ,50 MHz	U=0.011 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		VSWR		1~1.5,300 kHz~2 GHz,50 Ω	U=0.001		
				1~1.5,2 GHz~40 GHz,50 Ω	U=0.009		
				1~1.5,40 GHz~50 GHz,50 Ω	U=0.022		
				1.5~2.33,300 kHz~10 MHz,50 Ω	U=0.010		
				1.5~2.33,10 MHz~50 MHz,50 Ω	U=0.008		
				1.5~2.33,50 MHz~2 GHz,50 Ω	U=0.009		
				1.5~2.33,2 GHz~8 GHz,50 Ω	U=0.013 dB		
				1.5~2.33,8 GHz~40 GHz,50 Ω	U=0.014 dB		
				1.5~2.33,40 GHz~50 GHz,50 Ω	U=0.034		
				2.33~4,300 kHz~10 MHz,50 Ω	U=0.028		
				2.33~4,10 MHz~50 MHz,50 Ω	U=0.015		
				2.33~4,50 MHz~2 GHz,50 Ω	U=0.017		
				2.33~4,2 GHz~8 GHz,50 Ω	U=0.028		
				2.33~4,8 GHz~20 GHz,50 Ω	U=0.030		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				2.33~4,20 GHz~40 GHz,50 Ω	U=0.033		
				2.33~4,40 GHz~50 GHz,50 Ω	U=0.063		
				4~10,300 kHz~10 MHz,50 Ω	U=0.081		
				4~10,10 MHz~50 MHz,50 Ω	U=0.044		
				4~10,50 MHz~2 GHz,50 Ω	U=0.043		
				4~10,2 GHz~20 GHz,50 Ω	U=0.082		
				4~10,20 GHz~40 GHz,50 Ω	U=0.11		
				4~10,40 GHz~50 GHz,50 Ω	U=0.16		
				1~1.5,300 kHz~1.3 GHz,75 Ω	U=0.005		
				1~1.5,1.3 GHz~3 GHz,75 Ω	U=0.009		
				1.5~10,300 kHz~1.3 GHz,75 Ω	U=0.008		
				1.5~10,1.3 GHz~3 GHz,75 Ω	U=0.014		
				26	*Frequency Counter		
				-20 dBm~0 dBm,500 MHz~1 GHz	U=0.24 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~0 dBm,1 GHz~20 GHz	U=0.28 dB		
				-20 dBm~0 dBm,20 GHz~26.5 GHz	U=0.38 dB		
				-20 dBm~0 dBm,26.5 GHz~40 GHz	U=0.43 dB		
				-17 dBm~0 dBm,40 GHz~46 GHz	U=0.62 dB		
		Frequency		0.1 Hz ~ 1 Hz	U=6.7×10 ⁻¹¹ Hz		
				1 Hz ~ 10 Hz	U=8.8×10 ⁻¹⁰ Hz		
				10 Hz ~ 100 Hz	U=4.7×10 ⁻⁹ Hz		
				100 Hz ~ 12.4 GHz	U _{rel} =2.4×10 ⁻¹¹		
		Signal Period		4 ns~1 μs,-1 V~-0.1V, 0.1V~1 V	U=0.0026 ns		
		Pulse Width		5 ns,-1 V~-0.1V,0.1V~1 V	U=0.0028 ns		
		Peak Volts		-1 V~-0.1V,0.1V~1 V,2 MHz	U=0.02 V		
		Time Interval		49.9 ns~150 ns,1 V	U=0.0037 ns		
		Trigger Level		0.01 V~0.08 V,1 MHz	U=0.0045 V		
		Termination Resistor		50 Ω	U _{rel} =0.004%		
				1 MΩ	U _{rel} =0.001%		
		Time-base Output Frequency		10 MHz	U _{rel} =1×10 ⁻¹¹		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rise Fall Time		1 ns~30 ns , 1 kHz ~ 20MHz	$U=0.92$ ns		
27	*Digital LCR Meter	Capacitance	Digital LCR Meter Test Procedure CAL-CHA012-01	1 pF, 1 kHz~1 MHz	$U_{rel}=0.012\%$		
				1 pF, 1 MHz~13 MHz	$U_{rel}=0.050\%$		
				10 pF, 1 kHz~2 MHz	$U_{rel}=0.0046\%$		
				10 pF, 2 MHz~13 MHz	$U_{rel}=0.011\%$		
				100 pF, 1 kHz~300 kHz	$U_{rel}=0.0086\%$		
				100 pF, 300 kHz~2 MHz	$U_{rel}=0.007\%$		
				100 pF, 2 MHz~13 MHz	$U_{rel}=0.016\%$		
				1000 pF, 20 Hz~100 kHz	$U_{rel}=0.0081\%$		
				1000 pF, 100 kHz~2 MHz	$U_{rel}=0.010\%$		
				1000 pF, 2 MHz~13 MHz	$U_{rel}=0.032\%$		
				10 nF, 100 Hz~120 Hz	$U_{rel}=0.0056\%$		
				10 nF, 1 kHz	$U_{rel}=0.0044\%$		
				10 nF, 1 kHz~100 kHz	$U_{rel}=0.0055\%$		
				100 nF, 120 Hz~100 kHz	$U_{rel}=0.0042\%$		
1 μ F, 100 Hz~120 Hz	$U_{rel}=0.0069\%$						
1 μ F, 1 kHz	$U_{rel}=0.0044\%$						
1 μ F, 10 kHz	$U_{rel}=0.0064\%$						




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				1 μ F, 100 kHz	$U_{rel}=0.0096\%$		
				10 μ F, 20 Hz~100 Hz	$U_{rel}=0.0082\%$		
				10 μ F, 100 Hz~1 kHz	$U_{rel}=0.0052\%$		
				10 μ F, 10 kHz	$U_{rel}=0.017\%$		
				10 μ F, 100 kHz	$U_{rel}=0.070\%$		
		Dissipation		0.0~0.2, 1 pF~1000 pF, 20 Hz~1 MHz	$U=0.00002$		
				0.0~0.024, 1 pF~1000 pF, 1 MHz~3 MHz	$U=0.000025$		
				0.0~0.024, 1 pF~1000 pF, 3 MHz~13 MHz	$U=0.000075$		
				0.0~0.024, 0.01 μ F~1 μ F, 120 Hz~100 kHz	$U=0.00002$		
				0.0~0.024, 10 μ F, 120 Hz~1 kHz	$U=0.000065$		
				0.0~0.024, 10 μ F, 10 kHz	$U=0.00029$		
				0.0~0.024, 10 μ F, 100 kHz	$U=0.00072$		
		Signal Level		0.005 V~0.012 V, 20 Hz~300 kHz	$U_{rel}=0.065\%$		
				0.005 V~0.012 V, 300 kHz~1 MHz	$U_{rel}=1.4\%$		
				0.012 V~1 V, 20 Hz~ 300 kHz	$U_{rel}=0.012\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.012 V~1 V,300 kHz~1 MHz	$U_{rel}=0.68\%$		
				1 V~20 V,20 Hz~300 kHz	$U_{rel}=0.054\%$		
				1 V~20 V,300 kHz~1 MHz	$U_{rel}=1.2\%$		
				0.1 V~10 V,1 MHz~30 MHz	$U_{rel}=1.0\%$		
		Signal Level Monitor		0.005 V~0.012 V,20 Hz~300 kHz	$U_{rel}=0.064\%$		
				0.005 V~0.012 V,300 kHz~1 MHz	$U_{rel}=1.4\%$		
				0.012 V~1 V,20 Hz~300 kHz	$U_{rel}=0.011\%$		
				0.012 V~1 V,300 kHz~1 MHz	$U_{rel}=0.68\%$		
				1 V~20 V,20 Hz~300 kHz	$U_{rel}=0.034\%$		
				1 V~20 V,300 kHz~1 MHz	$U_{rel}=1.05\%$		
				0.1 V~10 V,1 MHz~30 MHz	$U_{rel}=1.0\%$		
			DC Bias Level		0.1 V~10 V	$U_{rel}=0.0011\%$	
				10 V~40 V	$U_{rel}=0.0013\%$		
		DC Bias Current		0.02 mA~2 mA	$U_{rel}=0.0062\%$		
				2 mA~20 mA	$U_{rel}=0.0089\%$		
				20 mA~100 mA	$U_{rel}=0.019\%$		




在线扫码获取验证

No. CNAS L0640

第 154 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Impedance	ilac-MRA	10 mΩ ,1 kHz	$U_{rel}=0.8\%$		
				100 mΩ ,DC~1 kHz	$U_{rel}=0.028\%$		
				1 Ω ,120 Hz~1 kHz	$U_{rel}=0.0087\%$		
				10 Ω ,120 Hz~1 kHz	$U_{rel}=0.0045\%$		
				100 Ω ,20 Hz~1 MHz	$U_{rel}=0.03\%$		
				1 kΩ ,20 Hz~1 MHz	$U_{rel}=0.017\%$		
				100 kΩ ,DC~100 kHz	$U_{rel}=0.012\%$		
		Frequency		20 Hz~30 MHz	$U_{rel}=3.8 \times 10^{-9}$		
28	*RF Impedance Analyzer	Frequency	RF Impedance Analyer Test Procedure CAL-CHA013-01	1 MHz~3 GHz	$U_{rel}=1.3 \times 10^{-9}$		
		Amplitude		-40 dBm~-19 dBm,1 MHz	$U=0.18$ dB		
				-15 dBm,1 MHz	$U=0.11$ dB		
				-10 dBm,1 MHz	$U=0.18$ dB		
				-7.9 dBm~-6 dBm,1 MHz	$U=0.31$ dB		
				-5 dBm,1 MHz	$U=0.18$ dB		
				-4 dBm~-1 dBm,1 MHz	$U=0.31$ dB		
				0 dBm~1 dBm,1 MHz	$U=0.16$ dB		
				-40 dBm~-35 dBm,10 MHz~300 MHz	$U=0.27$ dB		
				-30 dBm~1 dBm,10 MHz~300 MHz	$U=0.19$ dB		




No. CNAS L0640

第 155 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-40 dBm~1 dBm,500 MHz	U=0.15 dB		
				-40 dBm~7 dBm,1 GHz	U=0.16 dB		
				-40 dBm~-35 dBm,1.5 GHz~1.8 GHz	U=0.29 dB		
				-30 dBm~1 dBm,1.5 GHz~1.8 GHz	U=0.19 dB		
				-40 dBm,2 GHz	U=0.34 dB		
				-35 dBm~1 dBm,2 GHz	U=0.24 dB		
				-40 dBm~-35 dBm,2.2 GHz~3 GHz	U=0.33 dB		
				-30 dBm~1 dBm,2.2 GHz~3 GHz	U=0.25 dB		
		Impedance		0 Ω,-40 dBm,1 MHz	U=0.014 Ω		
				0 Ω,-33 dBm~13 dBm,1 MHz	U=0.0025 Ω		
				0 Ω,-40 dBm,10 MHz	U=0.0099 Ω		
				0 Ω,-33 dBm~13 dBm,10 MHz	U=0.0035 Ω		
				0 Ω,-40 dBm~13 dBm,100 MHz	U=0.014 Ω		
				0 Ω,-40 dBm~13 dBm,200 MHz	U=0.025 Ω		
				0 Ω,-40 dBm~13 dBm,300 MHz~1 GHz	U=0.03 Ω		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 Ω, -40 dBm~13 dBm, 1.3~1.8 GHz	U=0.1 Ω		
				0 Ω, -40 dBm~13 dBm, 2~3 GHz	U=0.2 Ω		
				50 Ω, -40 dBm~13 dBm, 1 MHz~100 MHz	U=0.05 Ω		
				50 Ω, -40 dBm~13 dBm, 200 MHz~300 MHz	U=0.1 Ω		
				50 Ω, -40 dBm~13 dBm, 500 MHz	U=0.15 Ω		
				50 Ω, -40 dBm~13 dBm, 600 MHz~1 GHz	U=0.2 Ω		
				50 Ω, -40 dBm~-20 dBm, 1.3 GHz~1.8 GHz	U=0.3 Ω		
				50 Ω, -15 dBm~5 dBm, 1.3 GHz~1.8 GHz	U=0.2 Ω		
				50 Ω, -40 dBm~1 dBm, 2 GHz~3 GHz	U=0.41 Ω		
				0 Ω ~ 22 k Ω, -40 dBm, 1 MHz	U=2.8 k Ω		
				0 k Ω ~ 22 k Ω, -33 dBm~13 dBm, 1 MHz	U=35 Ω		
				0 k Ω ~ 22 k Ω, -40 dBm, 10 MHz	U=24 Ω		
				0 k Ω ~ 22 k Ω, -33 dBm~13 dBm, 10 MHz	U=3.5 Ω		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 100 MHz	U=0.34 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 200 MHz	U=0.2 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 300 MHz	U=0.12 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 500 MHz	U=0.068 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 600 MHz	U=0.049 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 800 MHz	U=0.068 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 1 GHz	U=0.22 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 1.6 GHz	U=0.6 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 1.8 GHz	U=0.23 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 2.0 GHz	U=0.08 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 2.2 GHz	U=0.18 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 2.4 GHz	U=0.55 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 2.6 GHz	U=2.2 Ω		
				0 kΩ ~ 22 kΩ, -40 dBm ~ 13 dBm, 3 GHz	U=1 Ω		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 Ω ~ 156 Ω, -40 dBm, 1 MHz	U=0.002 Ω		
				0 Ω ~ 156 Ω, -33 dBm ~ 13 dBm, 1 MHz	U=0.00019 Ω		
				0 Ω ~ 156 Ω, -40 dBm, 10 MHz	U=0.004 Ω		
				0 Ω ~ 156 Ω, -33 dBm ~ 13 dBm, 10 MHz	U=0.0018 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 100 MHz	U=0.019 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 200 MHz	U=0.045 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 300 MHz	U=0.073 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 500 MHz	U=0.26 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 600 MHz	U=0.62 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 800 MHz ~ 1 GHz	U=0.43 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 1.3 GHz	U=0.13 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 1.6 GHz	U=0.08 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 1.8 GHz	U=0.26 Ω		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 2 GHz	U=0.91 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 2.4 GHz	U=1.8 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 2.6 GHz	U=0.6 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 2.8 GHz	U=0.24 Ω		
				0 Ω ~ 156 Ω, -40 dBm ~ 13 dBm, 3 GHz	U=0.3 Ω		
		Admittance		0.01 μS ~ 10000 μS, -15 dBm ~ 5 dBm, 1 MHz	U=0.016 μS		
				0.01 μS ~ 10000 μS, -33 dBm ~ -20 dBm, 1 MHz	U=0.37 μS		
				0.01 μS ~ 10000 μS, -40 dBm, 1 MHz	U=8.6 μS		
				0.01 μS ~ 10000 μS, -15 dBm ~ 5 dBm, 10 MHz	U=0.16 μS		
				0.01 μS ~ 10000 μS, -33 dBm ~ -20 dBm, 10 MHz	U=0.32 μS		
				0.01 μS ~ 10000 μS, -40 dBm, 10 MHz	U=4.7 μS		
				0.01 μS ~ 10000 μS, -15 dBm ~ 5 dBm, 100 MHz	U=1.6 μS		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.01 μS~10000 μS,-33 dBm~20 dBm,100 MHz	U=2.2 μS		
				0.01 μS~10000 μS,-40 dBm,100 MHz	U=6.3 μS		
				0.01 μS~10000 μS,-33 dBm~5 dBm,200 MHz	U=6.3 μS		
				0.01 μS~10000 μS,-40 dBm,200 MHz	U=8.6 μS		
				0.01 μS~10000 μS,-33 dBm~5 dBm,300 MHz	U=9.4 μS		
				0.01 μS~10000 μS,-40 dBm,300 MHz	U=13 μS		
				0.01 μS~10000 μS,-40 dBm~5 dBm,500 MHz	U=16 μS		
				0.01 μS~10000 μS,-40 dBm~5 dBm,600 MHz	U=19 μS		
				0.01 μS~10000 μS,-40 dBm~5 dBm,800 MHz	U=25 μS		
				0.01 μS~10000 μS,-40 dBm~5 dBm,1 GHz	U=32 μS		
				0.01 μS~10000 μS,-40 dBm~5 dBm,1.3 GHz	U=42 μS		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 1.6 GHz	$U=58 \mu$ S		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 1.8 GHz	$U=66 \mu$ S		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 2 GHz	$U=91 \mu$ S		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 2.2 GHz	$U=0.10$ mS		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 2.4 GHz	$U=0.11$ mS		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 2.6 GHz	$U=0.13$ mS		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 2.8 GHz	$U=0.15$ mS		
				0.01 μ S ~ 10000 μ S, -40 dBm ~ 5 dBm, 3 GHz	$U=0.16$ mS		
				29	*High Resistance Meter		
10 M Ω	$U_{rel}=0.18\%$						
100 M Ω	$U_{rel}=0.19\%$						
1 G Ω	$U_{rel}=0.19\%$						
10 G Ω	$U_{rel}=0.37\%$						
100 G Ω	$U_{rel}=0.62\%$						
DC Current	100 pA	$U=0.34$ pA					
	1 nA	$U=0.0019$ nA					



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	10 nA	U=0.017 nA		
				100 nA	U=0.13 nA		
				1 μA	U=0.00073 μA		
				10 μA	U=0.0073 μA		
				100 μA	U=0.075 μA		
		DC Voltage		10mV	U=0.06mV		
				10 V	U=0.00068 V		
				25 V	U=0.0010 V		
				50 V	U=0.0015 V		
				100 V	U=0.0025 V		
				200 V	U=0.0051 V		
				201 V	U=0.0055 V		
				250 V	U=0.0065 V		
				500 V	U=0.013 V		
				1000 V	U=0.034 V		
30	*Function Generator	Time-base Frequency	Function Generator Test Procedure CAL-CHA015-01	10 MHz	U=0.0021 Hz	合格评定 国家认可 CNAS 认可证书专用章	
		DC Voltage		10 V	U=0.00028 V		
				0.5 V	U=0.0000073 V		
				10mV	U=0.06mV		

No. CNAS L0640


第 163 页 共 209



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Amplitude		-10 V	$U=0.00059\text{ V}$		
				5.7 V~10 V,40 Hz~100 Hz	$U_{rel}=0.035\%$		
				0.018 V~1 V,100 Hz~1 kHz	$U_{rel}=0.011\%$		
				1 V~3.5 V,100 Hz~1 kHz	$U_{rel}=0.018\%$		
				3.5 V~5.7 V,100 Hz~1 kHz	$U_{rel}=0.044\%$		
				5.7 V~7 V,100 Hz~1 kHz	$U_{rel}=0.011\%$		
		Flatness		-10 dBm~0 dBm,100 kHz~5 MHz	$U=0.027\text{ dB}$		
				-10 dBm~ 0 dBm,5 MHz~10 MHz	$U=0.031\text{ dB}$		
				-10 dBm~ 0 dBm,10 MHz~22 MHz	$U=0.033\text{ dB}$		
				-10 dBm~ 0 dBm,22 MHz~30 MHz	$U=0.034\text{ dB}$		
				0 dBm~ 10 dBm, 100 kHz~5 MHz	$U=0.026\text{ dB}$		
				0 dBm~ 10 dBm, 5 MHz~60 MHz	$U=0.030\text{ dB}$		
				0 dBm~10 dBm, 6 MHz~90 MHz	$U=0.030\text{ dB}$		
				0 dBm~ 10 dBm, 90 MHz~120 MHz	$U=0.032\text{ dB}$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	10 dBm~25 dBm, 100 kHz~8 MHz	U=0.009 dB		
				10 dBm~25 dBm, 8 MHz~40 MHz	U=0.011 dB		
				10 dBm~25 dBm, 40 MHz~80 MHz	U=0.024 dB		
				10 dBm~25 dBm, 80 MHz~120 MHz	U=0.033 dB		
		Harmonic Distortion		-100 dBc~0 dBc, 100 Hz~40 kHz	U=1.2 dB		
				-100 dBc~0 dBc, 40 kHz~60 kHz	U=1.4 dB		
				-100 dBc~0 dBc, 60 kHz~80 kHz	U=1.8 dB		
				-100 dBc~0 dBc, 80 kHz~500 kHz	U=1.2 dB		
				-100 dBc~0 dBc, 500 kHz~5 MHz	U=1.0 dB		
				-100 dBc~0 dBc, 5 MHz~75 MHz	U=1.6 dB		
31	*Noise Source	ENR	Noise Source Test Procedure CAL-CHA016-01	1 dB~30 dB, 1 GHz	U=0.09 dB		
				1 dB~30 dB, 4 GHz	U=0.091 dB		
				1 dB~30 dB, 2 GHz	U=0.093 dB		
				1 dB~30 dB, 10 MHz, 100 MHz, 7 GHz	U=0.10 dB		
				1 dB~30 dB, 3 GHz, 6 GHz, 8 GHz	U=0.11 dB		

No. CNAS L0640


第 165 页 共 209



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				1 dB~30 dB, 5 GHz, 19 GHz, 25 GHz	U=0.13 dB		
				1 dB~30 dB, 9 GHz, 10 GHz, 11 GHz, 12 GHz, 13 GHz, 14 GHz, 20 GHz, 21 GHz, 24 GHz, 26 GHz, 26.5 GHz	U=0.14 dB		
				1 dB~30 dB, 15 GHz, 18 GHz	U=0.15 dB		
				1 dB~30 dB, 16 GHz, 17 GHz, 22 GHz	U=0.16 dB		
				1 dB~30 dB, 23 GHz	U=0.17 dB		
		Reflection		0.0001~1, 10 MHz, 100 MHz, 1 GHz, 2 GHz	U=0.0030		
				0.0001~1, 3 GHz, 4 GHz, 14 GHz	U=0.0049		
				0.0001~1, 5 GHz, 6 GHz, 7 GHz, 8 GHz, 9 GHz, 10 GHz, 11 GHz, 12 GHz, 16 GHz	U=0.0050		
				0.0001~1, 13 GHz, 15 GHz, 17 GHz, 18 GHz	U=0.0051		
				0.0001~1, 19 GHz	U=0.0052		
				0.0001~1, 20 GHz, 23 GHz	U=0.0053		
				0.0001~1, 24 GHz	U=0.0054		
				0.0001~1, 22 GHz, 25 GHz	U=0.0055		



No. CNAS L0640


The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Reflection(So urce On- Source Off)	ilac-M	0.0001~1, 21 GHz, 26 GHz, 26.5 GHz	U=0.0056		
				0.0001~1, 1 GHz, 2 GHz, 5 GHz, 10 GHz, 15 GHz, 18 GHz	U=0.000058		
				0.0001~1, 100 MHz, 3 GHz, 4 GHz, 6 GHz, 11 GHz, 12 GHz	U=0.00013		
				0.0001~1, 9 GHz, 13 GHz, 14 GHz, 16 GHz, 17 GHz	U=0.00016		
				0.0001~1, 7 GHz	U=0.0002		
				0.0001~1, 8 GHz	U=0.00023		
				0.0001~1, 10 MHz	U=0.00054		
				32	*Noise Figure Analyzer	Reference Frequency	Noise Figure Analyzer Test Procedure CAL-CHA017-01
Input VSWR	1.01~2, 10 MHz~0.5 GHz	U=0.09					
	1.01~2, 0.5 GHz~1.0 GHz	U=0.11					
	1.01~2, 1.0 GHz~1.5 GHz	U=0.13					
	1.01~2, 1.5 GHz~3 GHz	U=0.11					
	1.01~2, 3 GHz~6.7 GHz	U=0.05					
	1.01~2, 6.7 GHz~20 GHz	U=0.11					



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Noise Source Supply		1.01~2,20 GHz~26.5 GHz	$U=0.21$		
				0.01 V~1V	$U_{rel}=0.03\%$		
		Frequency		28 V	$U=0.31$ mV		
				14 MHz~26.5 GHz,Resolution	$U=0.010$ MHz		
				Bandwidth:4 MHz			
				1 GHz,Resolution	$U=0.002$ MHz		
		Instrument Noise Figure		0 dB~20 dB ,10 MHz~26.5 GHz	$U=0.014$ dB		
		Measurement Jitter		0 dB~0.5 dB ,1 GHz	$U=0.006$ dB		
Noise Figure Range And Measurement	0 dB~22 dB,50 MHz	$U=0.014$ dB					
Gain Measurement Uncertainty	-20 dB~40 dB	$U=0.010$ dB					
33	*Multiplexer Module	Relay Contact Resistance	Multiplexer Module Test Procedure CAL-CHA018-01	0.05 Ω ~2 Ω	$U=0.00054$ Ω		
		Thermocouple Reference Junction		-209 $^{\circ}\text{C}$	$U=0.13$ $^{\circ}\text{C}$		
				-150 $^{\circ}\text{C}$ ~450 $^{\circ}\text{C}$	$U=0.09$ $^{\circ}\text{C}$		
				600 $^{\circ}\text{C}$ ~750 $^{\circ}\text{C}$	$U=0.10$ $^{\circ}\text{C}$		
				1000 $^{\circ}\text{C}$	$U=0.12$ $^{\circ}\text{C}$		
				1199 $^{\circ}\text{C}$	$U=0.13$ $^{\circ}\text{C}$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
34	*Power Splitter	Reflection Coefficient	Power Splitter Test Procedure CAL-CHA024-01	0~1, 50 MHz~2 GHz	U=0.0036		
				0~1, 2.1 GHz~3 GHz	U=0.0049		
				0~1, 3.1 GHz~6 GHz	U=0.0067		
				0~1, 6.1 GHz~8 GHz	U=0.0071		
				0~1, 8.1 GHz~13.5 GHz	U=0.0082		
				0~1, 13.6 GHz~18 GHz	U=0.0085		
				0~1, 18.1 GHz~26.5 GHz	U=0.01		
				0~1, 26.6 GHz~33 GHz	U=0.014		
				0~1, 33.1 GHz~40 GHz	U=0.015		
		0~1, 40.1 GHz~50 GHz	U=0.018				
		0 dB~8 dB, 50 MHz	U=0.031 dB				
		0 dB~8 dB, 50.1 MHz~500 MHz	U=0.027 dB				
		0 dB~8 dB, 500.1 MHz~1 GHz	U=0.03 dB				
		0 dB~8 dB, 1.1 GHz~2 GHz	U=0.026 dB				
		0 dB~8 dB, 2.1 GHz~3 GHz	U=0.028 dB				
		0 dB~8 dB, 3.1 GHz~6 GHz	U=0.034 dB				
		0 dB~8 dB, 6.1 GHz~9.5 GHz	U=0.066 dB				




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB~8 dB, 9.6 GHz~12.5 GHz	U=0.064 dB		
				0 dB~8 dB, 12.6 GHz~16 GHz	U=0.062 dB		
				0 dB~8 dB, 16.1 GHz~25 GHz	U=0.06 dB		
				0 dB~8 dB, 25.1 GHz~26.5 GHz	U=0.059 dB		
				0 dB~8 dB, 26.6 GHz~28 GHz	U=0.094 dB		
				0 dB~8 dB, 28.1 GHz~30 GHz	U=0.092 dB		
				0 dB~8 dB, 30.1 GHz~32 GHz	U=0.09 dB		
				0 dB~8 dB, 32.1 GHz~34.5 GHz	U=0.088 dB		
				0 dB~8 dB, 34.6 GHz~37.3 GHz	U=0.086 dB		
				0 dB~8 dB, 37.4 GHz~40 GHz	U=0.084 dB		
				0 dB~8.5 dB, 40.1 GHz~43.5 GHz	U=0.14 dB		
				0 dB~8.5 dB, 43.6 GHz~50 GHz	U=0.13 dB		
		Port 2 To Port 3 Tracking		-1 dB~1 dB, 50 MHz~3 GHz	U=0.030 dB		
				-1 dB~1 dB, 3.1 GHz~4 GHz	U=0.033 dB		
				-1 dB~1 dB, 4.1 GHz~5 GHz	U=0.030 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-1 dB~1 dB, 5.1 GHz~6.5 GHz	U=0.033 dB		
				-1 dB~1 dB, 6.6 GHz~8 GHz	U=0.043 dB		
				-1 dB~1 dB, 8.1 GHz~9 GHz	U=0.030 dB		
				-1 dB~1 dB, 9.1 GHz~12.5 GHz	U=0.047 dB		
				-1 dB~1 dB, 12.6 GHz~16 GHz	U=0.049 dB		
				-1 dB~1 dB, 16.1 GHz~18 GHz	U=0.048 dB		
				-1 dB~1 dB, 18.1 GHz~20 GHz	U=0.049 dB		
				-1 dB~1 dB, 20.1 GHz~23.5 GHz	U=0.047 dB		
				-1 dB~1 dB, 23.6 GHz~26.5 GHz	U=0.066 dB		
				-1 dB~1 dB, 26.6 GHz~32 GHz	U=0.070 dB		
				-1 dB~1 dB, 32.1 GHz~37 GHz	U=0.069 dB		
				-1 dB~1 dB, 37.1 GHz~40 GHz	U=0.068 dB		
				-1 dB~1 dB, 40.1 GHz~43.6 GHz	U=0.086 dB		
				-1 dB~1 dB, 43.7 GHz~47 GHz	U=0.085 dB		
				-1 dB~1 dB, 47.1 GHz~50 GHz	U=0.093 dB		
		Equivalent Source Match		0~1, 50 MHz~500 MHz	U=0.0051		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0~1, 501 MHz~2 GHz	$U=0.0041$		
				0~1, 2.1 GHz~10 GHz	$U=0.0055$		
				0~1, 10.1 GHz~14 GHz	$U=0.0073$		
				0~1, 14.1 GHz~18 GHz	$U=0.0074$		
				0~1, 18.1 GHz~22 GHz	$U=0.0075$		
				0~1, 22.1 GHz~26.5 GHz	$U=0.0074$		
				0~1, 26.6 GHz~ 40 GHz	$U=0.013$		
				0~1, 40.1 GHz~50 GHz	$U=0.015$		
35	Semiconductor Device Parameter Analyzer	DC Voltage Output	Semiconductor Device Parameter Analyzer Test Procedure CAL-CHA025-01	0.01 V~0.22 V	$U_{rel}=0.0013\%$		
				0.22 V~2.2 V	$U_{rel}=0.0009\%$		
				2.2 V~22 V	$U_{rel}=0.0008\%$		
				22 V~500 V	$U_{rel}=0.0012\%$		
		DC Current Output		1 pA~11.5 pA	$U_{rel}=0.13\%$		
				11.5 pA~115 pA	$U_{rel}=0.058\%$		
				115 pA~1.15 nA	$U_{rel}=0.031\%$		
				1.15 nA~11.5 nA	$U_{rel}=0.016\%$		
				11.5 nA~115 nA	$U_{rel}=0.0039\%$		
				115 nA~1.15 μ A	$U_{rel}=0.0021\%$		
				1.15 μ A~11.5 μ A	$U_{rel}=0.0022\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				20 μA ~ 115 μA	$U_{rel}=0.0025\%$		
				0.2 mA ~ 1.15 mA	$U_{rel}=0.0020\%$		
				2 mA ~ 11.5 mA	$U_{rel}=0.0024\%$		
				20 mA ~ 115 mA	$U_{rel}=0.0041\%$		
				0.2 A ~ 1.05 A	$U_{rel}=0.014\%$		
				1.5 A ~ 3.03 A	$U_{rel}=0.011\%$		
				3.03 A ~ 10.5 A	$U_{rel}=0.026\%$		
				10.5 A ~ 20 A	$U_{rel}=0.067\%$		
				DC Voltage Measurement	0.01 V ~ 0.22 V		
		0.22 V ~ 2.2 V	$U_{rel}=0.0009\%$				
		2.2 V ~ 22 V	$U_{rel}=0.0008\%$				
		22 V ~ 500 V	$U_{rel}=0.0012\%$				
		DC Current Measurement	1 pA ~ 11.5 pA	$U_{rel}=0.13\%$			
			11.5 pA ~ 115 pA	$U_{rel}=0.058\%$			
			115 pA ~ 1.15 nA	$U_{rel}=0.031\%$			
			1.15 nA ~ 11.5 nA	$U_{rel}=0.016\%$			
			11.5 nA ~ 115 nA	$U_{rel}=0.0039\%$			
			115 nA ~ 1.15 μA	$U_{rel}=0.0021\%$			
			1.15 μA ~ 11.5 μA	$U_{rel}=0.0022\%$			




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				20 μA ~ 115 μA	$U_{rel}=0.0025\%$		
				0.2 mA ~ 1.15 mA	$U_{rel}=0.0020\%$		
				2 mA ~ 11.5 mA	$U_{rel}=0.0024\%$		
				20 mA ~ 115 mA	$U_{rel}=0.0041\%$		
				0.2 A ~ 1.05 A	$U_{rel}=0.014\%$		
				1.5 A ~ 3.03 A	$U_{rel}=0.011\%$		
				3.03 A ~ 10.5 A	$U_{rel}=0.026\%$		
				10.5 A ~ 20 A	$U_{rel}=0.067\%$		
				Level Test	0.1 V ~ 2.2 V		
		2.2 V ~ 22 V	$U_{rel}=0.0033\%$				
		22 V ~ 50 V	$U_{rel}=0.0025\%$				
		Pulse Shape Test-Period	100 ns	$U=0.10$ ns			
			2 μs	$U=0.082$ μs			
			100 μs	$U=0.093$ μs			
			100 ms	$U=0.090$ ms			
		Pulse Shape Test-Delay	2 s	$U=0.082$ s			
			2.5 ns	$U=0.0036$ ns			
			2.5 μs	$U=0.003$ μs			
			2.5 ms	$U=0.051$ ms			




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date					
		Pulse Shape Test-Width	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	50 ns	$U=0.071$ ns							
				1 μ s	$U=0.0046$ μ s							
				50 μ s	$U=0.052$ μ s							
				50 ms	$U=0.052$ ms							
				1 s	$U=0.0041$ s							
		Pulse Shape Test-Rise/Fall Time		20 ns~100 ns	$U_{rel}=0.54\%$							
				0.35 μ s~10 μ s	$U_{rel}=0.34\%$							
				100 μ s~10 ms	$U_{rel}=0.30\%$							
		Output Impedance		50 Ω	$U_{rel}=0.002\%$							
		Pulse Overshoot		0.01 %~5.2%, +/-10 V	$U=0.29\%$							
				0.01%~5.1%, +/-20 V	$U=0.30\%$							
		36		*Pulse Pattern Generator	Frequency			Pulse Pattern Generator Test Procedure CAL-CHA026-01	0.2 Hz ~3.35 GHz including period test	$U_{rel}=8.5 \times 10^{-7}$		
									Pulse Width	100 ps~999 ns		
					1 μ s~500 ms					$U_{rel}=8.4 \times 10^{-6}$		
Pulse Delay	1 ns~500 ns		$U_{rel}=3.2 \times 10^{-5}$									
	3 μ s~500 ms		$U_{rel}=4.3 \times 10^{-5}$									
Double Pulse Delay	6 ns~100 ns		$U_{rel}=1.5 \times 10^{-4}$									
	500 ms~1 s		$U_{rel}=5.8 \times 10^{-5}$									



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Period Jitter	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	Period 0.3 ns~20 ns, including clock jitter and data jitter	U=1.9 ps		
				Period 20 ns~500 ns, including clock jitter and data jitter	U=1.4 ps		
				Width Jitter	width 50 ns~500 ns		
		Delay Jitter		Delay 50 ns~500 ns	U=1.3 ps		
		Output Levels		10 mV ~ 100 mV	U=0.82 mV		
				100 mV ~ 1 V	U=1.6 mV		
				1 V ~ 10 V	U=13 mV		
		Transition Time		50 ps ~ 1 ns	U=2.1 ps		
				1 ns ~ 500 ms	U=8 ps + 0.1%t _x		
		Pulse Aberration		10 mV ~ 200 mV, pulse level: 0.5 V~2 V	U=3 mV		
				10 mV ~ 200 mV, pulse level: 2 V~5 V	U=30 mV		
		37		*DC Electronic Load	CC Mode Current		
1 A~120 A	U _{rel} =0.01%						
CV Mode Voltage	1 V~10 V		U _{rel} =0.0017%				
	10 V~240 V		U _{rel} =0.0004%				
CR Mode Resistor	0.017 Ω~0.5 Ω		U=0.000055 Ω				
	0.5 Ω~2 Ω		U=0.0004 Ω				

No. CNAS L0640


第 176 页 共 209



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Slew		2 Ω ~ 2.5 kΩ	$U=0.0007 \Omega$		
				60 μs ~ 125 μs, 0.050 A/μs ~ 0.083 A/μs	$U=0.055 \mu s$		
				1.2 ms ~ 2 ms, 0.0025 A/μs ~ 0.0042 A/ms	$U=0.054 ms$		
		CC PARD		0.001 mA ~ 4 mA, 10kHz	$U=0.001 mA$		
		CC Mode Regulation		1 A ~ 10 A, Voltage Change: 4 V ~ 240 V	$U=0.53 mA$		
		CV Mode Regulation		5 V ~ 25 V, Current Change: 1A ~ 60A	$U=0.11 mV$		
		Power		0.6 W ~ 1.5 kW	$U_{rel}=0.012\%$		
		Current Rise/Fall Time		0.1 μs ~ 100 μs (5 V, Current Change: 3A ~ 27A)	$U=5.3 \mu s$		
38	*Wavelength Meter	Wavelength	Wavelength Meter Test Procedure CAL-CHA020-01	1550 nm	$U_{rel}=1.6 \times 10^{-7}$		
		Amplitude Accuracy		0 dBm (Wavelength: 1310 nm, 1550 nm)	$U=0.23 dB$		
39	*Optical power Meter	Optical power	Optical power Meter Test Procedure CAL-CHA021-01	10 μW (Wavelength: 1310 nm, 1550 nm)	$U_{rel}=1.9\%$		
		Linearity		10 dBm ~ -30 dBm (Wavelength : 1310 nm, 1550 nm)	$U_{rel}=0.20\%$		
				-30 dBm ~ -60 dBm (Wavelength : 1310 nm, 1550 nm)	$U_{rel}=0.21\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
40	*Optical Attenuator	Attenuation	Optical Attenuator Test Procedure CAL-CHA022-01	1 dB ~60 dB(Wavelength : 1310 nm, 1550 nm)	U=0.075 dB		
41	*Laser Source	Output Power	Laser Source Test Procedure CAL-CHA023-01	10 dBm~-60 dBm(Wavelength : 1310nm ~ 1630 nm)	U=0.11 dB		
		Wavelength		1310 nm~1630 nm	U _{rel} =3.5x10 ⁻⁷		
		Power Stability		10 dBm~-20 dBm(Wavelength : 1310 nm~1630 nm)	U=0.0026 dB		
42	*RF Amplifier	VSWR	RF Amplifier Test Procedure CAL-CHA032-01	1.001~2.2, Frequency: 100 kHz~400 MHz	U=0.021		
				1.001~2.2, Frequency: 400 MHz~1.3 GHz	U=0.030		
				10.0 dB~60.0 dB, Frequency: 100 kHz~400 MHz	U=0.060 dB		
				10.0 dB~60.0 dB, Frequency: 400 MHz~1.3 GHz	U=0.091 dB		
		Gain Flatness		10.0 dB~60.0 dB, Frequency: 1.3 GHz~8 GHz	U=0.25 dB		
				0 dB~3.0 dB, Frequency: 100 kHz~400 MHz	U=0.05 dB		
		0 dB~3.0 dB, Frequency: 400 MHz~1.3 GHz	U=0.82 dB				



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Harmonic Distortion		-100 dBc~0 dBc, Carrier Frequency: 100 MHz~400 MHz, Output Power: 0 dBm	$U=0.30$ dB		
		Gain Compression		0.0 dB~1.0 dB, Output Power: 6 dBm~16 dBm	$U=0.35$ dB		
		Reverse Isolation		-100 dB~0 dB, Frequency: 100 kHz~1.3 GHz	$U=0.30$ dB		
		Noise Figure		0.0 dB~35.0 dB, Frequency: 100 kHz~1.3 GHz	$U=0.43$ dB		
43	*RF/MW Vector Signal Generator with IQ ports	AM Distortion	RF/MW Vector Signal Generator with IQ ports Test Procedure CAL-CHA001-04	0.01%~3%, Carrier Frequency: 0.1 MHz~10 MHz, AM: 10%~90%	$U=0.03\%$		
				0.01%~3%, Carrier Frequency: 10 MHz~20 GHz, AM: 10%~90%	$U=0.04\%$		
		AM Frequency Response		0 dB~3 dB, Carrier Frequency: 100 MHz~2.7 GHz, Mod Rate: DC~10 kHz	$U=0.027$ dB		
				0 dB~3 dB, Carrier Frequency: 100 MHz~2.7 GHz, Mod Rate: 10 kHz~50 kHz	$U=0.04$ dB		




No. CNAS L0640

第 179 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB~3 dB, Carrier Frequency: 2.7 GHz~20 GHz, Mod Rate: DC~50 kHz	$U=0.13$ dB		
				0 dB~3 dB, Carrier Frequency: 100 MHz~ 20 GHz, Mod Rate: 50 kHz~100 kHz	$U=0.14$ dB		
		AM		5%~99%, Carrier Frequency: 100 kHz~ 10 MHz	$U_{rel}=0.75\%$		
				5%~20%, Carrier Frequency: 10 MHz~3 GHz	$U_{rel}=2.5\%$		
				20%~99%, Carrier Frequency: 10 MHz~3 GHz	$U_{rel}=0.5\%$		
				5%~20%, Carrier Frequency: 3 GHz~20 GHz	$U_{rel}=4.5\%$		
		Power Level		20%~99%, Carrier Frequency: 3 GHz~20 GHz	$U_{rel}=1.5\%$		
				-120 dBm~-20 dBm, Carrier Frequency: 100 kHz~250 kHz	$U=0.17$ dB		
				-20 dBm~20 dBm, Carrier Frequency: 100 kHz~250 kHz	$U=0.15$ dB		




在线扫码获取验证

No. CNAS L0640

第 180 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-120 dBm~-20 dBm,Carrier Frequency: 250 kHz~6 GHz	$U=0.11$ dB		
				-20 dBm~20 dBm,Carrier Frequency: 250 kHz~6 GHz	$U=0.063$ dB		
				-10 dBm~-120 dBm,Carrier Frequency: 6 GHz~20 GHz	$U=0.16$ dB		
				-10 dBm~20 dBm,Carrier Frequency: 6 GHz~20 GHz	$U=0.14$ dB		
		Carrier Frequency		10 kHz~20 GHz	$U_{rel}=1 \times 10^{-9}$		
		Carrier Frequency Offset with FM		1 MHz,Carrier Frequency: 50 MHz~6 GHz	$U=0.064$ kHz		
		FM Frequency Response		-3 dB~3 dB, Mod Rate: 30 Hz~400 kHz	$U=0.15$ dB		
		FM Distortion		0.01 %~1%, Carrier Frequency: 2 MHz~6 GHz,Mod Rate : 1 kHz~10 kHz, Mod index 0.2~100	$U=0.017\%$		
		FM Deviation		Deviation : 200 Hz~400 kHz,Carrier Frequency: 100 kHz~20 GHz	$U_{rel}=1.0\%$		




在线扫码获取验证

No. CNAS L0640

第 181 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Frequency Setting Time		0.0001 ms~2 ms,Carrier Frequency: 100 kHz~6 GHz	U=0.0061 ms		
		Harmonic Spurious		-70 dBc~-40 dBc,Carrier Frequency: 10 kHz~3 GHz	U=0.49 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 3 GHz~6.6 GHz	U=1.9 dB		
				-70 dBc~-40 dBc,Carrier Frequency: 6.6 GHz~22 GHz	U=2.5 dB		
		IQ Error Vector		0.01%~1.5%,Carrier Frequency: 100 kHz~6 GHz	U=0.74%		
		IQ Input Impedance		1.01~1.2,Frequency: 1 MHz~50 MHz	U=0.004		
				1.01~1.2,Frequency: 50 MHz~1000 MHz	U=0.007		
		IQ Residual Carrier And Leakage		-70 dBc~-40 dBc,Carrier Frequency: 100 kHz~6 GHz	U=0.17 dBc		
		Level Frequency Response		0 dB~1 dB,Frequency:9 kHz~20 MHz,-20 dBm~20 dBm	U=0.09 dB		
				0 dB~1 dB,Frequency:20 MHz~50 MHz,-20 dBm~20 dBm	U=0.13 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				0 dB~1 dB, Frequency: 50 MHz~2 GHz, -20 dBm ~ 20 dBm	$U=0.10$ dB		
				0 dB~1 dB, Frequency: 2 GHz~6 GHz, -20 dBm ~ 20 dBm	$U=0.13$ dB		
				0 dB~1 dB, Frequency: 6 GHz~20 GHz, -20 dBm ~ 20 dBm	$U=0.17$ dB		
		Level Setting Time		3 ms~5 ms, Amplitude: -130 dBm~20 dBm, Frequency: 300 kHz~6 GHz	$U=0.0061$ ms		
		Modulation Generator Distortion		0.001%~0.1%, Frequency: 100 Hz~20 kHz	$U=0.016\%$		
				0.001%~0.1%, Frequency: 20 kHz~100 kHz	$U=0.032\%$		
		Modulation Generator Frequency Response		(0~2) dB, Frequency: 10 Hz~100 kHz	$U=0.047$ dB		
				(0~2) dB, Frequency: 100 kHz~10 MHz	$U=0.15$ dB		
		Modulation Generator Output Level		3 mV, Frequency: 1kHz	$U_{rel}=1.7\%$		
				10 mV, Frequency: 1kHz	$U_{rel}=0.56\%$		
				30 mV, Frequency: 1kHz	$U_{rel}=0.24\%$		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				100 mV, Frequency: 1kHz	$U_{rel}=0.12\%$		
				100 mV~300 mV, Frequency: 1kHz	$U_{rel}=0.19\%$		
				300 mV~1 V, Frequency: 1kHz	$U_{rel}=0.11\%$		
				1 V~3 V, Frequency: 1kHz	$U_{rel}=0.26\%$		
		Modulation Generator Frequency		1 kHz~1 MHz	$U=0.00053$ Hz		
		Non-Harmonic Spurious		-70 dBc~-40 dBc, Frequency: 10 kHz~1 MHz	$U=1.0$ dB		
				-70 dBc~-40 dBc, Frequency: 1 MHz~3 GHz	$U=0.21$ dB		
				-70 dBc~-40 dBc, Frequency: 3 GHz~6.6 GHz	$U=0.74$ dB		
				-70 dBc~-40 dBc, Frequency: 6.6 GHz~13.2 GHz	$U=1.3$ dB		
				-70 dBc~-40 dBc, Frequency: 13.2 GHz~18 GHz	$U=1.7$ dB		
				-70 dBc~-40 dBc, Frequency: 18 GHz~20 GHz	$U=1.7$ dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Output Impedance		1.01~1.9, Frequency: 300 kHz~6 GHz	$U=0.021$		
		Phase Modulation Distortion		0~1%, Carrier Frequency: 100 kHz~6000 MHz, Phase deviation: (1~90) rad	$U=0.03\%$		
		Phase Modulation Flatness		0 dB~3 dB, Mod Rate: DC~100 kHz	$U=0.16$ dB		
		Phase Modulation on AM		0.0001 rad~0.1 rad, Carrier Frequency: 100 kHz~6 GHz	$U=0.0038$ rad		
		Phase Modulation		0.7 rad~100 rad, 100 kHz~6 GHz	$U_{rel}=1\%$		
				0.3 rad~0.7 rad, 100 kHz~6 GHz	$U_{rel}=3\%$		
				1.2 rad~100 rad, 6 GHz~20 GHz	$U_{rel}=3\%$		
		Pulse Modulation Video Crosstalk		-30 dBc, Carrier Frequency: 100 kHz~6 GHz	$U=0.17$ dB		
		Pulse Modulation On/Off Ratio		50 dB ~ 200 dB, Carrier Frequency: 100 kHz~6 GHz	$U=0.16$ dB		
				50 dB ~ 200 dB, Carrier Frequency: 6 GHz~20 GHz	$U=0.17$ dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Pulse Modulation: Rise/Fall Time		(0.1~12) ns, Carrier Frequency: 500 MHz~20 GHz	U=0.2 ns		
				(0.1~12) ns, Carrier Frequency: 100 kHz~500 MHz	U=0.4 ns		
		Residual FM		(0.01~4) Hz, Carrier Frequency: 1 GHz	U=0.61 Hz		
		Residual AM		0.00001%~0.02%, Carrier Frequency: 100 MHz~6 GHz	U=0.00061%		
				0.00001%~0.02%, Carrier Frequency: 100 kHz~100 MHz	U=0.0012%		
		Single-Sideband Phase Noise		-110 dBc/Hz~-70 dBc/Hz, Carrier Frequency: 10 MHz~20 GHz, 10 Hz~1 MHz offset	U=2.3 dB		
		Sub-Harmonic Spurious		-20 dBc~-80 dBc, Carrier Frequency: 100 kHz~2 GHz	U=0.5 dB		
				-20 dBc~-80 dBc, Carrier Frequency: 2 GHz~4.5 GHz	U=1.9 dB		
				-20 dBc~-80 dBc, Carrier Frequency: 4.5 GHz~20 GHz	U=2.5 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Synchronous AM with FM		0.0001%~0.1%,Carrier Frequency: 10 MHz~2 GHz	U=0.0014%		
		Synchronous PHIM with AM		0.0001%~0.1%,Carrier Frequency: 2 GHz~6 GHz	U=0.0016%		
		Wideband Noise		-120 dBc/Hz~-160 dBc/Hz,Carrier Frequency: 20 MHz~3 GHz	U=0.49 dBc/Hz		
		Wideband Noise		-120 dBc/Hz~-160 dBc/Hz,Carrier Frequency: 3 GHz~6 GHz	U=1.8 dBc/Hz		
		Sensitivity Of Electronic Tuning Of Internal Reference Frequency		Sensitivity $5 \times 10^{-9}/V \sim 5 \times 10^{-8}/V$, Frequency: 10 MHz	U= $6.1 \times 10^{-11}/V$		
		Additive white Gaussian noise		0 dB~0.1 dB, system bandwidth = symbolrate, symbol rate < 4 MHz, -24 dB < C/N < 30 dB and 0 dB < crest factor < 12 dB	U=0.0061 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		I/Q modulation RF frequency response		-2 dB~ 2 dB, Frequency: 300 kHz~ 6 GHz, external/internal I/Q inputs	U=0.17 dB		
		I/Q outputs DC voltage		10 mV ~ 0.3 V	U=0.061 mV		
				0.3 V ~ 1 V	U=0.084 mV		
				1 V ~ 3.6 V	U=0.23 mV		
		I/Q Outputs SFDR		50 dBc ~ 80 dBc, 2 MHz ~ 20 MHz offset	U=0.17 dB		
		D/A Converter Interpolation Spectra Of Aliasing Filter		-100 dB ~ -70 dB, Internal baseband, 10MHz ~ 100 MHz offset	U=0.49 dB		
		Internal Reference		10 MHz	$U_{rel} = 1 \times 10^{-11}$		
		I/Q Output Frequency Response		-2 dB~ 2 dB, IQ Output Frequency: 1 MHz ~ 60 MHz, reference 1 MHz	U=0.085 dB		
		I/Q Outputs Wideband Noise		-170 dBc~ -150 dBc, Generated Signal 10 dBm, Offset > 30 MHz	U=0.17 dB		
		IQ Adjacent Channel Power For 3GPP FDD		-70 dB ~ -40 dB, Bandwidth 5 MHz ~ 10 MHz	U=0.16 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		IQ Output EVM for WCDMA 3GPP		0.01%~5%, Carrier Frequency: 100 MHz~6 GHz	U=0.09%		
44	*Wideband Wireless Communications Test Set	Reference Frequency Amplitude	Broadband Wireless Communications Test Set Test Procedure CAL-CHA004-03	1.4 V~20 V, 1 MHz~80 MHz	U=0.021 V		
		Reference Frequency Accuracy		10 MHz	U=0.014 Hz		
		Audio Generator THD+N		0.01%~ 0.1%, 100 mV~20 V, 100 Hz~10 kHz	U=0.007%		
		Audio Generator THD		0.01%~ 0.1%, 100 mV~20 V, 100 Hz~10 kHz	U=0.0036%		
		Audio Generator Level		0.1 V ~5 V, 20 Hz ~10 kHz	U=0.1 mV		
		Audio Generator Output Impedance		10 mV ~100 mV, 20 Hz ~10 kHz	U=0.06 mV		
		Audio Generator Max Output Current		1 mV ~10 mV, 20 Hz ~10 kHz	U=0.002 mV		
				50 Ω ~75 Ω, 1 kHz ~10 kHz	U=0.006 Ω		
		5 mA~20 mA, 20 Hz ~10 kHz	U=0.22 mA				



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Audio Analyzer Frequency Response		-40 mV~-10mV, 10mV~40 mV,2 V Input,50 Hz ~20 kHz	$U_{rel}=0.023\%$		
		Audio Analyzer Inherent Distortion		0.01%~0.1%,0.5 V~5 V,100 Hz~10 kHz	$U=0.007\%$		
		Audio Analyzer THD		0.01%~0.1%,0 mV~5000 mV,100 Hz~10 kHz	$U=0.0036\%$		
		Audio Analyzer Linearity		0.01%~2%,0.02 V ~5 V,10 kHz~20 kHz	$U=0.024\%$		
		RX SSB Phase Noise		0.01%~2%,1 mV ~0.02 V,10 kHz~20 kHz	$U=0.18\%$		
		RX Average Noise Level		-150 dBc/Hz~-50 dBc/Hz,50 MHz ~ 6000 MHz	$U=1.8$ dB		
		RX Dynamic Range		-150 dBm~-50 dBm	$U=0.9$ dB		
		RX Power Level		-150 dBc~-50 dBc	$U=0.6$ dB		
				20 dBm~-20 dBm,100 kHz~2000 MHz	$U=0.073$ dB		
				20 dBm~-20 dBm,2000 MHz~4000 MHz	$U=0.08$ dB		
				20 dBm~-20 dBm,4000 MHz~6000 MHz	$U=0.14$ dB		
				-20 dBm~-60 dBm,100 kHz~4000 MHz	$U=0.17$ dB		




No. CNAS L0640

第 190 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~-60 dBm,4000 MHz~6000 MHz	U=0.18 dB		
		RX Linearity		-2 dB~2 dB,0 dBm~-60 dBm	U=0.06 dB		
		RX Residual Response		-60 dBm~-100 dBm	U=1.0 dB		
		RX Harmonics		-20 dBc~-80 dBc	U=0.6 dB		
		RX Spurious /Image Response		-20 dBc~-80 dBc	U=0.6 dB		
		RX VSWR		1.01~4,50 MHz~6000 MHz	U=0.01		
		TX Frequency		50 MHz~6000 MHz	U _{rel} =0.12%		
		TX VSWR		1.01~4,50 MHz~6000 MHz	U=0.01		
		TX Signal To Noise Ratio		60 dB~100 dB,50 MHz~3000 MHz, Resolution Bandwidth 1 kHz	U=0.66 dB		
				60 dB~100 dB,3000 MHz~6000 MHz, Resolution Bandwidth 1 kHz	U=2.6 dB		
		TX Output Power		20 dBm~-20 dBm,100 kHz~2000 MHz	U=0.07 dB		
				20 dBm~-20 dBm,2000 MHz~4000 MHz	U=0.08 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				20 dBm~-20 dBm,4000 MHz~6000 MHz	U=0.12 dB		
				-20 dBm~-80 dBm,100 kHz~6000 MHz	U=0.13 dB		
				-80 dBm~-110 dBm,100 kHz~6000 MHz	U=0.14 dB		
				-110 dBm~-120 dBm,100 kHz~6000 MHz	U=0.18 dB		
		TX Output Linearity		0 dB ~-46 dB,380 MHz~6000 MHz	U=0.06 dB		
		TX SSB Phase Noise		-150 dBc/Hz~-50 dBc/Hz,50 MHz ~ 6000 MHz	U=0.06 dB		
		TX Harmonics		-70 dB~-30 dB,50 MHz ~ 6000 MHz	U=0.47 dB		
		TX IN-Band Spurious		-70 dB~-30 dB,50 MHz ~ 6000 MHz	U=0.66 dB		
		1XEV-DO TX RHO		0.9 ~1,870 MHz~2150 MHz	U=0.0019		
		1XEV-DO RX ACP		-90 dB~-10 dB,870 MHz~2150 MHz,0 dBm~-10 dBm	U=1.1 dB		
		1XEV-DO RX EVM		0.01%~4% rms,870 MHz~2150 MHz	U=0.1 %		
		1XEV-DO RX Frequency Error		-15 Hz~15 Hz,870 MHz~2150 MHz	U=3.1 Hz		




在线扫码获取验证

No. CNAS L0640

第 192 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		1XEV-DO IQ Imbalance		-90dB~-40 dB,870 MHz~2150 MHz	U=4.5 dB		
		1XEV-DO Carrier Feedthrough		-90dB~-40 dB, 870 MHz~2150 MHz	U=3.2 dB		
		TX Frequency Accuracy (Bluetooth)		2402 MHz ~ 2495 MHz	U=6.7 Hz		
		TX Bluetooth Level Accuracy		- 105 dBm ~ 5 dBm,2402 MHz~2495 MHz	U=0.085 dB		
		TX Carrier Suppression		- 105 dBm~5 dBm	U=0.2 dB		
		RX Bluetooth Frequency Offset		0.1 kHz~1 MHz	U=0.1 kHz		
		RX Bluetooth Level Accuracy		- 105 dBm ~ 5 dBm,2402 MHz~2495 MHz	U=0.3 dB		
		RX Bluetooth Leakage Power		-100 dB~-55 dB,2402 MHz~2495 MHz	U=0.1 dB		
		Rx Bluetooth Frequency Deviation		0.01%~1%, Deviation: 100 kHz ~ 210 kHz, modulation index 0.22~ 0.42	U=0.059%		
		Rx Bluetooth Frequency Drift		-75 kHz~75 kHz, Test pattern 10101010	U=0.23 kHz		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Rx Bluetooth RMS EVM		0.01%~3 %, PRBS pattern	U=0.43%		
		Rx Bluetooth PEAK EVM		0.01%~3 %, PRBS pattern	U=0.61%		
		Rx Bluetooth Frequency Stability Wi		Wi (-75~75) kHz, Deviation (0~160) kHz	U=0.24 kHz		
		Rx Bluetooth Frequency Stability Wo		Wo (-10~10) kHz, Deviation (0~160) kHz	U=0.17 kHz		
		C2K TX ACP		-90 dB ~ -10 dB,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.1 dB		
		C2K TX EVM		0.01%~25%,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.95%		
		C2K TX RHO		0.9~1,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.001		
		C2K RX ACP		-90 dB ~ -10 dB,0 dBm~-20 dBm,380 MHz~2800 MHz	U=1.1 dB		
		C2K RX EVM		0.01%~25%,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.09 %		
		C2K RX RHO		0.9~1,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.0012		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		C2K RX Frequency Error		0.1 Hz~10 kHz,380 MHz~2800 MHz	U=3.3 Hz		
		C2K RX IQ Imbalance		-90 dB~-40 dB,0 dBm~ -20 dBm,380 MHz~2800 MHz	U=8.3 dB		
		C2K RX Carrier Feedthrough		-90 dB~-40 dB,0 dBm~-20 dBm,380 MHz~2800 MHz	U=3.4 dB		
		EDGE RX Origin Offset		-50 dB~-10 dB,0 dBm~ -20 dBm,380 MHz~2800 MHz	U=0.48 dB		
		EDGE RX EVM		0.01%~10% ,0 dBm~- 20 dBm,380 MHz~2800 MHz	U=0.15%		
		EDGE RX ORFS		-90 dB~-10 dB,0 dBm~-20 dBm,380 MHz~2800 MHz	U=0.8 dB		
		EDGE RX Frequency Error		0.1 Hz ~10 kHz,380 MHz~2800 MHz	U=1 Hz		
		EDGE TX EVM		0.01%~10% ,0 dBm~- 20 dBm,380 MHz~2800 MHz	U=0.48%		
		TX GSM Frequency Error		-80 kHz ~80 kHz,710 MHz~1990 MHz,0 dBm~-20 dBm	U=6.1 Hz		
		TX GSM Phase Error (Peak, RMS)		-20° ~20° ,710 MHz~1990 MHz	U=0.36°		



在线扫码获取验证

No. CNAS L0640

第 195 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		TX GSM 8PSK EVM		0.01%~10%, 710 MHz~1990 MHz	U=0.58%		
		TX GSM Power		-10 dBm ~ 0 dBm, 710 MHz~1990 MHz	U=0.14 dB		
				-90 dBm ~ -10 dBm, 710 MHz~1990 MHz	U=0.17 dB		
		RX GSM Origin Offset		-90 dB ~ -40 dB, 0 dBm ~ -20 dBm, 710 MHz~1990 MHz	U=2.1 dB		
		RX GSM 8PSK EVM		0.01%~8%, 0dBm ~ -20dBm, 710MHz~1990M Hz	U=0.19%		
		RX GSM ORFS		-90 dB ~ -10 dB, 0 dBm ~ -20 dBm, 710 MHz~1990 MHz	U=0.31 dB		
		RX GSM Frequency Error		0.01 Hz~10 kHz, 710 MHz~1990 MHz	U=0.85 Hz		
		RX GSM Phase Error (Peak)		-4.0° pk~4.0° peak, 710 MHz~1990 MHz	U=0.1°		
		RX GSM Phase Error(RMS)		-1.0° rms~1.0° rms, 710 MHz~1990 MHz	U=0.26°		
		TX LTE EVM		0.01%~8% , 0 dBm ~ -20 dBm, 380 MHz~3800 MHz	U=0.45%		
		TX LTE Power		0 dBm ~ -20 dBm, 380 MHz~3800 MHz	U=0.14 dB		



No. CNAS L0640

第 196 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		RX LTE ACP		-80 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 3800 MHz	$U=0.99$ dB		
		RX LTE CHP		0 dBm ~ -20 dBm, 380 MHz ~ 3800 MHz	$U=0.05$ dB		
		RX LTE EVM		0.01% ~ 8%, 0 dBm ~ -20 dBm, 380 MHz ~ 3800 MHz	$U=0.27\%$		
		RX LTE I/Q Offset		-50 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 3800 MHz	$U=2.7$ dB		
		RX LTE Frequency Error		0.1 Hz ~ 10 kHz, 380 MHz ~ 3800 MHz	$U=3.1$ Hz		
		TX TD-SCDMA Modulation (EVM)		0.01% ~ 18%, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=0.47\%$		
		TX TD-SCDMA Modulation (Rho)		0.9 ~ 1, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=0.0019$		
		TX TD-SCDMA Power		0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=0.16$ dB		
		RX TD-SCDMA ACP		-80 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=1.1$ dB		



No. CNAS L0640

在线扫码获取验证

第 197 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		RX TD-SCDMA EVM		0.01% ~ 10%, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=0.09\%$		
		RX TD-SCDMA I/Q Imbalance		-90 dB ~ -40 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=5.7$ dB		
		RX TD-SCDMA I/Q Origin Offset		-90 dB ~ -40 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=3.0$ dB		
		RX TD-SCDMA Frequency Error		380 MHz ~ 2800 MHz, 0 dBm ~ -20 dBm	$U=2.5$ Hz		
		TX WCDMA HSDPA EVM		0.01% ~ 10%, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=1.2\%$		
		TX WCDMA EVM		0.01% ~ 10%, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=1.2\%$		
		TX WCDMA Level Error		0 dBm ~ -20 dBm, 380 MHz ~ 2800 MHz	$U=0.14$ dB		
		RX WCDMA I/Q Origin Offset		-50 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=1.8$ dB		
		RX WCDMA I/Q Imbalance		-50 dB ~ -10 dB, 0 dBm ~ -20 dBm, 380 MHz ~ 6000 MHz	$U=3.5$ dB		




No. CNAS L0640

第 198 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX WCDMA Peak Code Domain Error		-1 dB~1 dB,0 dBm~-20 dBm,380 MHz~6000 MHz	U=0.1 dB		
		RX WCDMA ACP		-80 dB~-10 dB,0 dBm~- 20 dBm,380 MHz~6000 MHz	U=1.1 dB		
		RX WCDMA EVM		0.01%~25%,0 dBm~- 20 dBm,380 MHz~6000 MHz	U=0.08%		
		RX WCDMA Frequency Error		380 MHz~2800 MHz,0 dBm~-20 dBm	U=1.2 Hz		
		TX WLAN 802.11a EVM		-53 dB~-22 dB,0 dBm~- 20 dBm	U=0.29 dB		
		TX WLAN 802.11ac EVM		-53 dB~-22 dB,0 dBm~- 20 dBm,80 MHz Bandwidth	U=0.39 dB		
				-53 dB~-22 dB,0 dBm~- 20 dBm,160 MHz Bandwidth	U=0.38 dB		
		TX WLAN 802.11b EVM		0.01%~8%,0 dBm~- 20 dBm	U=0.24%		
		TX WLAN 802.11g EVM		-53 dB~-22 dB,0 dBm~- 20 dBm	U=0.31 dB		
		TX WLAN 802.11n EVM		-53 dB~-22 dB,0 dBm~- 20 dBm,20 MHz Bandwidth	U=2.2 dB		
				-53 dB~-22 dB,0 dBm~- 20 dBm,40 MHz Bandwidth	U=0.29 dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX WLAN 802.11a (dB) EVM		-53 dB~-22 dB,0 dBm~- -20 dBm	U=1.3 dB		
		RX WLAN 802.11ac EVM		-53 dB~-22 dB,0 dBm~- -20 dBm,80 MHz Bandwidth	U=0.39 dB		
		RX WLAN 802.11ac EVM		-53 dB~-22 dB,0 dBm~- -20 dBm,160 MHz Bandwidth	U=0.38 dB		
		RX WLAN 802.11b EVM		0.01~8%,0 dBm~-20 dBm	U=0.1%		
		RX WLAN 802.11g EVM		-53 dB~-22 dB,0 dBm~- -20 dBm,DSSS	U=0.1%		
		RX WLAN 802.11g EVM		0.01%~8%,0 dBm~- -20 dBm,OFDM	U=0.48%		
		RX WLAN 802.11b/g Frequency Error		0 dBm~-20 dBm, 2484 MHz	U=3.1 Hz		
		RX WLAN 802.11b/g Chip Clock Error		$1 \times 10^{-9} \sim 1 \times 10^{-6}$, 2484 MHz Carrier	$U=5.3 \times 10^{-8}$		
		RX WLAN 802.11b/g IQ Offset		0 dBm~-20 dBm, 2484 MHz	U=1.1 dB		
		RX WLAN 802.11b/g IQ Spectrum Flatness		0 dBm~-20 dBm, 2484 MHz	U=0.4 dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		RX WLAN 802.11n EVM	ilac-M	-53 dB~-22 dB, 0 dBm~-20 dBm, 20 MHz Bandwidth	U=3.4 dB		
		RX WLAN 802.11n EVM		-53 dB~-22 dB, 0 dBm~-20 dBm, 40 MHz Bandwidth	U=3.5 dB		
		RX WLAN 802.11n Frequency Error		2484 MHz, 0 dBm~-20 dBm	U=5.0 Hz		
		RX WLAN 802.11n Chip Clock Error		$1 \times 10^{-9} \sim 1 \times 10^{-6}$, 2484 MHz Carrier	$U=2.5 \times 10^{-7}$		
		RX WLAN 802.11n IQ Offset		-90 dB~-40 dB, 0 dBm~-20 dBm	U=1.8 dB		
		RX WLAN 802.11n IQ Spectrum Flatness		-1 dB~1 dB, 0 dBm~-20 dBm	U=0.4 dB		
		TX WIMAX EVM		-50 dB~-30 dB, 2500 MHz	U=3.3 dB		
45	*High Frequency Function Generator	Timebase Frequency	High Frequency Function Generator Test Procedure CAL-CHA015-02	1 MHz ~ 10 MHz	$U_{rel}=2.1 \times 10^{-7}$		
		DC Voltage		-10 V ~ -3.5 V	$U_{rel}=5.5 \times 10^{-5}$		
				-3.5 V ~ -1.5 V	$U_{rel}=6.5 \times 10^{-5}$		
				-1.5 V ~ -0.25 V	$U_{rel}=1.5 \times 10^{-5}$		
				-0.25 V ~ -0.15 V	$U_{rel}=6.9 \times 10^{-5}$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-0.15 V ~-0.03 V	$U_{rel}=6.9 \times 10^{-5}$		
				0.03 V ~0.1 V	$U_{rel}=1.4 \times 10^{-4}$		
				0.11 V ~1.1 V	$U_{rel}=2.2 \times 10^{-4}$		
				1.2 V ~10 V	$U_{rel}=2.2 \times 10^{-4}$		
		AC Amplitude		Vpp0.4185V~2.5V,1MHz~40MHz	$U=1.6\%$		
		Flatness		(0~1) dB, 500 kHz~0.9 MHz	$U=0.055$ dB		
				(0~1) dB, 1 MHz~4.9 MHz	$U=0.053$ dB		
				(0~1) dB, 5 MHz~10 MHz	$U=0.058$ dB		
				(0~1) dB, 11 MHz~50 MHz	$U=0.060$ dB		
				(0~1) dB, 51 MHz~240 MHz	$U=0.067$ dB		
		Harmonic Distortion		-80 dB ~ -40 dB, 500 kHz~240 MHz	$U=0.64$ dB		
				-80 dB ~ -40 dB, 240 MHz~375 MHz	$U=0.73$ dB		
		Total Harmonics Distortion		0.01%~0.2%, Signal Frequency 20 kHz,peak to peak 1V	$U=0.0095\%$		
		Spurious		-80 dB~-40 dB, Fundamental Frequency 500 kHz~250 MHz	$U=0.78$ dB		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-80 dB~-40 dB, Fundamental Frequency 250 MHz~375 MHz	$U=0.89$ dB		
		Pulse Rise-Fall Time		(1~3) ns, Signal Frequency 2 MHz~10 MHz	$U=0.11$ ns		
46	*Serial/Par BERT Modules	Frequency	Serial/Par BERT Modules Test Procedure CAL- CHA027-01	300 MHz ~12.4GHz	$U_{rel}=1.7 \times 10^{-7}$		
		Output Transition Time		0.2 ps~75 ps	$U=0.19$ ps		
		Output Termination Voltage		-2V ~-0.8 V,1V~5V	$U=1.0$ mV		
		Output Delay		-0.75 ns~0.75 ns	$U=0.19$ ps		
		Output Jitter		0.03 ps~10 ps	$U=0.71$ ps		
		Output Overshoot		1mV~500 mV	$U=7.4$ mV		
		Generator Width		100 ps~999 ns	$U=2.0$ ps		
		Output Phase Noise		-100 dBc/Hz~-50 dBc/Hz, 10 GHz(10 kHz Offset)	$U=1.3$ dB		
		Input Delay		-0.75 ns~0.75 ns	$U=1.0$ ps		
		Input Phase Margin		0.03 ps~12 ps	$U=1.1$ ps		
		Input Norm Sensitivity		1 mV ~150 mV	$U=2.1$ mV		
		Delay Control Input		-110 ps~110 ps	$U=1.0$ ps		



在线扫码获取验证

No. CNAS L0640

第 203 页 共 209

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
47	*Oscilloscope Voltage Probe	Input Resistance	Oscilloscope Voltage Probe Test Procedure CAL-CHA029-01	20 k Ω ~ 1M Ω	$U_{rel}=0.053\%$		
		Bandwidth		(1.8~3)GHz, power level:0dBm~-10dBm	$U=0.070$ GHz		
				(3~5)GHz, power level:0dBm~-10dBm	$U=0.085$ GHz		
				(5~10)GHz, power level:0dBm~-10dBm	$U=0.090$ GHz		
		(10~12)GHz, power level:0dBm~-10dBm	$U=0.094$ GHz				
48	*Oscilloscope Differential Probe	DC Voltage	Oscilloscope Differential Probe Calibration Procedure CAL-CHA029-02	-50 mV~ 50 mV	$U=0.053$ mV		
		Gain		1~500	$U_{rel}=0.053\%$		
		Rise Time		40 ps~265 ps	$U=9.1$ ps		
				10 ps~40 ps	$U=4.7$ ps		
		Differential Signal Range		1~10	$U_{rel}=0.6\%$		
		Common Mode Rejection Voltage		10 mV~167 mV, Common-Mode Rejection Ratio: 60 Hz ~ 1 MHz	$U=7.0$ mV		
Common Mode Rejection Ratio	18 dB~153 dB, frequency: 1 MHz~1000 MHz	$U=0.77$ dB					
Frequency response	0.1 GHz~1 GHz	$U=0.18$ dB					
49	*Oscilloscope Current Probe	DC Current	Oscilloscope Current Probe Test Procedure CAL-CHA030-01	DC current: 50 mA ~ 100 A, Convert rate 10mV/A~1 V/A	$U_{rel}=0.22\%$		



No. CNAS L0640

第 204 页 共 209

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		AC Current	ilac-MRA	50 mA ~ 10 A, Frequency: 10 Hz ~ 50 Hz	U=0.024 A		
		AC Current (To Voltage)		50 mA ~ 10 A, Frequency: 50 Hz ~ 100 kHz Convert rate 10 mV/A ~ 100 mV/A	U=4.1 mA		
		Bandwidth		Bandwidth DC ~ 100 kHz Convert Rate 10 mV/A ~ 1V/A	U=0.12		
				Bandwidth 100 kHz ~ 100 MHz, Convert Rate 10 mV/A ~ 1V/A	U=0.065		
		Rise time		Rise time 0.1 ns ~ 23 ns Convert Rate 10 mV/A ~ 200 mV/A	U=0.080 ns		
				上升时间 0.1 ns ~ 23 ns 转换比 200 mV/A ~ 1 V/A	U=0.058 ns		
50	*Clamp Meter	DC Voltage	Clamp Meter Test Procedure CAL-CHA031-01	60 V, 300 V	U=0.053 V	合格评定 国家认可 母 认可证书专用章	
				540 V	U=0.055 V		
		AC Voltage		60 V, 50 Hz	U=0.056 V		
				300 V, 50 Hz	U=0.078 V		
				540 V, 50 Hz	U=0.15 V		
				60 V, 450 Hz	U=0.056 V		
				300 V, 450 Hz	U=0.078 V		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		DC Current	Signal Source Analyzer Test Procedure CAL-CHA033-01	540 V,450 Hz	U=0.15 V		
				4 A	U=0.012 A		
				36 A	U=0.22 A		
				540 A	U=2.6 A		
				900 A	U=4.1 A		
		AC Current		4 A,50 Hz	U=0.015 A		
				36 A,50 Hz	U=0.30 A		
				4 A,440 Hz	U=0.021 A		
				36 A,440 Hz	U=0.41 A		
		Resistance		40 Ω	U=0.055 Ω		
				360 Ω	U=0.060 Ω		
				3600 Ω	U=0.56 Ω		
		Frequency		9 Hz~450 Hz	U=0.053 Hz		
51	*Signal Source Analyzer	RF IN Port VSWR	1.001~1.6,10 MHz~30 MHz	U=0.042			
			1.001~1.2,30 MHz~2 GHz	U=0.015			
			1.001~1.3,2 GHz~3 GHz	U=0.028			
		10 MHz~30 MHz,Resolution 64 kHz,-20 dBm~0 dBm	U _{rel} =2.3x10 ⁻⁴				




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.


在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				10 MHz~30 MHz, Resolution 64 kHz, 0 dBm~20 dBm	$U_{rel}=8.9 \times 10^{-5}$		
				30 MHz~2 GHz, Resolution 64 kHz, -20 dBm~0 dBm	$U_{rel}=2.3 \times 10^{-6}$		
				30 MHz~2 GHz, Resolution 64 kHz, 0 dBm~20 dBm	$U_{rel}=1.5 \times 10^{-6}$		
				2 GHz~7 GHz, Resolution 64 kHz, -20 dBm~0 dBm	$U_{rel}=1.4 \times 10^{-7}$		
				2 GHz~7 GHz, Resolution 64 kHz, 0 dBm~20 dBm	$U_{rel}=6.3 \times 10^{-7}$		
				10 MHz~30 MHz, Resolution 1 kHz, -20 dBm~0 dBm	$U_{rel}=1.8 \times 10^{-8}$		
				10 MHz~30 MHz, Resolution 1 kHz, 0 dBm~20 dBm	$U_{rel}=4.4 \times 10^{-7}$		
				30 MHz~2 GHz, Resolution 1 kHz, -20 dBm~0 dBm	$U_{rel}=4.0 \times 10^{-9}$		
				30 MHz~2 GHz, Resolution 1 kHz, 0 dBm~20 dBm	$U_{rel}=6.3 \times 10^{-9}$		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				2 GHz~7 GHz, Resolution 1 kHz, -20 dBm~0 dBm	$U_{rel}=7.4 \times 10^{-10}$		
				2 GHz~7 GHz, Resolution 1 kHz, 0 dBm~20 dBm	$U_{rel}=2.6 \times 10^{-9}$		
				10 MHz~30 MHz, Resolution 10 Hz, -20 dBm~0 dBm	$U_{rel}=4.0 \times 10^{-9}$		
				10 MHz~30 MHz, Resolution 10 Hz, 0 dBm~20 dBm	$U_{rel}=1.8 \times 10^{-8}$		
				30 MHz~2 GHz, Resolution 10 Hz, -20 dBm~0 dBm	$U_{rel}=1.7 \times 10^{-10}$		
				30 MHz~2 GHz, Resolution 10 Hz, 0 dBm~20 dBm	$U_{rel}=4.7 \times 10^{-10}$		
				2 GHz~7 GHz, Resolution 10 Hz, -20 dBm~0 dBm	$U_{rel}=4.0 \times 10^{-11}$		
				2 GHz~7 GHz, Resolution 10 Hz, 0 dBm~20 dBm	$U_{rel}=5.9 \times 10^{-11}$		
		Power Measurement		-20 dBm~20 dBm, 10 MHz~30 MHz	$U=0.14$ dB		
				-20 dBm~20 dBm, 30 MHz~3 GHz	$U=0.13$ dB		




No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				-20 dBm~20 dBm, 3 GHz~7 GHz	U=0.16 dB		
		SA Relative Level		-2 dB~2 dB, CF 10 MHz~3 GHz, offset -7.5 MHz~7.5 MHz	U=0.17 dB		
		Phase Noise		-3 dB~3 dB, CF 1 GHz, offset -100 MHz~100 MHz	U=0.23 dB		
		Phase Noise Sensitivity		-3 dB~3 dB, CF 10 MHz, offset -1 MHz~1 MHz CF 70 MHz, offset 1 MHz~7 MHz, -180 dBc/Hz ~-160 dBc/Hz	U=0.22 dB U=2.6 dB		
		DC Power Voltage		1 V	U=0.065 mV		
				2 V	U=0.031 mV		
				5 V	U=0.30 mV		
				10 V	U=0.44 mV		
				12 V	U=0.86 mV		
		DC Current		1 mA~5 mA	U=2.5 μA		
				5 mA~80 mA	U=8 μA		



No. CNAS L0640

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证