

Agilent 34410A/11A Command Quick Reference

Syntax Conventions

Braces ({ }) enclose the parameter choices for a given command string. The braces are not sent with the command string.

A vertical bar (|) separates multiple parameter choices for a given command string. The bar is not sent with the command string.

Triangle brackets (< >) indicate that you must specify a value for the enclosed parameter. For example, the above syntax statement shows the <range> parameter enclosed in triangle brackets. The brackets are not sent with the command string. You must specify a value for the parameter (e.g., "VOLT:DC:RANG 10").

Some parameters are enclosed in square brackets ([]). The square brackets indicate that the parameter is optional and can be omitted. The brackets are not sent with the command string. If you do not specify a value for an optional parameter, the instrument chooses a default value.

Measurement Commands

```
MEASURE:CAPacitance? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:CONTinuity?
MEASURE:CURREnt:AC? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:CURREnt[:DC]? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:DIODe?
MEASURE:FREQuency? [{<range>}|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:FRESistance? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:PERiod? [{<range>}|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:RESistance? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE:TEMPerature? {FRTD|RTD|THERmistor|DEF}, {<type>}|DEF} [,1 [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE[:VOLTage]:AC? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
MEASURE[:VOLTage][:DC]? [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
```

Temperature Configuration Commands

```
CONFIGure:TEMPerature {FRTD|RTD|THERmistor|DEF}, {<type>}|DEF} [,1 [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFIGure?
[SENSe:]TEMPerature:APERture {<seconds>}|MIN|MAX|DEF}
[SENSe:]TEMPerature:APERture? [{MIN|MAX}]
[SENSe:]TEMPerature:APERture:ENABLEd?
[SENSe:]TEMPerature:NPLC {<PLCs>}|MIN|MAX|DEF}
[SENSe:]TEMPerature:NPLC? [{MIN|MAX}]
[SENSe:]TEMPerature:NULL[:STATe] {ON|OFF}
[SENSe:]TEMPerature:NULL[:STATe]?
```

```
[SENSe:]TEMPerature:NULL:VALue {<value>|MIN|MAX}
[SENSe:]TEMPerature:NULL:VALue? [{MIN|MAX}]

[SENSe:]TEMPerature:TRANsducer:TYPE {FRTD|RTD|THERmistor}
[SENSe:]TEMPerature:TRANsducer:TYPE?

[SENSe:]TEMPerature:ZERO:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]TEMPerature:ZERO:AUTO?

UNIT:TEMPerature {C|F|K}
UNIT:TEMPerature?
```

RTD Configuration

```
[SENSe:]TEMPerature:TRANsducer:FRTD:OCOMPensated {OFF|0|ON|1}
[SENSe:]TEMPerature:TRANsducer:FRTD:OCOMPensated?

[SENSe:]TEMPerature:TRANsducer:FRTD:RESistance[:REFERENCE] {<reference>|MIN|MAX|DEF}
[SENSe:]TEMPerature:TRANsducer:FRTD:RESistance[:REFERENCE]? [{MIN|MAX}]

[SENSe:]TEMPerature:TRANsducer:FRTD:TYPE {85}
[SENSe:]TEMPerature:TRANsducer:FRTD:TYPE?

[SENSe:]TEMPerature:TRANsducer:RTD:OCOMPensated {OFF|0|ON|1}
[SENSe:]TEMPerature:TRANsducer:RTD:OCOMPensated?

[SENSe:]TEMPerature:TRANsducer:RTD:RESistance[:REFERENCE] {<reference>|MIN|MAX|DEF}
[SENSe:]TEMPerature:TRANsducer:RTD:RESistance[:REFERENCE]? [{MIN|MAX}]

[SENSe:]TEMPerature:TRANsducer:RTD:TYPE {85}
[SENSe:]TEMPerature:TRANsducer:RTD:TYPE?
```

Thermistor Configuration

```
[SENSe:]TEMPerature:TRANsducer:FTHermistor:TYPE {2252|5000|10000}
[SENSe:]TEMPerature:TRANsducer:FTHermistor:TYPE?

[SENSe:]TEMPerature:TRANsducer:THERmistor:TYPE {2252|5000|10000}
[SENSe:]TEMPerature:TRANsducer:THERmistor:TYPE?
```

Voltage Configuration Commands

DC Voltage Configuration

```
CONFIGure[:VOLTage][:DC] [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF} ]
CONFIGure?

[SENSe:]VOLTage[:DC]:APERture {<seconds>|MIN|MAX|DEF}
[SENSe:]VOLTage[:DC]:APERture? [{MIN|MAX}]

[SENSe:]VOLTage[:DC]:APERture:ENABLEd?

[SENSe:]VOLTage[:DC]:IMPedance:AUTO {OFF|0|ON|1}
[SENSe:]VOLTage[:DC]:IMPedance:AUTO?

[SENSe:]VOLTage[:DC]:NPLC {<PLCs>|MIN|MAX|DEF}
[SENSe:]VOLTage[:DC]:NPLC? [{MIN|MAX}]

[SENSe:]VOLTage[:DC]:NULL[:STATe] {ON|OFF}
[SENSe:]VOLTage[:DC]:NULL[:STATe]?

[SENSe:]VOLTage[:DC]:NULL:VALue {<value>|MIN|MAX}
[SENSe:]VOLTage[:DC]:NULL:VALue? [{MIN|MAX}]

[SENSe:]VOLTage[:DC]:PEAK:STATe {ON|OFF}
[SENSe:]VOLTage[:DC]:PEAK:STATe?

[SENSe:]VOLTage[:DC]:RANGE:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]VOLTage[:DC]:RANGE:AUTO?

[SENSe:]VOLTage[:DC]:RANGE[:UPPer] {<range>|MIN|MAX|DEF}
[SENSe:]VOLTage[:DC]:RANGE[:UPPer]? [{MIN|MAX}]
```

```
[SENSe:]VOLTage[:DC]:RESolution {<resolution>}|MIN|MAX|DEF
[SENSe:]VOLTage[:DC]:RESolution? [{MIN|MAX}]

[SENSe:]VOLTage[:DC]:ZERO:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]VOLTage[:DC]:ZERO:AUTO?
```

AC Voltage Configuration

```
CONFigure[:VOLTage]:AC [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFigure?

[SENSe:]VOLTage:AC:BANDwidth {3|20|200|MIN|MAX|DEF}
[SENSe:]VOLTage:AC:BANDwidth? [{MIN|MAX}]

[SENSe:]VOLTage:AC:NULL[:STATE] {ON|OFF}
[SENSe:]VOLTage:AC:NULL[:STATE]?

[SENSe:]VOLTage:AC:NULL:VALue {<value>}|MIN|MAX
[SENSe:]VOLTage:AC:NULL:VALue? [{MIN|MAX}]

[SENSe:]VOLTage:AC:PEAK:STATe {ON|OFF}
[SENSe:]VOLTage:AC:PEAK:STATe?

[SENSe:]VOLTage:AC:RANGe:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]VOLTage:AC:RANGe:AUTO?

[SENSe:]VOLTage:AC:RANGe[:UPPer] {<range>}|MIN|MAX|DEF
[SENSe:]VOLTage:AC:RANGe[:UPPer]? [{MIN|MAX}]
```

Resistance Configuration Commands

2-Wire Resistance Configuration

```
CONFigure:RESistance [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFigure?

[SENSe:]RESistance:APERture {<seconds>}|MIN|MAX|DEF
[SENSe:]RESistance:APERture? [{MIN|MAX}]

[SENSe:]RESistance:APERture:ENABled?

[SENSe:]RESistance:NPLC {<PLCs>}|MIN|MAX|DEF
[SENSe:]RESistance:NPLC? [{MIN|MAX}]

[SENSe:]RESistance:NULL[:STATE] {ON|OFF}
[SENSe:]RESistance:NULL[:STATE]?

[SENSe:]RESistance:NULL:VALue {<value>}|MIN|MAX
[SENSe:]RESistance:NULL:VALue? [{MIN|MAX}]

[SENSe:]RESistance:OCOMpensated {OFF|0|ON|1}
[SENSe:]RESistance:OCOMpensated?

[SENSe:]RESistance:RANGe:AUTO {OFF|0|ON|1}
[SENSe:]RESistance:RANGe:AUTO?

[SENSe:]RESistance:RANGe[:UPPer] {<range>}|MIN|MAX|DEF
[SENSe:]RESistance:RANGe[:UPPer]? [{MIN|MAX}]

[SENSe:]RESistance:RESolution {<resolution>}|MIN|MAX|DEF
[SENSe:]RESistance:RESolution? [{MIN|MAX}]

[SENSe:]RESistance:ZERO:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]RESistance:ZERO:AUTO?
```

4-Wire Resistance Configuration

```
CONFigure:FRESistance [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFigure?

[SENSe:]FRESistance:APERture {<seconds>}|MIN|MAX|DEF
[SENSe:]FRESistance:APERture? [{MIN|MAX}]
```

```

[SENSe:]FRESistance:APERture:ENABLEd?
[SENSe:]FRESistance:NPLC {<PLCs>}|MIN|MAX|DEF}
[SENSe:]FRESistance:NPLC? [{MIN|MAX}]

[SENSe:]FRESistance:NULL[:STATe] {ON|OFF}
[SENSe:]FRESistance:NULL[:STATe]?

[SENSe:]FRESistance:NULL:VALue {<value>}|MIN|MAX}
[SENSe:]FRESistance:NULL:VALue? [{MIN|MAX}]

[SENSe:]FRESistance:OCOMPensated {OFF|0|ON|1}
[SENSe:]FRESistance:OCOMPensated?

[SENSe:]FRESistance:RANGE:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]FRESistance:RANGE:AUTO?

[SENSe:]FRESistance:RANGE[:UPPer] {<range>}|MIN|MAX|DEF}
[SENSe:]FRESistance:RANGE[:UPPer]? [{MIN|MAX}]

[SENSe:]FRESistance:RESolution {<resolution>}|MIN|MAX|DEF}
[SENSe:]FRESistance:RESolution? [{MIN|MAX}]

```

Current Configuration Commands

DC Current Configuration

```

CONFIGure:CURRent[:DC] [{<range>}|AUTO|MIN|MAX|DEF} [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFIGure?

[SENSe:]CURRent[:DC]:APERture {<seconds>}|MIN|MAX|DEF}
[SENSe:]CURRent[:DC]:APERture? [{MIN|MAX}]

[SENSe:]CURRent[:DC]:APERture:ENABLEd?

[SENSe:]CURRent[:DC]:NPLC {<PLCs>}|MIN|MAX|DEF}
[SENSe:]CURRent[:DC]:NPLC? [{MIN|MAX}]

[SENSe:]CURRent[:DC]:NULL[:STATe] {ON|OFF}
[SENSe:]CURRent[:DC]:NULL[:STATe]?

[SENSe:]CURRent[:DC]:NULL:VALue {<value>}|MIN|MAX}
[SENSe:]CURRent[:DC]:NULL:VALue? [{MIN|MAX}]

[SENSe:]CURRent[:DC]:PEAK:STATe {ON|OFF}
[SENSe:]CURRent[:DC]:PEAK:STATe?

[SENSe:]CURRent[:DC]:RANGE:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]CURRent[:DC]:RANGE:AUTO?

[SENSe:]CURRent[:DC]:RANGE[:UPPer] {<range>}|MIN|MAX|DEF}
[SENSe:]CURRent[:DC]:RANGE[:UPPer]? [{MIN|MAX}]

[SENSe:]CURRent[:DC]:RESolution {<resolution>}|MIN|MAX|DEF}
[SENSe:]CURRent[:DC]:RESolution? [{MIN|MAX}]

[SENSe:]CURRent[:DC]:ZERO:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]CURRent[:DC]:ZERO:AUTO?

```

AC Current Configuration

```

CONFIGure:CURRent:AC [{<range>}|AUTO|MIN|MAX|DEF} [,,{<resolution>}|MIN|MAX|DEF}] ]
CONFIGure?

[SENSe:]CURRent:AC:BANDwidth {3|20|200|MIN|MAX|DEF}
[SENSe:]CURRent:AC:BANDwidth? [{MIN|MAX}]

[SENSe:]CURRent:AC:NULL[:STATe] {ON|OFF}
[SENSe:]CURRent:AC:NULL[:STATe]?

[SENSe:]CURRent:AC:NULL:VALue {<value>}|MIN|MAX}
[SENSe:]CURRent:AC:NULL:VALue? [{MIN|MAX}]

```

```

[SENSe:]CURRent:AC:PEAK:STATe {ON|OFF}
[SENSe:]CURRent:AC:PEAK:STATe?

[SENSe:]CURRent:AC:RANGe:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]CURRent:AC:RANGe:AUTO?

[SENSe:]CURRent:AC:RANGe[:UPPer] {<range>|MIN|MAX|DEF}
[SENSe:]CURRent:AC:RANGe[:UPPer]? [{MIN|MAX}]

```

Capacitance Configuration Commands

```

CONFigure:CAPacitance [{<range>}|AUTO|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}]

[SENSe:]CAPacitance:NULL[:STATe] {ON|OFF}
[SENSe:]CAPacitance:NULL[:STATe]?

[SENSe:]CAPacitance:NULL:VALue {<value>|MIN|MAX}
[SENSe:]CAPacitance:NULL:VALue? [{MIN|MAX}]

[SENSe:]CAPacitance:RANGe:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]CAPacitance:RANGe:AUTO?

[SENSe:]CAPacitance:RANGe[:UPPer] {<range>|MIN|MAX|DEF}
[SENSe:]CAPacitance:RANGe[:UPPer]? [{MIN|MAX}]

```

Continuity and Diode Configuration Commands

```

CONFigure:CONTinuity
CONFigure:DIODE

```

Frequency and Period Configuration Commands

Frequency Configuration

```

CONFigure:FREQuency [{<range>}|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}]

CONFigure?

[SENSe:]FREQuency:APERture {<seconds>|MIN|MAX|DEF}
[SENSe:]FREQuency:APERture? [{MIN|MAX}]

[SENSe:]FREQuency:NULL[:STATe] {ON|OFF}
[SENSe:]FREQuency:NULL[:STATe]?

[SENSe:]FREQuency:NULL:VALue {<value>|MIN|MAX}
[SENSe:]FREQuency:NULL:VALue? [{MIN|MAX}]

[SENSe:]FREQuency:RANGE:LOWER {3|20|200|MIN|MAX|DEF}
[SENSe:]FREQuency:RANGE:LOWER? [{MIN|MAX}]

[SENSe:]FREQuency:VOLTage:RANGE:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]FREQuency:VOLTage:RANGE:AUTO?

[SENSe:]FREQuency:VOLTage:RANGE[:UPPer] {<voltage_range>|MIN|MAX|DEF}
[SENSe:]FREQuency:VOLTage:RANGE[:UPPer]? [{MIN|MAX}]

```

Period Configuration

```

CONFigure:PERiod [{<range>}|MIN|MAX|DEF] [,,{<resolution>}|MIN|MAX|DEF}]

CONFigure?

[SENSe:]PERiod:APERture {<seconds>|MIN|MAX|DEF}
[SENSe:]PERiod:APERture? [{MIN|MAX}]

[SENSe:]PERiod:NULL[:STATe] {ON|OFF}
[SENSe:]PERiod:NULL[:STATe]?

```

```

[SENSe:]PERiod:NULl:VALue {<value>}|MIN|MAX}
[SENSe:]PERiod:NULl:VALue? [{MIN|MAX}]

[SENSe:]PERiod:RANGe:LOWer {3|20|200|MIN|MAX|DEF}
[SENSe:]PERiod:RANGe:LOWer? [{MIN|MAX}]

[SENSe:]PERiod:VOLTage:RANGe:AUTO {OFF|0|ON|1|ONCE}
[SENSe:]PERiod:VOLTage:RANGe:AUTO?

[SENSe:]PERiod:VOLTage:RANGe[:UPPer] {<voltage_range>}|MIN|MAX|DEF}
[SENSe:]PERiod:VOLTage:RANGe[:UPPer]? [{MIN|MAX}]

```

Measurement Configuration Commands

```

ABORT

INITiate[:IMMEDIATE]

FORMAT:BORDer {NORMal|SWAPPed}
FORMAT:BORDer?

FORMAT[:DATA] {ASCII | REAL} [, <length>]
FORMAT[:DATA]?

OUTPUT:TRIGger:SLOPe {POSitive|NEGative}
OUTPUT:TRIGger:SLOPe?

READ?

ROUTe:TERMinals?

SAMPLE:COUNT {<count>}|MIN|MAX|INFinity}
SAMPLE:COUNT? [{MIN|MAX}]

SAMPLE:COUNT:PRETrigger {<PTcount>}|MIN|MAX}
SAMPLE:COUNT:PRETrigger? [{MIN|MAX}]

SAMPLE:SOURce {AUTO|TImer}
SAMPLE:SOURce?

SAMPLE:TImer {<interval>}|MIN|MAX}
SAMPLE:TImer? [{MIN|MAX}]

TRIGger:COUNT {<count>}|MIN|MAX|DEF|INFinity}
TRIGger:COUNT? [{MIN|MAX}]

TRIGger:DELay {<seconds>}|MIN|MAX}
TRIGger:DELay? [{MIN|MAX}]

TRIGger:DELay:AUTO {OFF|0|ON|1}>
TRIGger:DELay:AUTO?

TRIGger:LEVel{<level>}|MIN|MAX}
TRIGger:LEVel? [{MIN|MAX}]

TRIGger:SLOPe {POSitive|NEGative}
TRIGger:SLOPe?

TRIGger:SOURce {IMMEDIATE|EXTernal|BUS}
TRIGger:SOURce?

```

Triggering Commands

```

*TRG

INITiate[:IMMEDIATE]

OUTPUT:TRIGger:SLOPe {POSitive|NEGative}
OUTPUT:TRIGger:SLOPe?

READ?

TRIGger:COUNT {<count>}|MIN|MAX|DEF|INFinity}
TRIGger:COUNT? [{MIN|MAX}]

```

```

TRIGger:DELay {<seconds>}|MIN|MAX}
TRIGger:DELay? [{MIN|MAX}]

TRIGger:DELay:AUTO {OFF|0|ON|1}
TRIGger:DELay:AUTO?

TRIGger:LEVel{</eve/>}|MIN|MAX}
TRIGger:LEVel? [{MIN|MAX}]

TRIGger:SLOPe {POSitive|NEGative}
TRIGger:SLOPe?

TRIGger:SOURce {IMMediate|EXTernal|BUS}
TRIGger:SOURce?

```

Calculation (Math) Commands

```

CALCulate:AVERage:AVERage?
CALCulate:AVERage:CLEar
CALCulate:AVERage:COUNT?
CALCulate:AVERage:MAX?
CALCulate:AVERage:MIN?
CALCulate:AVERage:PTPeak?
CALCulate:AVERage:SDEViation?
CALCulate:DB:REFerence {<value> | MIN | MAX}
CALCulate:DB:REFerence? {MIN | MAX}

CALCulate:DBM:REFerence {<value> | MIN | MAX}
CALCulate:DBM:REFerence? {MIN | MAX}

CALCulate:FUNCTION {NULL | DB | DBM | AVERage | LIMit}
CALCulate:FUNCTION?

CALCulate:LIMit:LOWER {<value> | MIN | MAX}
CALCulate:LIMit:LOWER? {MIN | MAX}

CALCulate:LIMit:UPPer {<value> | MIN | MAX}
CALCulate:LIMit:UPPer? {MIN | MAX}

CALCulate:NULL:OFFSet {<value> | MIN | MAX}
CALCulate:NULL:OFFSet? {MIN | MAX}

CALCulate[:STATe] {OFF | ON}
CALCulate:STATe?

```

Reading Memory Commands

```

DATA:LAST?
DATA:COPY NVMEM, RDG_STORE
DATA:DATA? NVMEM
DATA:DELet NVMEM
DATA:POINTS:EVENT:THreshold <num_readings>
DATA:POINTS:EVENT:THreshold?

DATA:POINTs?
DATA:REMove? <num_readings>
FETCH?
FETCH:CURREnt:AC:PTPeak?
FETCH:CURREnt[:DC]:PEAK:MAX?

```

```

FETCH:CURREnt[:DC]:PEAK:MIN?
FETCH:CURREnt[:DC]:PTPeak?
FETCH:VOLTage:AC:PTPeak?
FETCH:VOLTage[:DC]:PEAK:MAX?
FETCH:VOLTage[:DC]:PEAK:MIN?
FETCH:VOLTage[:DC]:PTPeak?
R? [<max_count>]

```

Calibration Commands

```

CALibration:ADC?
CALibration[:ALL]?
CALibration:COUNt?
CALibration:LFREquency {50|60}
CALibration:LFREquency?
CALibration:LFREquency:ACTual?
CALibration:SECure:CODE <new_code>
CALibration:SECure:STATE {OFF|0|ON|1}, <code>
CALibration:SECure:STATE?
CALibration:STORe
CALibration:STRing "<string>"
CALibration:STRing?
CALibration:VALue <value>
CALibration:VALue?

```

State Storage Commands

```

*RCL {0|1|2|3|4}
*SAV {0|1|2|3|4}
MEMory:NSTates?
MEMory:STATE:CATalog?
MEMory:STATE:DELetE {0|1|2|3|4}
MEMory:STATE:DELetE:ALL
MEMory:STATE:NAME {0|1|2|3|4} [,<name>]
MEMory:STATE:NAME? {0|1|2|3|4}
MEMory:STATE:RECall:AUTO {OFF|0|ON|1}
MEMory:STATE:RECall:AUTO?
MEMory:STATE:RECall:SELect {0|1|2|3|4}
MEMory:STATE:RECall:SELect?
MEMory:STATE:VALid? {0|1|2|3|4}

```

IEEE-488 Commands

```

*CLS
*ESE <enable_value>
*ESE?
*ESR?

```

*IDN?
*LRN?
*OPC
*OPC?
*PSC {0|1}
*PSC?
*RCL {0|1|2|3|4}
*RST
*SAV {0|1|2|3|4}
*SRE <enable_value>
*SRE?
*STB?
*TRG
*TST?
*WAI

System-Related Commands

*IDN?
*RST
*TST?
CALibration:LFREquency?
DISPLAY[:WINDOW[1|2][:STATE]] {OFF|0|ON|1}
DISPLAY[:WINDOW[1|2][:STATE]]?
DISPLAY[:WINDOW{1|2}]:TEXT:CLEar
DISPLAY[:WINDOW{1|2}]:TEXT[:DATA] "<string>"
DISPLAY[:WINDOW{1|2}]:TEXT[:DATA]?
DISPLAY:WINDOW2:TEXT:FEED "<feed>"
DISPLAY:WINDOW2:TEXT:FEED?
SYSTem:BEEPer[:IMMediate]
SYSTem:BEEPer:STATe {OFF|0|ON|1}
SYSTem:BEEPer:STATe?
SYSTem:ERRor[:NEXT]?
SYSTem:HELP?
SYSTem:LANGuage "{34401A|34410A|34411A}"
SYSTem:LANGuage?
SYSTem:LFREquency:ACTual?
SYSTem:LFREquency?
SYSTem:PRESet
SYSTem:SECurity:IMMediate
SYSTem:VERSion?

Remote Interface Configuration Commands

```

SYSTEM:COMMunicate:ENABLE {OFF|0|ON|1}, {GPIB|USB|LAN|SOCKets|TELNet|VXI11|WEB}
SYSTEM:COMMunicate:ENABLE? {GPIB|USB|LAN|SOCKets|TELNet|VXI11|WEB}

SYSTEM:COMMunicate:GPIB[:SELF]:ADDRess {<address>}
SYSTEM:COMMunicate:GPIB[:SELF]:ADDRess?

SYSTEM:LOCK:NAME?
SYSTEM:LOCK:OWNer?
SYSTEM:LOCK:RELease
SYSTEM:LOCK:REQuest?

```

LAN Configuration Commands

```

SYSTEM:COMMunicate:LAN:AUTOip[:STATe] {OFF|0|ON|1}
SYSTEM:COMMunicate:LAN:AUTOip[:STATe]?

SYSTEM:COMMunicate:LAN:BSTatus?

SYSTEM:COMMunicate:LAN:CONTrol?

SYSTEM:COMMunicate:LAN:DDNS {OFF|0|ON|1}
SYSTEM:COMMunicate:LAN:DDNS?

SYSTEM:COMMunicate:LAN:DHCP {OFF|0|ON|1}
SYSTEM:COMMunicate:LAN:DHCP?

SYSTEM:COMMunicate:LAN:DNS <address>
SYSTEM:COMMunicate:LAN:DNS?

SYSTEM:COMMunicate:LAN:DOMain "<name>" 
SYSTEM:COMMunicate:LAN:DOMain? [{CURRent|STATic}]

SYSTEM:COMMunicate:LAN:GATEway <address>
SYSTEM:COMMunicate:LAN:GATEway? [{CURRent|STATic}]

SYSTEM:COMMunicate:LAN:HISTory:CLEar
SYSTEM:COMMunicate:LAN:HISTory?

SYSTEM:COMMunicate:LAN:HOSTname "<name>" 
SYSTEM:COMMunicate:LAN:HOSTname? [{CURRent|STATic}]

SYSTEM:COMMunicate:LAN:IPAddress <address>
SYSTEM:COMMunicate:LAN:IPAddress? [{CURRent|STATic}]

SYSTEM:COMMunicate:LAN:KEEPalive {<seconds>}|MIN|MAX
SYSTEM:COMMunicate:LAN:KEEPalive? [{MIN|MAX}]

SYSTEM:COMMunicate:LAN:LIPaddress?

SYSTEM:COMMunicate:LAN:MEDIasense {OFF|0|ON|1}
SYSTEM:COMMunicate:LAN:MEDIasense?

SYSTEM:COMMunicate:LAN:MAC?

SYSTEM:COMMunicate:LAN:NETBios {OFF|0|ON|1}
SYSTEM:COMMunicate:LAN:NETBios?

SYSTEM:COMMunicate:LAN:SMASK <mask>
SYSTEM:COMMunicate:LAN:SMASK? [{CURRent|STATic}]

SYSTEM:COMMunicate:LAN:TELNet:PROMpt "<string>" 
SYSTEM:COMMunicate:LAN:TELNet:PROMpt?

SYSTEM:COMMunicate:LAN:TELNet:WMESsage "<string>" 
SYSTEM:COMMunicate:LAN:TELNet:WMESsage?

```

Status System Commands

*CLS

```
*ESE <enable_value>
*ESE?
*ESR?
*PSC {0|1}
*PSC?
*SRE <enable_value>
*SRE?
*STB?
STATUS:OPERation:CONDition?
STATUS:OPERation:ENABLE <enable_value>
STATUS:OPERation:ENABLE?

STATUS:OPERation[:EVENT]?
STATUS:PRESet
STATUS:QUEstionable:CONDition?
STATUS:QUEstionable:ENABLE <enable_value>
STATUS:QUEstionable:ENABLE?

STATUS:QUEstionable[:EVENT]?
```

Copyright © 2005, 2006 Agilent Technologies, Inc.
January 2006